

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.764	4.666	4.866	475.480	500.000	-4.9
Aroclor-1016-2	4.783	4.684	4.884	494.560	500.000	-1.1
Aroclor-1016-3	4.840	4.741	4.941	485.280	500.000	-2.9
Aroclor-1016-4	4.960	4.861	5.061	484.640	500.000	-3.1
Aroclor-1016-5	5.217	5.118	5.318	505.120	500.000	1.0
Aroclor-1260-1	6.254	6.157	6.357	486.680	500.000	-2.7
Aroclor-1260-2	6.444	6.346	6.546	527.440	500.000	5.5
Aroclor-1260-3	6.811	6.713	6.913	531.610	500.000	6.3
Aroclor-1260-4	7.070	6.972	7.172	535.590	500.000	7.1
Aroclor-1260-5	7.313	7.216	7.416	516.860	500.000	3.4
Decachlorobiphenyl	8.708	8.612	8.812	45.020	50.000	-10.0
Tetrachloro-m-xylene	3.676	3.577	3.777	51.620	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.749	4.651	4.851	492.000	500.000	-1.6
Aroclor-1016-2	4.767	4.669	4.869	498.410	500.000	-0.3
Aroclor-1016-3	4.942	4.844	5.044	493.840	500.000	-1.2
Aroclor-1016-4	4.984	4.887	5.087	490.000	500.000	-2.0
Aroclor-1016-5	5.196	5.099	5.299	486.890	500.000	-2.6
Aroclor-1260-1	6.226	6.129	6.329	482.400	500.000	-3.5
Aroclor-1260-2	6.414	6.316	6.516	502.890	500.000	0.6
Aroclor-1260-3	6.566	6.469	6.669	475.680	500.000	-4.9
Aroclor-1260-4	7.036	6.939	7.139	469.160	500.000	-6.2
Aroclor-1260-5	7.277	7.181	7.381	474.960	500.000	-5.0
Decachlorobiphenyl	8.656	8.561	8.761	46.490	50.000	-7.0
Tetrachloro-m-xylene	3.672	3.573	3.773	48.950	50.000	-2.1