

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.641	5.546	5.746	485.990	500.000	-2.8
Aroclor-1016-2	5.662	5.567	5.767	533.210	500.000	6.6
Aroclor-1016-3	5.725	5.629	5.829	520.180	500.000	4.0
Aroclor-1016-4	5.822	5.727	5.927	549.860	500.000	10.0
Aroclor-1016-5	6.114	6.019	6.219	524.760	500.000	5.0
Aroclor-1260-1	7.231	7.136	7.336	560.880	500.000	12.2
Aroclor-1260-2	7.485	7.389	7.589	543.950	500.000	8.8
Aroclor-1260-3	7.842	7.747	7.947	544.400	500.000	8.9
Aroclor-1260-4	8.066	7.971	8.171	548.260	500.000	9.7
Aroclor-1260-5	8.384	8.288	8.488	516.290	500.000	3.3
Decachlorobiphenyl	10.181	10.087	10.287	50.830	50.000	1.7
Tetrachloro-m-xylene	4.490	4.394	4.594	51.690	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.862	4.768	4.968	473.800	500.000	-5.2
Aroclor-1016-2	4.880	4.786	4.986	482.540	500.000	-3.5
Aroclor-1016-3	5.056	4.962	5.162	483.380	500.000	-3.3
Aroclor-1016-4	5.098	5.004	5.204	477.720	500.000	-4.5
Aroclor-1016-5	5.312	5.218	5.418	546.190	500.000	9.2
Aroclor-1260-1	6.342	6.249	6.449	495.490	500.000	-0.9
Aroclor-1260-2	6.531	6.438	6.638	513.470	500.000	2.7
Aroclor-1260-3	6.683	6.589	6.789	485.640	500.000	-2.9
Aroclor-1260-4	7.152	7.060	7.260	489.320	500.000	-2.1
Aroclor-1260-5	7.395	7.302	7.502	489.940	500.000	-2.0
Decachlorobiphenyl	8.788	8.697	8.897	50.670	50.000	1.3
Tetrachloro-m-xylene	3.783	3.688	3.888	47.840	50.000	-4.3