

LAB CHRONICLE

OrderID:	Q2425	OrderDate:	6/25/2025 2:03:00 PM
Client:	Kleinfelder	Project:	AS Jenks School
Contact:	Mark Warchol	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2425-01	COMP-1	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25
Q2425-02	COMP-2	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25
Q2425-03	COMP-3	SOIL	PCB Group1	8082A	06/24/25	06/26/25	06/26/25	06/25/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.:	Q2425	Order ID:	Q2425
Client:	Kleinfelder	Project ID:	AS Jenks School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				Total Concentration:	0.000			



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-1	SDG No.:	Q2425
Lab Sample ID:	Q2425-01	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	76.9
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111866.D	1	06/26/25 08:20	06/26/25 13:12	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.10	U	5.10	22.1	ug/kg
11097-69-1	Aroclor-1254	4.20	U	4.20	22.1	ug/kg
11096-82-5	Aroclor-1260	4.20	U	4.20	22.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.4		32 - 144	82%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-2	SDG No.:	Q2425
Lab Sample ID:	Q2425-02	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	79.7
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111867.D	1	06/26/25 08:20	06/26/25 13:30	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.3	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.3	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/25/25
Client Sample ID:	COMP-3	SDG No.:	Q2425
Lab Sample ID:	Q2425-03	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	82.6
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111868.D	1	06/26/25 08:20	06/26/25 13:48	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.5	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.5	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.1		32 - 144	106%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.7		32 - 175	98%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit



QC SUMMARY

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO111586.D	PIBLK-PO111586.D	Tetrachloro-m-xylene	1	20	20.3	102		60	140
		Decachlorobiphenyl	1	20	20.7	104		60	140
		Tetrachloro-m-xylene	2	20	19.0	95		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140
I.BLK-PO111862.D	PIBLK-PO111862.D	Tetrachloro-m-xylene	1	20	20.5	102		60	140
		Decachlorobiphenyl	1	20	19.8	99		60	140
		Tetrachloro-m-xylene	2	20	18.4	92		60	140
		Decachlorobiphenyl	2	20	19.4	97		60	140
Q2425-01	COMP-1	Tetrachloro-m-xylene	1	20	16.4	82		32	144
		Decachlorobiphenyl	1	20	15.8	79		32	175
		Tetrachloro-m-xylene	2	20	14.6	73		32	144
		Decachlorobiphenyl	2	20	16.5	82		32	175
Q2425-02	COMP-2	Tetrachloro-m-xylene	1	20	18.6	93		32	144
		Decachlorobiphenyl	1	20	16.5	83		32	175
		Tetrachloro-m-xylene	2	20	16.9	85		32	144
		Decachlorobiphenyl	2	20	16.8	84		32	175
Q2425-03	COMP-3	Tetrachloro-m-xylene	1	20	21.1	106		32	144
		Decachlorobiphenyl	1	20	19.2	96		32	175
		Tetrachloro-m-xylene	2	20	19.4	97		32	144
		Decachlorobiphenyl	2	20	19.7	98		32	175
I.BLK-PO111875.D	PIBLK-PO111875.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	18.1	90		60	140
		Decachlorobiphenyl	2	20	19.5	97		60	140
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073283.D	PIBLK-PP073283.D	Tetrachloro-m-xylene	1	20	18.9	94		60	140
		Decachlorobiphenyl	1	20	18.6	93		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	20.9	105		60	140
PB168622BL	PB168622BL	Tetrachloro-m-xylene	1	20	19.3	96		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.8	94		32	144
		Decachlorobiphenyl	2	20	21.1	106		32	175
PB168622BS	PB168622BS	Tetrachloro-m-xylene	1	20	20.2	101		32	144
		Decachlorobiphenyl	1	20	20.0	100		32	175
		Tetrachloro-m-xylene	2	20	18.6	93		32	144
		Decachlorobiphenyl	2	20	23.0	115		32	175
Q2409-02MS	COP-SOIL-PILEMS	Tetrachloro-m-xylene	1	20	20.0	100		32	144

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2409-02MS	COP-SOIL-PILEMS	Decachlorobiphenyl	1	20	22.5	112		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
		Decachlorobiphenyl	2	20	22.7	113		32	175
Q2409-02MSD	COP-SOIL-PILEMSD	Tetrachloro-m-xylene	1	20	19.3	97		32	144
		Decachlorobiphenyl	1	20	23.0	115		32	175
		Tetrachloro-m-xylene	2	20	19.2	96		32	144
		Decachlorobiphenyl	2	20	22.7	114		32	175
		Tetrachloro-m-xylene	1	20	18.3	91		60	140
I.BLK-PP073298.D	PIBLK-PP073298.D	Decachlorobiphenyl	1	20	18.9	95		60	140
		Tetrachloro-m-xylene	2	20	18.7	93		60	140
		Decachlorobiphenyl	2	20	20.5	103		60	140
		Tetrachloro-m-xylene	1	20	18.3	91		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PP073292.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2409-02MS (Column 1)	COP-SOIL-PILEMS AR1016	172.7	0	163	ug/kg	94				55	146	
	AR1260	172.7	31	195	ug/kg	95				54	119	
Client Sample ID: Q2409-02MS (Column 2)	COP-SOIL-PILEMS AR1016	172.7	0	154	ug/kg	89				55	146	
	AR1260	172.7	28	169	ug/kg	98				54	119	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	DataFile :	<u>PP073293.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2409-02MSD (Column 1)	COP-SOIL-PILEMSD AR1016	172.5	0	163	ug/kg	94		0		55	146	15
	AR1260	172.5	31	202	ug/kg	99		4		54	119	15
Client Sample ID: Q2409-02MSD (Column 2)	COP-SOIL-PILEMSD AR1016	172.5	0	158	ug/kg	92		3		55	146	15
	AR1260	172.5	28	174	ug/kg	101		3		54	119	15

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PP073290.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168622BS (Column 1)	AR1016	166.6	160	ug/kg	96				71	120		
	AR1260	166.6	159	ug/kg	95				65	130		
PB168622BS (Column 2)	AR1016	166.6	149	ug/kg	89				71	120		
	AR1260	166.6	157	ug/kg	94				65	130		

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168622BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q2425

SAS No.: Q2425 SDG NO.: Q2425

Lab Sample ID: PB168622BL

Lab File ID: PP073289.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 06/26/2025

Date Analyzed (1): 06/26/2025

Date Analyzed (2): 06/26/2025

Time Analyzed (1): 12:25

Time Analyzed (2): 12:25

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
COMP-1	Q2425-01	PO111866.D	06/26/2025	06/26/2025
COMP-2	Q2425-02	PO111867.D	06/26/2025	06/26/2025
COMP-3	Q2425-03	PO111868.D	06/26/2025	06/26/2025
PB168622BS	PB168622BS	PP073290.D	06/26/2025	06/26/2025
COP-SOIL-PILEMS	Q2409-02MS	PP073292.D	06/26/2025	06/26/2025
COP-SOIL-PILEMSD	Q2409-02MSD	PP073293.D	06/26/2025	06/26/2025

COMMENTS: _____



CALIBRATION SUMMARY

RETENTION TIMES OF INITIAL CALIBRATION

Contract:	<u>POWE02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2425</u>	SAS No.:	<u>Q2425</u>	SDG NO.:	<u>Q2425</u>
Instrument ID:	<u>ECD_O</u>	Calibration Date(s):		<u>06/11/2025</u>	<u>06/11/2025</u>		
		Calibration Times:		<u>10:40</u>	<u>18:31</u>		

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PO111587.D</u>	RT 750 =	<u>PO111588.D</u>
	RT 500 =	PO111589.D	RT 250 =	PO111590.D
			RT 050 =	PO111591.D

[illegible]

A
B
C
D
E
F
G

Calibration Times: 10:40 18:31

LAB FILE ID:	RT 1000 =	<u>PO111587.D</u>	RT 750 =	<u>PO111588.D</u>
	RT 500 =	PO111589.D	RT 250 =	PO111590.D
			RT 050 =	PO111591.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_O **Calibration Date(s):** 06/11/2025 06/11/2025
Calibration Times: 10:40 18:31

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO111587.D	CF 750 =	PO111588.D			
CF 500 =		PO111589.D	CF 250 =	PO111590.D	CF 050 =	PO111591.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	220511854	230179169	244104994	255923844	242047420	238553456	6
Aroclor-1016-2	(2)	319525390	333752768	345607420	360231916	304600680	332743635	7
Aroclor-1016-3	(3)	218699222	228973452	241664208	255477324	205889780	230140797	8
Aroclor-1016-4	(4)	174886969	182974663	191580278	204290796	173581000	185462741	7
Aroclor-1016-5	(5)	177464979	184697736	193390828	207039800	170624200	186643509	8
Aroclor-1260-1	(1)	321794228	337454645	358028956	426721964	332609080	355321775	12
Aroclor-1260-2	(2)	433743230	452724105	474832224	506755684	494307920	472472633	6
Aroclor-1260-3	(3)	393522712	410790572	429624930	441759844	447671720	424673956	5
Aroclor-1260-4	(4)	282329615	299358140	315388464	327883556	347275780	314447111	8
Aroclor-1260-5	(5)	781713901	810719289	832144010	867226572	826835640	823727882	4
Decachlorobiphenyl		4901811500	5109294200	5313242720	5581051480	5340821400	5249244260	5
Tetrachloro-m-xylene		5558158730	5743696387	5927517640	6035511000	5532196200	5759415991	4
Aroclor-1254-1	(1)	375291496	390445984	409053510	418326128	532236220	425070668	15
Aroclor-1254-2	(2)	334554268	347730616	365467502	374564136	475325660	379528436	15
Aroclor-1254-3	(3)	528273778	541799637	566990470	566909168	652766900	571347991	8
Aroclor-1254-4	(4)	330811190	344690741	356705876	353681912	358524400	348882824	3
Aroclor-1254-5	(5)	481176015	491854427	511981248	504869196	570769000	512129977	7
Decachlorobiphenyl		5009210160	5167761240	5403649020	5380598200	6191612600	5430566244	8
Tetrachloro-m-xylene		5724223470	6030704747	6021777640	5926293200	6653717200	6071343251	6

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_O **Calibration Date(s):** 06/11/2025 06/11/2025
Calibration Times: 10:40 18:31

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO111587.D	CF 750 =	PO111588.D			
CF 500 =		PO111589.D	CF 250 =	PO111590.D	CF 050 =	PO111591.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	180839726	189231809	196263600	204676936	212823680	196767150	6
Aroclor-1016-2	(2)	269525798	277038755	287808724	296313292	293716580	284880630	4
Aroclor-1016-3	(3)	141401189	146814288	152988496	158088464	151904200	150239327	4
Aroclor-1016-4	(4)	112384684	117918833	124672358	128074364	125886140	121787276	5
Aroclor-1016-5	(5)	145637432	151733320	158595618	164650128	175375940	159198488	7
Aroclor-1260-1	(1)	232933493	242290633	253160914	265400132	260795120	250916058	5
Aroclor-1260-2	(2)	272930024	283658681	293882326	306210008	313403160	294016840	6
Aroclor-1260-3	(3)	248293091	256892847	265761412	279567532	285382000	267179376	6
Aroclor-1260-4	(4)	171150736	180595663	190038668	201212768	207483680	190096303	8
Aroclor-1260-5	(5)	395557404	413148045	424269490	442012232	450112560	425019946	5
Decachlorobiphenyl		1664373530	1742647173	1808169220	1884747160	1788371000	1777661617	5
Tetrachloro-m-xylene		5585912110	5701804613	5792994800	5795065440	5210609600	5617277313	4
Aroclor-1254-1	(1)	296936934	303927659	319123596	320561628	374491800	323008323	9
Aroclor-1254-2	(2)	255453905	262352463	275811104	277251804	325698480	279313551	10
Aroclor-1254-3	(3)	392851690	400205872	416928596	411547332	473734340	419053566	8
Aroclor-1254-4	(4)	214635691	222689013	230857806	231773900	239050800	227801442	4
Aroclor-1254-5	(5)	301414788	308036695	323788090	319823096	374889540	325590442	9
Decachlorobiphenyl		1681329490	1723219680	1812787460	1794853560	2027152800	1807868598	7
Tetrachloro-m-xylene		5554099840	5630982267	5714907100	5492046840	5804310800	5639269369	2

A
B
C
D
E
F
G

Calibration Times: 10:04 19:37

LAB FILE ID:	RT 1000 =	<u>PP072991.D</u>	RT 750 =	<u>PP072992.D</u>
	RT 500 =	PP072993.D	RT 250 =	PP072994.D
			RT 050 =	PP072995.D

[illegible]

A
B
C
D
E
F
G

Calibration Times: 10:04 19:37

LAB FILE ID:	RT 1000 =	<u>PP072991.D</u>	RT 750 =	<u>PP072992.D</u>
	RT 500 =	PP072993.D	RT 250 =	PP072994.D
			RT 050 =	PP072995.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 19:37

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP072991.D	CF 750 =	PP072992.D			
CF 500 =		PP072993.D	CF 250 =	PP072994.D	CF 050 =	PP072995.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	63671603	67644257	70148492	76963128	84930220	72671540	12
Aroclor-1016-2	(2)	97440661	102354848	107008936	114458460	98840440	104020669	7
Aroclor-1016-3	(3)	59983524	63267185	65041372	73554480	64895380	65348388	8
Aroclor-1016-4	(4)	50589003	52912481	54400336	56287652	47214800	52280854	7
Aroclor-1016-5	(5)	46240143	47840919	49349310	51433860	55161280	50005102	7
Aroclor-1260-1	(1)	82538787	86792979	90284630	94211564	87238820	88213356	5
Aroclor-1260-2	(2)	125600539	132072107	137363486	143645412	140737120	135883733	5
Aroclor-1260-3	(3)	106201830	111185451	115341698	120182220	117849880	114152216	5
Aroclor-1260-4	(4)	98670506	102821608	101520658	108282220	103932620	103045522	3
Aroclor-1260-5	(5)	227637841	236485680	245198018	255372264	238418920	240622545	4
Decachlorobiphenyl		1620648950	1677444467	1749703060	1784234960	1712434400	1708893167	4
Tetrachloro-m-xylene		1977356170	2059482187	2109826840	2243117400	2164249200	2110806359	5
Aroclor-1254-1	(1)	73763285	78885443	81757104	88891384	97572240	84173891	11
Aroclor-1254-2	(2)	111451506	118674340	123357828	133620984	139899720	125400876	9
Aroclor-1254-3	(3)	119235877	126285657	130365552	141133040	139034740	131210973	7
Aroclor-1254-4	(4)	108754762	114515340	117946542	128101128	129602040	119783962	7
Aroclor-1254-5	(5)	106940669	112254864	115991910	123909512	112185340	114256459	6
Decachlorobiphenyl		1648218000	1739306053	1775656020	1935470960	1659491600	1751628527	7
Tetrachloro-m-xylene		1872735760	1960277133	2019595300	2159825040	2074265000	2017339647	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q2425 **SAS No.:** Q2425 **SDG NO.:** Q2425

Instrument ID: ECD_P **Calibration Date(s):** 06/17/2025 06/17/2025
Calibration Times: 10:04 19:37

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP072991.D</u>		CF 750 = <u>PP072992.D</u>				
CF 500 = <u>PP072993.D</u>		CF 250 = <u>PP072994.D</u>		CF 050 = <u>PP072995.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	58734267	64000205	67674152	72695560	80484240	68717685	12
Aroclor-1016-2	(2)	90382349	94344453	98038690	104433472	115993580	100638509	10
Aroclor-1016-3	(3)	48136215	50748025	53595712	56975444	64125620	54716203	11
Aroclor-1016-4	(4)	38107893	40907291	43245840	46954372	53532680	44549615	13
Aroclor-1016-5	(5)	49203503	51615780	54910048	58505108	62177080	55282304	9
Aroclor-1260-1	(1)	84892100	88579871	95448406	100236824	104495860	94730612	9
Aroclor-1260-2	(2)	105588352	108160505	115763440	121715240	130953840	116436275	9
Aroclor-1260-3	(3)	96095206	97763095	104346708	109521728	113078520	104161051	7
Aroclor-1260-4	(4)	77788306	79915727	86740682	92092068	102155800	87738517	11
Aroclor-1260-5	(5)	196041650	198840268	209548214	218270728	244301260	213400424	9
Decachlorobiphenyl		1225771100	1248497173	1342901640	1435827000	1420581600	1334715703	7
Tetrachloro-m-xylene		1662896120	1767446653	1788437460	1893713440	1896490200	1801796775	5
Aroclor-1254-1	(1)	94131500	102369100	108358230	120483728	122907600	109650032	11
Aroclor-1254-2	(2)	80645865	88030495	92596756	103733056	107725280	94546290	12
Aroclor-1254-3	(3)	128776319	138178049	147317104	161124892	162339800	147547233	10
Aroclor-1254-4	(4)	83562949	90268741	96675708	106896836	108101020	97101051	11
Aroclor-1254-5	(5)	116424396	122854644	129598642	142067368	138868340	129962678	8
Decachlorobiphenyl		1234350420	1314370427	1354328000	1438665160	1427909200	1353924641	6
Tetrachloro-m-xylene		1670033590	1775346533	1836392320	1937020080	1788575000	1801473505	5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:06 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-2	(2)	4.79	4.78	4.68	4.88	-0.01
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.26	6.26	6.16	6.36	0.00
Aroclor-1260-2	(2)	6.45	6.45	6.35	6.55	0.00
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.32	7.32	7.22	7.42	0.00
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 10:06 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-4	(4)	7.04	7.04	6.94	7.14	0.00
Aroclor-1260-5	(5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.767	4.666	4.866	493.110	500.000	-1.4
Aroclor-1016-2	4.786	4.684	4.884	509.120	500.000	1.8
Aroclor-1016-3	4.842	4.741	4.941	505.140	500.000	1.0
Aroclor-1016-4	4.962	4.861	5.061	503.830	500.000	0.8
Aroclor-1016-5	5.220	5.118	5.318	530.310	500.000	6.1
Aroclor-1260-1	6.258	6.157	6.357	513.670	500.000	2.7
Aroclor-1260-2	6.448	6.346	6.546	553.480	500.000	10.7
Aroclor-1260-3	6.813	6.713	6.913	575.780	500.000	15.2
Aroclor-1260-4	7.074	6.972	7.172	564.630	500.000	12.9
Aroclor-1260-5	7.316	7.216	7.416	545.190	500.000	9.0
Decachlorobiphenyl	8.712	8.612	8.812	48.170	50.000	-3.7
Tetrachloro-m-xylene	3.679	3.577	3.777	53.720	50.000	7.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111858.D Time Analyzed: 10:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.651	4.851	495.750	500.000	-0.9
Aroclor-1016-2	4.766	4.669	4.869	504.470	500.000	0.9
Aroclor-1016-3	4.941	4.844	5.044	501.230	500.000	0.2
Aroclor-1016-4	4.984	4.887	5.087	498.050	500.000	-0.4
Aroclor-1016-5	5.196	5.099	5.299	502.020	500.000	0.4
Aroclor-1260-1	6.226	6.129	6.329	501.900	500.000	0.4
Aroclor-1260-2	6.414	6.316	6.516	526.020	500.000	5.2
Aroclor-1260-3	6.566	6.469	6.669	497.850	500.000	-0.4
Aroclor-1260-4	7.036	6.939	7.139	497.700	500.000	-0.5
Aroclor-1260-5	7.277	7.181	7.381	508.850	500.000	1.8
Decachlorobiphenyl	8.658	8.561	8.761	47.550	50.000	-4.9
Tetrachloro-m-xylene	3.671	3.573	3.773	49.570	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 15:37 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.84	4.84	4.74	4.94	0.00
Aroclor-1016-4	(4)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-5	(5)	5.22	5.22	5.12	5.32	0.00
Aroclor-1260-1	(1)	6.25	6.26	6.16	6.36	0.01
Aroclor-1260-2	(2)	6.44	6.45	6.35	6.55	0.01
Aroclor-1260-3	(3)	6.81	6.81	6.71	6.91	0.00
Aroclor-1260-4	(4)	7.07	7.07	6.97	7.17	0.00
Aroclor-1260-5	(5)	7.31	7.32	7.22	7.42	0.01
Tetrachloro-m-xylene		3.68	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.71	8.71	8.61	8.81	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/11/2025 06/11/2025

Continuing Calib Time: 15:37 Initial Calibration Time(s): 10:40 18:31

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.75	4.65	4.85	0.00
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.94	4.94	4.84	5.04	0.00
Aroclor-1016-4	(4)	4.98	4.99	4.89	5.09	0.01
Aroclor-1016-5	(5)	5.20	5.20	5.10	5.30	0.00
Aroclor-1260-1	(1)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-4	(4)	7.04	7.04	6.94	7.14	0.00
Aroclor-1260-5	(5)	7.28	7.28	7.18	7.38	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.00
Decachlorobiphenyl		8.66	8.66	8.56	8.76	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.764	4.666	4.866	475.480	500.000	-4.9
Aroclor-1016-2	4.783	4.684	4.884	494.560	500.000	-1.1
Aroclor-1016-3	4.840	4.741	4.941	485.280	500.000	-2.9
Aroclor-1016-4	4.960	4.861	5.061	484.640	500.000	-3.1
Aroclor-1016-5	5.217	5.118	5.318	505.120	500.000	1.0
Aroclor-1260-1	6.254	6.157	6.357	486.680	500.000	-2.7
Aroclor-1260-2	6.444	6.346	6.546	527.440	500.000	5.5
Aroclor-1260-3	6.811	6.713	6.913	531.610	500.000	6.3
Aroclor-1260-4	7.070	6.972	7.172	535.590	500.000	7.1
Aroclor-1260-5	7.313	7.216	7.416	516.860	500.000	3.4
Decachlorobiphenyl	8.708	8.612	8.812	45.020	50.000	-10.0
Tetrachloro-m-xylene	3.676	3.577	3.777	51.620	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/11/2025 06/11/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO111871.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.749	4.651	4.851	492.000	500.000	-1.6
Aroclor-1016-2	4.767	4.669	4.869	498.410	500.000	-0.3
Aroclor-1016-3	4.942	4.844	5.044	493.840	500.000	-1.2
Aroclor-1016-4	4.984	4.887	5.087	490.000	500.000	-2.0
Aroclor-1016-5	5.196	5.099	5.299	486.890	500.000	-2.6
Aroclor-1260-1	6.226	6.129	6.329	482.400	500.000	-3.5
Aroclor-1260-2	6.414	6.316	6.516	502.890	500.000	0.6
Aroclor-1260-3	6.566	6.469	6.669	475.680	500.000	-4.9
Aroclor-1260-4	7.036	6.939	7.139	469.160	500.000	-6.2
Aroclor-1260-5	7.277	7.181	7.381	474.960	500.000	-5.0
Decachlorobiphenyl	8.656	8.561	8.761	46.490	50.000	-7.0
Tetrachloro-m-xylene	3.672	3.573	3.773	48.950	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.73	5.73	5.63	5.83	0.01
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.11	6.12	6.02	6.22	0.01
Aroclor-1260-1	(1)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-2	(2)	7.49	7.49	7.39	7.59	0.00
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.00
Aroclor-1260-5	(5)	8.38	8.39	8.29	8.49	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 09:04 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.86	4.87	4.77	4.97	0.01
Aroclor-1016-2	(2)	4.88	4.89	4.79	4.99	0.01
Aroclor-1016-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.32	5.22	5.42	0.01
Aroclor-1260-1	(1)	6.34	6.35	6.25	6.45	0.01
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.68	6.69	6.59	6.79	0.01
Aroclor-1260-4	(4)	7.15	7.16	7.06	7.26	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.01
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.641	5.546	5.746	485.990	500.000	-2.8
Aroclor-1016-2	5.662	5.567	5.767	533.210	500.000	6.6
Aroclor-1016-3	5.725	5.629	5.829	520.180	500.000	4.0
Aroclor-1016-4	5.822	5.727	5.927	549.860	500.000	10.0
Aroclor-1016-5	6.114	6.019	6.219	524.760	500.000	5.0
Aroclor-1260-1	7.231	7.136	7.336	560.880	500.000	12.2
Aroclor-1260-2	7.485	7.389	7.589	543.950	500.000	8.8
Aroclor-1260-3	7.842	7.747	7.947	544.400	500.000	8.9
Aroclor-1260-4	8.066	7.971	8.171	548.260	500.000	9.7
Aroclor-1260-5	8.384	8.288	8.488	516.290	500.000	3.3
Decachlorobiphenyl	10.181	10.087	10.287	50.830	50.000	1.7
Tetrachloro-m-xylene	4.490	4.394	4.594	51.690	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073279.D Time Analyzed: 09:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.862	4.768	4.968	473.800	500.000	-5.2
Aroclor-1016-2	4.880	4.786	4.986	482.540	500.000	-3.5
Aroclor-1016-3	5.056	4.962	5.162	483.380	500.000	-3.3
Aroclor-1016-4	5.098	5.004	5.204	477.720	500.000	-4.5
Aroclor-1016-5	5.312	5.218	5.418	546.190	500.000	9.2
Aroclor-1260-1	6.342	6.249	6.449	495.490	500.000	-0.9
Aroclor-1260-2	6.531	6.438	6.638	513.470	500.000	2.7
Aroclor-1260-3	6.683	6.589	6.789	485.640	500.000	-2.9
Aroclor-1260-4	7.152	7.060	7.260	489.320	500.000	-2.1
Aroclor-1260-5	7.395	7.302	7.502	489.940	500.000	-2.0
Decachlorobiphenyl	8.788	8.697	8.897	50.670	50.000	1.3
Tetrachloro-m-xylene	3.783	3.688	3.888	47.840	50.000	-4.3

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 14:37 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.64	5.65	5.55	5.75	0.01
Aroclor-1016-2	(2)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-3	(3)	5.73	5.73	5.63	5.83	0.00
Aroclor-1016-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1016-5	(5)	6.12	6.12	6.02	6.22	0.00
Aroclor-1260-1	(1)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-2	(2)	7.49	7.49	7.39	7.59	0.00
Aroclor-1260-3	(3)	7.84	7.85	7.75	7.95	0.01
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.00
Aroclor-1260-5	(5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

Continuing Calib Date: 06/26/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 14:37 Initial Calibration Time(s): 10:04 19:37

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.87	4.87	4.77	4.97	0.01
Aroclor-1016-2	(2)	4.88	4.89	4.79	4.99	0.01
Aroclor-1016-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1016-4	(4)	5.10	5.10	5.00	5.20	0.00
Aroclor-1016-5	(5)	5.31	5.32	5.22	5.42	0.01
Aroclor-1260-1	(1)	6.34	6.35	6.25	6.45	0.01
Aroclor-1260-2	(2)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.69	6.69	6.59	6.79	0.01
Aroclor-1260-4	(4)	7.16	7.16	7.06	7.26	0.00
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.79	3.79	3.69	3.89	0.00
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.643	5.546	5.746	477.730	500.000	-4.5
Aroclor-1016-2	5.665	5.567	5.767	514.610	500.000	2.9
Aroclor-1016-3	5.727	5.629	5.829	507.950	500.000	1.6
Aroclor-1016-4	5.824	5.727	5.927	534.040	500.000	6.8
Aroclor-1016-5	6.116	6.019	6.219	509.130	500.000	1.8
Aroclor-1260-1	7.233	7.136	7.336	554.370	500.000	10.9
Aroclor-1260-2	7.486	7.389	7.589	530.980	500.000	6.2
Aroclor-1260-3	7.844	7.747	7.947	525.830	500.000	5.2
Aroclor-1260-4	8.068	7.971	8.171	527.090	500.000	5.4
Aroclor-1260-5	8.386	8.288	8.488	505.830	500.000	1.2
Decachlorobiphenyl	10.182	10.087	10.287	51.630	50.000	3.3
Tetrachloro-m-xylene	4.492	4.394	4.594	49.270	50.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/26/2025

Lab Sample No.: AR1660CCC500 Data File : PP073294.D Time Analyzed: 14:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.865	4.768	4.968	475.950	500.000	-4.8
Aroclor-1016-2	4.882	4.786	4.986	486.130	500.000	-2.8
Aroclor-1016-3	5.059	4.962	5.162	484.990	500.000	-3.0
Aroclor-1016-4	5.100	5.004	5.204	478.600	500.000	-4.3
Aroclor-1016-5	5.314	5.218	5.418	545.720	500.000	9.1
Aroclor-1260-1	6.344	6.249	6.449	518.030	500.000	3.6
Aroclor-1260-2	6.534	6.438	6.638	531.220	500.000	6.2
Aroclor-1260-3	6.685	6.589	6.789	499.440	500.000	-0.1
Aroclor-1260-4	7.155	7.060	7.260	498.560	500.000	-0.3
Aroclor-1260-5	7.396	7.302	7.502	503.650	500.000	0.7
Decachlorobiphenyl	8.790	8.697	8.897	52.420	50.000	4.8
Tetrachloro-m-xylene	3.786	3.688	3.888	49.210	50.000	-1.6

Analytical Sequence

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/11/2025	10:21	PO111586.D	8.71	3.68
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.71	3.68
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.71	3.68
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.71	3.68
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.71	3.68
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.71	3.68
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.71	3.68
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.71	3.68
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.71	3.68
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.71	3.68
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.71	3.68
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.71	3.68
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.71	3.68
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.71	3.68
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.71	3.68
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.71	3.68
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.71	3.68
AR1660CCC500	AR1660CCC500	06/26/2025	10:06	PO111858.D	8.71	3.68
IBLK	IBLK	06/26/2025	11:20	PO111862.D	8.71	3.68
COMP-1	Q2425-01	06/26/2025	13:12	PO111866.D	8.71	3.67
COMP-2	Q2425-02	06/26/2025	13:30	PO111867.D	8.71	3.68
COMP-3	Q2425-03	06/26/2025	13:48	PO111868.D	8.71	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	15:37	PO111871.D	8.71	3.68
IBLK	IBLK	06/26/2025	16:51	PO111875.D	8.71	3.68
IBLK	IBLK	06/17/2025	09:47	PP072990.D	10.19	4.49
AR1660ICC1000	AR1660ICC1000	06/17/2025	10:04	PP072991.D	10.19	4.49
AR1660ICC750	AR1660ICC750	06/17/2025	10:20	PP072992.D	10.19	4.49
AR1660ICC500	AR1660ICC500	06/17/2025	10:37	PP072993.D	10.19	4.49
AR1660ICC250	AR1660ICC250	06/17/2025	10:53	PP072994.D	10.19	4.49
AR1660ICC050	AR1660ICC050	06/17/2025	11:43	PP072995.D	10.19	4.50
AR1221ICC500	AR1221ICC500	06/17/2025	12:00	PP072996.D	10.18	4.49
AR1232ICC500	AR1232ICC500	06/17/2025	12:16	PP072997.D	10.19	4.50
AR1242ICC500	AR1242ICC500	06/17/2025	15:00	PP073000.D	10.19	4.49
AR1248ICC500	AR1248ICC500	06/17/2025	16:37	PP073005.D	10.19	4.49
AR1254ICC1000	AR1254ICC1000	06/17/2025	17:26	PP073008.D	10.19	4.49
AR1254ICC750	AR1254ICC750	06/17/2025	17:43	PP073009.D	10.19	4.49
AR1254ICC500	AR1254ICC500	06/17/2025	17:59	PP073010.D	10.19	4.49
AR1254ICC250	AR1254ICC250	06/17/2025	18:15	PP073011.D	10.19	4.49
AR1254ICC050	AR1254ICC050	06/17/2025	18:32	PP073012.D	10.19	4.49
AR1262ICC500	AR1262ICC500	06/17/2025	18:48	PP073013.D	10.19	4.49
AR1268ICC500	AR1268ICC500	06/17/2025	19:37	PP073016.D	10.18	4.49
AR1660CCC500	AR1660CCC500	06/26/2025	09:04	PP073279.D	10.18	4.49

Analytical Sequence

I.BLK	I.BLK	06/26/2025	10:09	PP073283.D	10.18	4.49
PB168622BL	PB168622BL	06/26/2025	12:25	PP073289.D	10.18	4.49
PB168622BS	PB168622BS	06/26/2025	12:42	PP073290.D	10.18	4.49
COP-SOIL-PILEMS	Q2409-02MS	06/26/2025	13:15	PP073292.D	10.18	4.49
COP-SOIL-PILEMSD	Q2409-02MSD	06/26/2025	13:31	PP073293.D	10.18	4.49
AR1660CCC500	AR1660CCC500	06/26/2025	14:37	PP073294.D	10.18	4.49
I.BLK	I.BLK	06/26/2025	15:42	PP073298.D	10.18	4.49

Analytical Sequence

Client: Kleinfelder

SDG No.: Q2425

Project: AS Jenks School

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 06/11/2025

06/11/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	06/11/2025	10:21	PO111586.D	8.66	3.67
AR1660ICC1000	AR1660ICC1000	06/11/2025	10:40	PO111587.D	8.66	3.67
AR1660ICC750	AR1660ICC750	06/11/2025	10:58	PO111588.D	8.66	3.67
AR1660ICC500	AR1660ICC500	06/11/2025	11:17	PO111589.D	8.66	3.67
AR1660ICC250	AR1660ICC250	06/11/2025	11:35	PO111590.D	8.66	3.67
AR1660ICC050	AR1660ICC050	06/11/2025	11:53	PO111591.D	8.66	3.67
AR1221ICC500	AR1221ICC500	06/11/2025	12:12	PO111592.D	8.66	3.67
AR1232ICC500	AR1232ICC500	06/11/2025	12:30	PO111593.D	8.66	3.67
AR1242ICC500	AR1242ICC500	06/11/2025	13:25	PO111596.D	8.66	3.67
AR1248ICC500	AR1248ICC500	06/11/2025	15:14	PO111601.D	8.66	3.67
AR1254ICC1000	AR1254ICC1000	06/11/2025	16:06	PO111604.D	8.66	3.67
AR1254ICC750	AR1254ICC750	06/11/2025	16:25	PO111605.D	8.66	3.67
AR1254ICC500	AR1254ICC500	06/11/2025	16:43	PO111606.D	8.66	3.67
AR1254ICC250	AR1254ICC250	06/11/2025	17:00	PO111607.D	8.66	3.67
AR1254ICC050	AR1254ICC050	06/11/2025	17:18	PO111608.D	8.66	3.67
AR1262ICC500	AR1262ICC500	06/11/2025	17:36	PO111609.D	8.66	3.67
AR1268ICC500	AR1268ICC500	06/11/2025	18:31	PO111612.D	8.66	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	10:06	PO111858.D	8.66	3.67
IBLK	IBLK	06/26/2025	11:20	PO111862.D	8.66	3.67
COMP-1	Q2425-01	06/26/2025	13:12	PO111866.D	8.66	3.67
COMP-2	Q2425-02	06/26/2025	13:30	PO111867.D	8.66	3.67
COMP-3	Q2425-03	06/26/2025	13:48	PO111868.D	8.66	3.67
AR1660CCC500	AR1660CCC500	06/26/2025	15:37	PO111871.D	8.66	3.67
IBLK	IBLK	06/26/2025	16:51	PO111875.D	8.66	3.67
IBLK	IBLK	06/17/2025	09:47	PP072990.D	8.80	3.79
AR1660ICC1000	AR1660ICC1000	06/17/2025	10:04	PP072991.D	8.80	3.79
AR1660ICC750	AR1660ICC750	06/17/2025	10:20	PP072992.D	8.80	3.79
AR1660ICC500	AR1660ICC500	06/17/2025	10:37	PP072993.D	8.80	3.79
AR1660ICC250	AR1660ICC250	06/17/2025	10:53	PP072994.D	8.80	3.79
AR1660ICC050	AR1660ICC050	06/17/2025	11:43	PP072995.D	8.80	3.79
AR1221ICC500	AR1221ICC500	06/17/2025	12:00	PP072996.D	8.80	3.79
AR1232ICC500	AR1232ICC500	06/17/2025	12:16	PP072997.D	8.80	3.79
AR1242ICC500	AR1242ICC500	06/17/2025	15:00	PP073000.D	8.80	3.79
AR1248ICC500	AR1248ICC500	06/17/2025	16:37	PP073005.D	8.80	3.79
AR1254ICC1000	AR1254ICC1000	06/17/2025	17:26	PP073008.D	8.80	3.79
AR1254ICC750	AR1254ICC750	06/17/2025	17:43	PP073009.D	8.80	3.79
AR1254ICC500	AR1254ICC500	06/17/2025	17:59	PP073010.D	8.80	3.79
AR1254ICC250	AR1254ICC250	06/17/2025	18:15	PP073011.D	8.80	3.79
AR1254ICC050	AR1254ICC050	06/17/2025	18:32	PP073012.D	8.80	3.79
AR1262ICC500	AR1262ICC500	06/17/2025	18:48	PP073013.D	8.80	3.79
AR1268ICC500	AR1268ICC500	06/17/2025	19:37	PP073016.D	8.80	3.79
AR1660CCC500	AR1660CCC500	06/26/2025	09:04	PP073279.D	8.79	3.78

Analytical Sequence

I.BLK	I.BLK	06/26/2025	10:09	PP073283.D	8.79	3.78
PB168622BL	PB168622BL	06/26/2025	12:25	PP073289.D	8.79	3.79
PB168622BS	PB168622BS	06/26/2025	12:42	PP073290.D	8.79	3.79
COP-SOIL-PILEMS	Q2409-02MS	06/26/2025	13:15	PP073292.D	8.79	3.79
COP-SOIL-PILEMSD	Q2409-02MSD	06/26/2025	13:31	PP073293.D	8.79	3.79
AR1660CCC500	AR1660CCC500	06/26/2025	14:37	PP073294.D	8.79	3.79
I.BLK	I.BLK	06/26/2025	15:42	PP073298.D	8.79	3.78



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168622BL	SDG No.:	Q2425
Lab Sample ID:	PB168622BL	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073289.D	1	06/26/25 08:20	06/26/25 12:25	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.3		32 - 144	96%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.1		32 - 175	106%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/11/25
Project:	AS Jenks School	Date Received:	06/11/25
Client Sample ID:	PIBLK-PO111586.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111586.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111586.D	1		06/11/25	po061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.0		60 - 140	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.5		60 - 140	103%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PO111862.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111862.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111862.D	1		06/26/25	po062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		60 - 140	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.4		60 - 140	97%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PO111875.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PO111875.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO111875.D	1		06/26/25	po062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140	90%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.2		60 - 140	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/17/25
Project:	AS Jenks School	Date Received:	06/17/25
Client Sample ID:	PIBLK-PP072990.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP072990.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP072990.D	1		06/17/25	pp061725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		60 - 140	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.2		60 - 140	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PP073283.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP073283.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073283.D	1		06/26/25	pp062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		60 - 140	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		60 - 140	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/26/25
Project:	AS Jenks School	Date Received:	06/26/25
Client Sample ID:	PIBLK-PP073298.D	SDG No.:	Q2425
Lab Sample ID:	I.BLK-PP073298.D	Matrix:	WATER
Analytical Method:	8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073298.D	1		06/26/25	pp062625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.9		60 - 140	95%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	AS Jenks School	Date Received:	
Client Sample ID:	PB168622BS	SDG No.:	Q2425
Lab Sample ID:	PB168622BS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	100
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073290.D	1	06/26/25 08:20	06/26/25 12:42	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	160		3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	159		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		32 - 175	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/24/25
Client Sample ID:	COP-SOIL-PILEMS	SDG No.:	Q2425
Lab Sample ID:	Q2409-02MS	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	96.4
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073292.D	1	06/26/25 08:20	06/26/25 13:15	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	163		4.10	17.6	ug/kg
11097-69-1	Aroclor-1254	135		3.30	17.6	ug/kg
11096-82-5	Aroclor-1260	195		3.30	17.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.0		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	22.7		32 - 175	113%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Kleinfelder	Date Collected:	06/24/25
Project:	AS Jenks School	Date Received:	06/24/25
Client Sample ID:	COP-SOIL-PILEMSD	SDG No.:	Q2425
Lab Sample ID:	Q2409-02MSD	Matrix:	SOIL
Analytical Method:	8082A	% Solid:	96.4
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP073293.D	1	06/26/25 08:20	06/26/25 13:31	PB168622

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	163		4.10	17.6	ug/kg
11097-69-1	Aroclor-1254	126		3.30	17.6	ug/kg
11096-82-5	Aroclor-1260	202		3.30	17.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.3		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.0		32 - 175	115%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit