

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.705	6.605	6.805	53.110	50.000	6.2
4,4'-DDE	6.196	6.096	6.296	52.180	50.000	4.4
4,4'-DDT	7.021	6.921	7.121	48.890	50.000	-2.2
Aldrin	5.271	5.171	5.371	54.570	50.000	9.1
Decachlorobiphenyl	9.072	8.972	9.172	49.280	50.000	-1.4
Dieldrin	6.347	6.247	6.447	53.280	50.000	6.6
Tetrachloro-m-xylene	3.551	3.450	3.650	53.870	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089159.D Time Analyzed: 11:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	51.260	50.000	2.5
4,4'-DDE	5.375	5.275	5.475	51.130	50.000	2.3
4,4'-DDT	6.183	6.084	6.284	48.440	50.000	-3.1
Aldrin	4.369	4.269	4.469	52.050	50.000	4.1
Decachlorobiphenyl	8.071	7.972	8.172	47.520	50.000	-5.0
Dieldrin	5.512	5.413	5.613	51.120	50.000	2.2
Tetrachloro-m-xylene	2.881	2.780	2.980	52.560	50.000	5.1