

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.605	6.805	52.210	50.000	4.4
4,4'-DDE	6.200	6.096	6.296	53.450	50.000	6.9
4,4'-DDT	7.025	6.921	7.121	46.980	50.000	-6.0
Aldrin	5.275	5.171	5.371	55.770	50.000	11.5
Decachlorobiphenyl	9.077	8.972	9.172	49.220	50.000	-1.6
Dieldrin	6.351	6.247	6.447	53.140	50.000	6.3
Tetrachloro-m-xylene	3.555	3.450	3.650	57.010	50.000	14.0

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/26/2025

Lab Sample No.: PSTDCCC050 Data File : PD089172.D Time Analyzed: 17:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	49.850	50.000	-0.3
4,4'-DDE	5.376	5.275	5.475	51.580	50.000	3.2
4,4'-DDT	6.184	6.084	6.284	46.860	50.000	-6.3
Aldrin	4.369	4.269	4.469	53.840	50.000	7.7
Decachlorobiphenyl	8.073	7.972	8.172	44.560	50.000	-10.9
Dieldrin	5.513	5.413	5.613	51.790	50.000	3.6
Tetrachloro-m-xylene	2.881	2.780	2.980	56.090	50.000	12.2