

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.030	50.000	14.1
4,4'-DDE	6.195	6.096	6.296	55.130	50.000	10.3
4,4'-DDT	7.020	6.921	7.121	43.230	50.000	-13.5
Aldrin	5.270	5.171	5.371	57.560	50.000	15.1
Decachlorobiphenyl	9.071	8.972	9.172	50.650	50.000	1.3
Dieldrin	6.346	6.247	6.447	56.130	50.000	12.3
Tetrachloro-m-xylene	3.550	3.450	3.650	58.070	50.000	16.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089192.D Time Analyzed: 18:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	53.960	50.000	7.9
4,4'-DDE	5.374	5.275	5.475	54.320	50.000	8.6
4,4'-DDT	6.182	6.084	6.284	43.220	50.000	-13.6
Aldrin	4.367	4.269	4.469	55.730	50.000	11.5
Decachlorobiphenyl	8.070	7.972	8.172	50.160	50.000	0.3
Dieldrin	5.511	5.413	5.613	54.310	50.000	8.6
Tetrachloro-m-xylene	2.881	2.780	2.980	58.440	50.000	16.9