

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.920	50.000	15.8
4,4'-DDE	6.194	6.096	6.296	54.700	50.000	9.4
4,4'-DDT	7.019	6.921	7.121	42.000	50.000	-16.0
Aldrin	5.270	5.171	5.371	56.890	50.000	13.8
Decachlorobiphenyl	9.070	8.972	9.172	50.550	50.000	1.1
Dieldrin	6.345	6.247	6.447	55.790	50.000	11.6
Tetrachloro-m-xylene	3.550	3.450	3.650	57.800	50.000	15.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	54.200	50.000	8.4
4,4'-DDE	5.373	5.275	5.475	53.740	50.000	7.5
4,4'-DDT	6.181	6.084	6.284	41.310	50.000	-17.4
Aldrin	4.367	4.269	4.469	55.520	50.000	11.0
Decachlorobiphenyl	8.070	7.972	8.172	50.420	50.000	0.8
Dieldrin	5.511	5.413	5.613	53.700	50.000	7.4
Tetrachloro-m-xylene	2.881	2.780	2.980	58.420	50.000	16.8