

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2425</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Kleinfelder</u>	Datafile :	<u>PD089167.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168623BS (Column 1)	Aldrin	16.65	16.2	ug/kg	97				82	124		
	Dieldrin	16.65	15.9	ug/kg	95				85	121		
	4,4'-DDE	16.65	15.9	ug/kg	95				81	123		
	4,4'-DDD	16.65	16.1	ug/kg	97				80	131		
	4,4'-DDT	16.65	13.9	ug/kg	83				70	129		
PB168623BS (Column 2)	Aldrin	16.65	15.9	ug/kg	95				82	124		
	Dieldrin	16.65	15.5	ug/kg	93				85	121		
	4,4'-DDE	16.65	15.5	ug/kg	93				81	123		
	4,4'-DDD	16.65	16.0	ug/kg	96				80	131		
	4,4'-DDT	16.65	13.9	ug/kg	83				70	129		