

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061925

Review By	Rahul	Review On	6/23/2025 3:55:52 PM
Supervise By	Jagrut	Supervise On	6/24/2025 3:17:49 PM
SubDirectory	BF061925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12670,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142787.D	19 Jun 2025 16:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142788.D	19 Jun 2025 17:07		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142789.D	19 Jun 2025 17:38		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF142790.D	19 Jun 2025 18:08		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142791.D	19 Jun 2025 18:39		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF142792.D	19 Jun 2025 19:09		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142793.D	19 Jun 2025 19:40		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142794.D	19 Jun 2025 20:10		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142795.D	19 Jun 2025 20:40	Compound #09 removed from 80 PPM.	RC/JU	Ok
10	SSTDICV040	ICVBF061925	BF142796.D	19 Jun 2025 21:10		RC/JU	Ok
11	PB168536BL	PB168536BL	BF142797.D	19 Jun 2025 22:10		RC/JU	Ok
12	PB168536BS	PB168536BS	BF142798.D	19 Jun 2025 22:40		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062625

Review By	Rahul	Review On	6/27/2025 11:23:56 AM
Supervise By	Jagrut	Supervise On	6/27/2025 11:38:19 AM
SubDirectory	BF062625	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12671,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142856.D	26 Jun 2025 09:10		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142857.D	26 Jun 2025 09:39		RC/JU	Ok
3	PB168610BL	PB168610BL	BF142858.D	26 Jun 2025 10:08		RC/JU	Ok
4	PB168610BS	PB168610BS	BF142859.D	26 Jun 2025 10:38		RC/JU	Ok,M
5	PB168614BL	PB168614BL	BF142860.D	26 Jun 2025 11:07		RC/JU	Ok
6	PB168614BS	PB168614BS	BF142861.D	26 Jun 2025 11:36		RC/JU	Ok,M
7	Q2403-05	Concrete-062325	BF142862.D	26 Jun 2025 12:09		RC/JU	Ok
8	Q2405-01	MH-M/H	BF142863.D	26 Jun 2025 12:39		RC/JU	Ok
9	Q2405-01MS	MH-M/HMS	BF142864.D	26 Jun 2025 13:08		RC/JU	Ok,M
10	Q2405-01MSD	MH-M/HMSD	BF142865.D	26 Jun 2025 13:38		RC/JU	Ok,M
11	Q2394-04	MH-K/L	BF142866.D	26 Jun 2025 14:07		RC/JU	Ok
12	Q2399-04	TP-13	BF142867.D	26 Jun 2025 14:37		RC/JU	Ok
13	Q2399-08	EP-7	BF142868.D	26 Jun 2025 15:07		RC/JU	Ok
14	Q2405-04	MH-M/N	BF142869.D	26 Jun 2025 15:37		RC/JU	Ok
15	Q2409-02	COP-SOIL-PILE	BF142870.D	26 Jun 2025 16:06		RC/JU	Ok
16	Q2416-01	MH-G/H	BF142871.D	26 Jun 2025 16:36		RC/JU	Ok
17	Q2416-01MS	MH-G/HMS	BF142872.D	26 Jun 2025 17:06		RC/JU	Ok,M
18	Q2416-01MSD	MH-G/HMSD	BF142873.D	26 Jun 2025 17:36		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062625

Review By	Rahul	Review On	6/27/2025 11:23:56 AM		
Supervise By	Jagrut	Supervise On	6/27/2025 11:38:19 AM		
SubDirectory	BF062625	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12671,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2425-01	COMP-1	BF142874.D	26 Jun 2025 18:05		RC/JU	Ok
20	Q2414-01	WC-1	BF142875.D	26 Jun 2025 18:35		RC/JU	Ok,M
21	Q2403-07	ARS 20-0006	BF142876.D	26 Jun 2025 19:04		RC/JU	Ok,M
22	Q2420-01	72-11933	BF142877.D	26 Jun 2025 19:34	Need 10X Dilution then decide further dilution	RC/JU	Dilution
23	Q2415-01	WC-1	BF142878.D	26 Jun 2025 20:03		RC/JU	Ok,M
24	Q2423-01	RT915	BF142879.D	26 Jun 2025 20:32	Internal standard fail	RC/JU	Ok,M
25	Q2419-01	EO-3-6-25-2025	BF142880.D	26 Jun 2025 21:02	Internal standard fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062725

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF062725	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12671,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142881.D	27 Jun 2025 09:21		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142882.D	27 Jun 2025 09:51		RC/JU	Ok
3	PB168571TB	PB168571TB	BF142883.D	27 Jun 2025 10:20		RC/JU	Ok
4	PB168625BL	PB168625BL	BF142884.D	27 Jun 2025 10:49		RC/JU	Ok
5	PB168625BS	PB168625BS	BF142885.D	27 Jun 2025 11:19		RC/JU	Ok,NR
6	Q2413-05	TP-73	BF142886.D	27 Jun 2025 11:52		RC/JU	Ok
7	Q2425-02	COMP-2	BF142887.D	27 Jun 2025 12:21		RC/JU	Ok
8	Q2425-03	COMP-3	BF142888.D	27 Jun 2025 12:50		RC/JU	Ok
9	Q2413-01	TP-35	BF142889.D	27 Jun 2025 13:19		RC/JU	Ok,NR
10	Q2413-03	TP-77	BF142890.D	27 Jun 2025 13:48		RC/JU	Ok
11	Q2413-04	TP-74	BF142891.D	27 Jun 2025 14:17		RC/JU	Ok,NR
12	Q2413-02	TP-34	BF142892.D	27 Jun 2025 14:47		RC/JU	Ok,NR
13	Q2413-06	TP-72	BF142893.D	27 Jun 2025 15:15		RC/JU	Ok
14	Q2420-01DL	72-11933DL	BF142894.D	27 Jun 2025 15:44	Need 5X Further Dilution	RC/JU	Dilution
15	Q2420-01DL2	72-11933DL2	BF142895.D	27 Jun 2025 16:13		RC/JU	Ok,NR
16	Q2429-01	TP-4	BF142896.D	27 Jun 2025 16:43		RC/JU	Ok
17	Q2414-04	WC-1	BF142897.D	27 Jun 2025 17:12		RC/JU	Ok
18	Q2414-04MS	WC-1MS	BF142898.D	27 Jun 2025 17:41		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF062725

Review By	Review On
Supervise By	Supervise On
SubDirectory	BF062725
HP Acquire Method	BNA_F
HP Processing Method	bf061925
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12671,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	Q2414-04MSD	WC-1MSD	BF142899.D	27 Jun 2025 18:10		RC/JU	Ok,NR
20	Q2416-04	MH-G/H	BF142900.D	27 Jun 2025 18:40		RC/JU	Ok
21	Q2420-02	72-11933	BF142901.D	27 Jun 2025 19:10		RC/JU	Ok
22	Q2415-04	WC-1	BF142902.D	27 Jun 2025 19:39		RC/JU	Ok
23	Q2427-01	RT2929	BF142903.D	27 Jun 2025 20:08		RC/JU	Ok
24	Q2355-07	20071	BF142904.D	27 Jun 2025 20:37	Internal standard fail & Need 2X Dilution	RC/JU	Dilution
25	Q2418-01	VAC-TRUCK-4074	BF142905.D	27 Jun 2025 21:07	Internal standard fail	RC/JU	ReRun

M : Manual Integration