

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG No.: Q2425

Instrument ID: MSVOA_Y Calibration Date/Time: 06/25/2025 09:13

Lab File ID: VY022811.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.672	0.648		-3.57	20
1,1,1-Trichloroethane	0.929	0.916		-1.4	20
Benzene	1.417	1.415		-0.14	20
Trichloroethene	0.356	0.364		2.25	20
Toluene	0.894	0.909		1.68	20
Ethyl Benzene	1.930	1.981		2.64	20
m/p-Xylenes	0.746	0.775		3.89	20
o-Xylene	0.703	0.725		3.13	20
Isopropylbenzene	3.698	3.805		2.89	20
1,2-Dichloroethane-d4	0.558	0.536		-3.94	20
Dibromofluoromethane	0.304	0.309		1.64	20
Toluene-d8	1.207	1.256		4.06	20
4-Bromofluorobenzene	0.388	0.398		2.58	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.