

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062325

Review By	Mahesh Dadoda	Review On	6/24/2025 8:24:26 AM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/24/2025 8:29:32 AM		
SubDirectory	VY062325	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk	VP134461
Initial Calibration Stds	VP134462,VP134463,VP134464,VP134465,VP134466,VP134467
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP134468
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022775.D	23 Jun 2025 10:17		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY022776.D	23 Jun 2025 13:38	LR- 58	SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY022777.D	23 Jun 2025 14:00		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY022778.D	23 Jun 2025 14:23		SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY022779.D	23 Jun 2025 14:46		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY022780.D	23 Jun 2025 15:08		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY022781.D	23 Jun 2025 15:31		SY/MD	Ok,M
8	VIBLK	VIBLK	VY022782.D	23 Jun 2025 15:54		SY/MD	Ok
9	VSTDICV050	ICVVY062325	VY022783.D	23 Jun 2025 16:17		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062525

Review By	Maresh Dadoda	Review On	6/26/2025 2:39:39 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/26/2025 2:42:28 PM
SubDirectory	VY062525	HP Acquire Method	MSVOA_Y HP Processing Method 82y062325s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134515
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134516,VP134517 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022810.D	25 Jun 2025 08:41		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022811.D	25 Jun 2025 09:13		SY/MD	Ok,M
3	VY0625SBL01	VY0625SBL01	VY022812.D	25 Jun 2025 09:46		SY/MD	Ok
4	VY0625SBS01	VY0625SBS01	VY022813.D	25 Jun 2025 10:18		SY/MD	Ok,M
5	VY0625SBSD01	VY0625SBSD01	VY022814.D	25 Jun 2025 10:40		SY/MD	Ok,M
6	Q2403-03RE	LAW-25-0093RE	VY022815.D	25 Jun 2025 11:24	vial-B Internal Standard Fail	SY/MD	Confirms
7	Q2398-06	ARS20-0036	VY022816.D	25 Jun 2025 11:48	vial-B Need MeOH	SY/MD	Dilution
8	Q2386-01	SU-1-062025	VY022817.D	25 Jun 2025 12:11	vial-B	SY/MD	Ok
9	Q2355-03	20071	VY022818.D	25 Jun 2025 12:35	vial-B Internal and surrogate fail	SY/MD	Ok
10	Q2413-01	TP-35	VY022819.D	25 Jun 2025 12:58	vial-A Surrogate Fail	SY/MD	ReRun
11	Q2413-02	TP-34	VY022820.D	25 Jun 2025 13:22	vial-A	SY/MD	Ok
12	Q2413-03	TP-77	VY022821.D	25 Jun 2025 13:45	vial-A	SY/MD	Ok
13	Q2413-04	TP-74	VY022822.D	25 Jun 2025 14:08	vial-A	SY/MD	Ok
14	Q2413-05	TP-73	VY022823.D	25 Jun 2025 14:32	vial-A	SY/MD	Ok
15	Q2413-06	TP-72	VY022824.D	25 Jun 2025 14:55	vial-A Internal Standard Fail	SY/MD	ReRun
16	Q2405-03	MH-M/N-VOC	VY022825.D	25 Jun 2025 15:19	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q2412-01	TR-05-062425	VY022826.D	25 Jun 2025 15:42	vial-A	SY/MD	Ok
18	Q2414-03	WC-1-VOC	VY022827.D	25 Jun 2025 16:05	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY062525

Review By	Mahesh Dadoda	Review On	6/26/2025 2:39:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	6/26/2025 2:42:28 PM		
SubDirectory	VY062525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y062325s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134515			
CCC		VP134516,VP134517			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2415-03	WC-1-VOC	VY022828.D	25 Jun 2025 16:29	vial-A	SY/MD	Ok
20	Q2416-03	MH-G/H-VOC	VY022829.D	25 Jun 2025 16:52	vial-A	SY/MD	Ok
21	Q2420-01	72-11933	VY022830.D	25 Jun 2025 17:16	vial-A	SY/MD	Ok
22	Q2425-01	COMP-1	VY022831.D	25 Jun 2025 17:39	vial-A	SY/MD	Ok
23	Q2425-02	COMP-2	VY022832.D	25 Jun 2025 18:02	vial-A	SY/MD	Ok
24	Q2425-03	COMP-3	VY022833.D	25 Jun 2025 18:26	vial-A	SY/MD	Ok
25	Q2423-01	RT915	VY022834.D	25 Jun 2025 18:50	vial-A	SY/MD	Ok
26	VIBLK	VIBLK	VY022835.D	25 Jun 2025 19:13		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY022836.D	25 Jun 2025 19:59		SY/MD	Ok,M

M : Manual Integration