

Surrogate Summary

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2425-01	COMP-1	1,2-Dichloroethane-d4	50	48.5	97	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	50.0	100	74	123
		4-Bromofluorobenzene	50	50.2	100	17	146
Q2425-02	COMP-2	1,2-Dichloroethane-d4	50	46.2	92	63	155
		Dibromofluoromethane	50	49.4	99	70	134
		Toluene-d8	50	49.6	99	74	123
		4-Bromofluorobenzene	50	50.2	100	17	146
Q2425-03	COMP-3	1,2-Dichloroethane-d4	50	45.7	91	63	155
		Dibromofluoromethane	50	49.5	99	70	134
		Toluene-d8	50	49.8	100	74	123
		4-Bromofluorobenzene	50	51.5	103	17	146
VY0625SBL01	VY0625SBL01	1,2-Dichloroethane-d4	50	48.2	96	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	49.3	99	74	123
		4-Bromofluorobenzene	50	53.3	107	17	146
VY0625SBS01	VY0625SBS01	1,2-Dichloroethane-d4	50	50.1	100	63	155
		Dibromofluoromethane	50	50.1	100	70	134
		Toluene-d8	50	51.1	102	74	123
		4-Bromofluorobenzene	50	49.7	99	17	146
VY0625SBSD01	VY0625SBSD01	1,2-Dichloroethane-d4	50	51.0	102	63	155
		Dibromofluoromethane	50	51.1	102	70	134
		Toluene-d8	50	51.5	103	74	123
		4-Bromofluorobenzene	50	49.5	99	17	146