

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2425

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY022814.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0625SBSD01	cis-1,2-Dichloroethene	20	20.4	ug/Kg	102	1		82	123	20
	1,1,1-Trichloroethane	20	20.5	ug/Kg	103	2		80	126	20
	Benzene	20	20.4	ug/Kg	102	1		84	121	20
	Trichloroethene	20	20.4	ug/Kg	102	2		83	122	20
	Toluene	20	20.0	ug/Kg	100	1		83	122	20
	Ethyl Benzene	20	20.6	ug/Kg	103	2		82	124	20
	m/p-Xylenes	40	41.0	ug/Kg	103	3		83	124	20
	o-Xylene	20	20.2	ug/Kg	101	2		83	123	20
	Isopropylbenzene	20	21.3	ug/Kg	106	4		82	124	20