

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: Q2425 SAS No.: Q2425 SDG NO.: Q2425
 Lab File ID: VY022811.D Date Analyzed: 06/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	316858	13.34				
UPPER LIMIT	633716	13.84				
LOWER LIMIT	158429	12.84				
EPA SAMPLE NO.						
COMP-1	315122	13.34				
COMP-2	306171	13.34				
COMP-3	284845	13.34				
VY0625SBL01	248657	13.35				
VY0625SBS01	303331	13.35				
VY0625SBSD01	290027	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.