



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	95 - 105	P	06/25/2025	13:08	LB136273

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	92.0	100	92	80 - 120	P	06/25/2025	13:20	LB136273

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5120	5000	102	90 - 110	P	06/25/2025	16:04	LB136273
CCV02	Iron	5120	5000	102	90 - 110	P	06/25/2025	16:57	LB136273
CCV03	Iron	5130	5000	103	90 - 110	P	06/25/2025	18:00	LB136273
CCV04	Iron	5150	5000	103	90 - 110	P	06/25/2025	18:55	LB136273
CCV05	Iron	5120	5000	102	90 - 110	P	06/25/2025	19:52	LB136273
CCV06	Iron	5040	5000	101	90 - 110	P	06/25/2025	20:46	LB136273
CCV07	Iron	5050	5000	101	90 - 110	P	06/25/2025	21:12	LB136273

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3810	4000	95	95 - 105	P	06/26/2025	17:07	LB136305



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	112	100	112	80 - 120	P	06/26/2025	17:19	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4790	5000	96	90 - 110	P	06/26/2025	17:55	LB136305
CCV02	Iron	4840	5000	97	90 - 110	P	06/26/2025	18:48	LB136305
CCV03	Iron	4860	5000	97	90 - 110	P	06/26/2025	19:27	LB136305
CCV04	Iron	4780	5000	96	90 - 110	P	06/26/2025	20:02	LB136305
CCV05	Iron	4810	5000	96	90 - 110	P	06/26/2025	20:46	LB136305

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	95 - 105	P	07/03/2025	14:50	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	98.7	100	99	80 - 120	P	07/03/2025	14:55	LB136377

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5160	5000	103	90 - 110	P	07/03/2025	16:06	LB136377
CCV02	Iron	5190	5000	104	90 - 110	P	07/03/2025	16:53	LB136377
CCV03	Iron	5230	5000	105	90 - 110	P	07/03/2025	17:28	LB136377