



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q2426
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2426 **SAS No.:** Q2426
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	107	100	107	65 - 135	P	06/25/2025	15:24	LB136273
CRI01	Iron	105	100	105	65 - 135	P	06/26/2025	17:28	LB136305
CRI01	Iron	112	100	112	65 - 135	P	07/03/2025	15:37	LB136377