

Continuing Calibration Report for SequenceID : FC063025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069326.D

Aliphatic C9-C12	9162395.000	60.000	3.308	6.605	152706.583	146679.627	-4.109
Aliphatic C12-C16	6226328.000	40.000	6.606	10.009	155658.200	159210.532	2.231
Aliphatic C16-C21	8817332.000	60.000	10.010	13.379	146955.533	154434.323	4.843
Aliphatic C21-C28	10431684.000	80.000	13.380	17.046	130396.050	133966.368	2.665
Aliphatic C28-C40	11562204.000	120.000	17.047	22.025	96351.700	92124.417	-4.589
Aliphatic EPH	46199943.000	360.000	3.308	22.025	128333.175	128354.160	0.016