

Continuing Calibration Report for SequenceID : FC063025AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069330.D

Aliphatic C9-C12	9081249.000	60.000	3.308	6.605	151354.150	146679.627	-3.187
Aliphatic C12-C16	6251541.000	40.000	6.606	10.009	156288.525	159210.532	1.835
Aliphatic C16-C21	8912187.000	60.000	10.010	13.379	148536.450	154434.323	3.819
Aliphatic C21-C28	10573737.000	80.000	13.380	17.046	132171.713	133966.368	1.340
Aliphatic C28-C40	11724019.000	120.000	17.047	22.025	97700.158	92124.417	-6.052
Aliphatic EPH	46542733.000	360.000	3.308	22.025	129285.369	128354.160	-0.725