

Continuing Calibration Report for SequenceID : FE062725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE054615.D

Aliphatic C9-C12	7567174.000	60.000	3.317	6.943	126119.567	136055.079	7.303
Aliphatic C12-C16	5235301.000	40.000	6.944	10.391	130882.525	140568.128	6.890
Aliphatic C16-C21	8076702.000	60.000	10.392	13.764	134611.700	144393.643	6.774
Aliphatic C21-C28	10798798.000	80.000	13.765	17.429	134984.975	144651.717	6.683
Aliphatic C28-C40	17217475.000	120.000	17.430	22.428	143478.958	138647.492	-3.485
Aliphatic EPH	48895450.000	360.000	3.317	22.428	135820.694	140720.791	3.482