

Continuing Calibration Report for SequenceID : FG062725AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016154.D

Aliphatic C9-C12	6416956.000	60.000	3.316	6.955	106949.267	109333.554	2.181
Aliphatic C12-C16	4427737.000	40.000	6.956	10.413	110693.425	111824.611	1.012
Aliphatic C16-C21	6800919.000	60.000	10.414	13.802	113348.650	114479.647	0.988
Aliphatic C21-C28	9051271.000	80.000	13.803	17.485	113140.888	114390.192	1.092
Aliphatic C28-C40	11162303.000	120.000	17.486	22.536	93019.192	91995.767	-1.112
Aliphatic EPH	37859186.000	360.000	3.316	22.536	105164.406	105812.455	0.612