

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC06302
Data File : FC069340.D
Signal(s) : FID1A.ch
Acq On : 30 Jun 2025 23:40
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 07/01/2025
Supervised By :mohammad ahmed 07/01/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 061825.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.406	3.370	3.447	BB	366123	2991297	88.56%	5.028%
2	4.479	4.447	4.532	BB	348971	3079667	91.17%	5.177%
3	6.074	6.035	6.147	BB	368302	3377869	100.00%	5.678%
4	6.504	6.455	6.540	BB	323131	3116957	92.28%	5.239%
5	7.134	7.085	7.227	BB	325247	3267917	96.74%	5.493%
6	8.304	8.269	8.375	BB	301103	3100898	91.80%	5.212%
7	9.909	9.869	9.979	BB	285845	3172818	93.93%	5.333%
8	11.353	11.317	11.405	BB	279746	3141133	92.99%	5.280%
9	11.676	11.637	11.722	BB	310044	3328551	98.54%	5.595%
10	12.665	12.625	12.714	BB	255660	2956239	87.52%	4.969%
11	13.111	13.070	13.159	BB	217965	2529589	74.89%	4.252%
12	13.278	13.239	13.332	BB	233133	2874850	85.11%	4.832%
13	13.865	13.792	13.904	BB	242867	2831081	83.81%	4.759%
14	14.971	14.932	15.017	BB	218238	2732822	80.90%	4.594%
15	15.995	15.952	16.049	BB	201458	2607679	77.20%	4.383%
16	16.944	16.859	16.992	BB	179020	2490965	73.74%	4.187%
17	17.834	17.780	17.882	BB	177658	2469385	73.10%	4.151%
18	18.668	18.595	18.727	BB	167568	2339744	69.27%	3.933%
19	19.453	19.400	19.494	BB	151397	2100828	62.19%	3.531%
20	20.194	20.140	20.230	BB	125958	1806356	53.48%	3.036%
21	20.962	20.882	21.030	BB	85737	1650385	48.86%	2.774%
22	21.930	21.880	22.017	BB	63337	1525495	45.16%	2.564%
Sum of corrected areas:						59492524		

Aliphatic EPH 061825.M Tue Jul 01 05:52:41 2025