

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q2432 SAS No.: Q2432 SDG No.: Q2432

Instrument ID: MSVOA_L Calibration Date/Time: 06/27/2025 08:23

Lab File ID: VL042670.D Init. Calib. Date(s): 05/29/2025 05/29/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:24

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.496	0.462		-6.86	30
1,1-Dichloroethane	0.957	1.056		10.35	30
cis-1,2-Dichloroethene	1.370	1.375		0.37	30
1,1,1-Trichloroethane	1.990	2.034		2.21	30
Trichloroethene	0.419	0.441		5.25	30
Tetrachloroethene	0.368	0.373		1.36	30
1-Bromo-4-Fluorobenzene	0.750	0.792		5.6	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.