

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
 Data File : VL042583.D
 Acq On : 29 May 2025 13:19
 Operator : SY/MD
 Sample : VSTDICC0.1
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/30/2025
 Supervised By :Mahesh Dadoda 05/30/2025

Quant Time: May 30 01:35:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2025
 QLast Update : Fri May 30 01:15:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.790	49	103370	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.962	114	302218	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	273881	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.380	95	205672	10.018	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.200%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	1.580	62	481	0.094	ppbv #	31
32) 1,1,1-Trichloroethane	3.373	97	2153	0.107	ppbv	90
35) Carbon Tetrachloride	3.768	117	2135	0.107	ppbv #	76
38) Trichloroethene	4.548	130	1305m	0.102	ppbv	
52) Tetrachloroethene	8.238	164	1230m	0.111	ppbv	
54) 1,2-Dibromoethane	7.658	107	1737m	0.103	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.969	83	2267	0.098	ppbv	96
79) Naphthalene	13.523	128	2980m	0.076	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL052925\
Data File : VL042583.D
Acq On : 29 May 2025 13:19
Operator : SY/MD
Sample : VSTDICC0.1
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.1

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/30/2025
Supervised By :Mahesh Dadoda 05/30/2025

Quant Time: May 30 01:35:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL052925AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri May 30 01:15:52 2025
Response via : Initial Calibration

