

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL052925

Review By	Semsettin Yesilyurt	Review On	5/30/2025 7:50:21 AM
Supervise By	Maresh Dadoda	Supervise On	5/30/2025 12:26:19 PM
SubDirectory	VL052925	HP Acquire Method	TO15AIR.M
		HP Processing Method	VL052925AIR.M
STD. NAME	STD REF.#		
Tune/Reschk	AP2629		
Initial Calibration Stds	AP2623,AP2625,AP2626		
CCC	AP2623		
Internal Standard/PEM	AP2629		
ICV/I.BLK	AP2627		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042578.D	29 May 2025 08:05		SY/MD	Ok
2	VSTDICCC010	VSTDICCC010	VL042579.D	29 May 2025 11:09		SY/MD	Ok,M
3	VSTDICC002	VSTDICC002	VL042580.D	29 May 2025 11:43		SY/MD	Ok,M
4	VSTDICC001	VSTDICC001	VL042581.D	29 May 2025 12:15		SY/MD	Ok,M
5	VSTDICC0.5	VSTDICC0.5	VL042582.D	29 May 2025 12:47		SY/MD	Ok,M
6	VSTDICC0.1	VSTDICC0.1	VL042583.D	29 May 2025 13:19		SY/MD	Ok,M
7	VSTDICC0.03	VSTDICC0.03	VL042584.D	29 May 2025 13:51		SY/MD	Ok,M
8	VSTDICC015	VSTDICC015	VL042585.D	29 May 2025 14:24		SY/MD	Ok,M
9	VSTDICV010	ICVVL052925	VL042586.D	29 May 2025 14:56	comp#13 fail	SY/MD	Ok,M
10	VL0529ABL01	VL0529ABL01	VL042587.D	29 May 2025 15:44	carryover	SY/MD	Not Ok
11	VL0529ABL02	VL0529ABL02	VL042588.D	29 May 2025 16:50		SY/MD	Ok
12	Q2124-01	SV1	VL042589.D	29 May 2025 17:22	Need 20X	SY/MD	Dilution
13	Q2124-01DL	SV1DL	VL042590.D	29 May 2025 17:55		SY/MD	Ok,M
14	VL0529ABS01	VL0529ABS01	VL042591.D	29 May 2025 18:27		SY/MD	Ok,M
15	Q2124-02	SV2	VL042592.D	29 May 2025 19:00		SY/MD	Ok,M
16	Q2124-02DL	SV2DL	VL042593.D	29 May 2025 19:32	Not Required	SY/MD	Not Ok
17	Q2124-03	SV3	VL042594.D	29 May 2025 20:05		SY/MD	Ok,M
18	Q2124-03DL	SV3DL	VL042595.D	29 May 2025 20:37	Not Required	SY/MD	Not Ok

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL052925

Review By	Semsettin Yesilyurt	Review On	5/30/2025 7:50:21 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/30/2025 12:26:19 PM		
SubDirectory	VL052925	HP Acquire Method	TO15AIR.M	HP Processing Method	VL052925AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2629			
Initial Calibration Stds		AP2623,AP2625,AP2626			
CCC		AP2623			
Internal Standard/PEM		AP2629			
ICV/I.BLK		AP2627			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2124-01DUP	SV1DUP	VL042596.D	29 May 2025 21:10	wrong purge	SY/MD	Not Ok
20	QC052725 CAN 10609	QC052725 CAN 10609	VL042597.D	29 May 2025 21:44	Contaminated	SY/MD	Not Ok
21	QC052725 CAN 10609	QC052725 CAN 10609	VL042598.D	29 May 2025 22:18		SY/MD	Ok
22	CAN10053	CAN10053	VL042599.D	29 May 2025 22:52		SY/MD	Ok
23	QC05225 CAN 10154	QC05225 CAN 10154	VL042600.D	29 May 2025 23:26		SY/MD	Ok
24	Q2124-01DUP	SV1DUP	VL042601.D	29 May 2025 23:58		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_L

Daily Analysis Runlog For Sequence/QC Batch ID # VL062725

Review By	Maresh Dadoda	Review On	6/30/2025 8:21:10 AM
Supervise By	Semsettin Yesilyurt	Supervise On	6/30/2025 8:23:55 AM
SubDirectory	VL062725	HP Acquire Method	TO15AIR.M HP Processing Method VL052925AIR.M

STD. NAME	STD REF.#
Tune/Reschk	AP2629
Initial Calibration Stds	AP2623,AP2625,AP2626
CCC	AP2623
Internal Standard/PEM	AP2629
ICV/I.BLK	AP2627
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VL042669.D	27 Jun 2025 07:40		SY/MD	Ok
2	VSTDCCC010	VSTDCCC010	VL042670.D	27 Jun 2025 08:23		SY/MD	Ok,M
3	VL0627ABL01	VL0627ABL01	VL042671.D	27 Jun 2025 09:11		SY/MD	Ok
4	VL0627ABS01	VL0627ABS01	VL042672.D	27 Jun 2025 10:01		SY/MD	Ok,M
5	Q2432-01	SV1	VL042673.D	27 Jun 2025 10:43	Run @ Lower Dilution	SY/MD	Not Ok
6	Q2432-01	SV1	VL042674.D	27 Jun 2025 11:30	Need 80X	SY/MD	Dilution
7	Q2432-01DUP	SV1DUP	VL042675.D	27 Jun 2025 12:02		SY/MD	Ok,M
8	Q2432-01DL	SV1DL	VL042676.D	27 Jun 2025 12:57	Not used,bad purge.	SY/MD	Not Ok
9	Q2432-01DL	SV1DL	VL042677.D	27 Jun 2025 14:01	Not used	SY/MD	Not Ok
10	Q2432-01DL	SV1DL	VL042678.D	27 Jun 2025 14:36		SY/MD	Ok,M

M : Manual Integration