

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2432

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042672.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0627ABS01	Vinyl Chloride	10	9.50	ppbv	95			70	130	
	1,1-Dichloroethane	10	11.3	ppbv	113			70	130	
	cis-1,2-Dichloroethene	10	10.2	ppbv	102			70	130	
	1,1,1-Trichloroethane	10	10.3	ppbv	103			70	130	
	Trichloroethene	10	10.3	ppbv	103			70	130	
	Tetrachloroethene	10	10.1	ppbv	101			70	130	