

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q2432
 Lab File ID: VL042670.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:23
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
SV1	0	0.00				
SV1DL	0	0.00				
SV1DUP	0	0.00				
VL0627ABL01	0	0.00				
VL0627ABS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +40% of internal standard area

AREA LOWER LIMIT = -40% of internal standard area

RT UPPER LIMIT = +0.33 minutes of internal standard RT

RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.