


LB136329

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.7741e-004

C= -3.2620e-002

Rho= 0.9994127

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/2)
0.0	0.000	-0.002	-0.002	171	0.000	171					
0.2	0.200	0.175	-0.025	1172	0.0 %	1172					-13
2.5	2.500	2.440	-0.060	13939	0.0 %	13939					-2
5.0	5.000	5.046	0.046	28627	0.0 %	28627					1
7.5	7.500	7.731	0.231	43760	0.0 %	43760					2
10.0	10.000	9.810	-0.190	55476	0.0 %	55476					-2

LB136329 INSTRUMENT ID : CV1

Method: 7471B Operator: Admin Date of Analysis: 30 Jun 2025 12:38:02

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Date	Type	Std	Conc	Method
S	0.0 - 1	50	-	171	PPB	30 Jun 2025 12:46:38	Std	0.0000	7471B	
S	0.2 - 1	50.2	-	1172	PPB	30 Jun 2025 12:52:20	Std	0.2000	7471B	
S	2.5 - 1	52.5	-	13939	PPB	30 Jun 2025 13:09:23	Std	2.5000	7471B	
S	5.0 - 1	55	-	28627	PPB	30 Jun 2025 13:14:25	Std	5.0000	7471B	
S	7.5 - 1	57.5	-	43760	PPB	30 Jun 2025 13:16:44	Std	7.5000	7471B	
S	10.0 - 1	516	-	55476	PPB	30 Jun 2025 13:26:17	Std	10.0000	7471B	
U	ICV12 - 1	ICV12	3.6368	20683	PPB	30 Jun 2025 13:29:23	SMPL	-	7471B	
U	ICB12 - 1	ICB12	-0.0617	-164	PPB	30 Jun 2025 13:31:38	SMPL	-	7471B	
U	CCV36 - 1	CCV36	4.9876	28297	PPB	30 Jun 2025 13:33:56	SMPL	-	7471B	
U	CCB36 - 1	CCB36	-0.0738	-232	PPB	30 Jun 2025 13:36:11	SMPL	-	7471B	
U	CRA - 1	CRA	0.1783	1189	PPB	30 Jun 2025 13:38:28	SMPL	-	7471B	
U	HighStd - 1	HighStd	9.9678	56368	PPB	30 Jun 2025 13:40:43	SMPL	-	7471B	
U	ChkStd - 1	ChkStd	7.2566	41086	PPB	30 Jun 2025 13:42:59	SMPL	-	7471B	
U	PB168665BL - 1	PBS	-0.0246	45	PPB	30 Jun 2025 13:48:05	SMPL	-	7471B	
U	Q2429-01 - 1	TP-4	0.2511	1599	PPB	30 Jun 2025 13:52:37	SMPL	-	7471B	
U	Q2430-01 - 1	MH-E/F	0.0520	477	PPB	30 Jun 2025 13:54:54	SMPL	-	7471B	
U	Q2436-01 - 1	TP-70	-0.0088	134	PPB	30 Jun 2025 13:57:11	SMPL	-	7471B	
U	Q2436-02 - 1	TP-69	0.3105	1934	PPB	30 Jun 2025 13:59:28	SMPL	-	7471B	
U	Q2436-03 - 1	TP-85	0.0536	486	PPB	30 Jun 2025 14:01:43	SMPL	-	7471B	
U	CCV37 - 1	CCV37	5.0829	28834	PPB	30 Jun 2025 14:03:59	SMPL	-	7471B	
U	CCB37 - 1	CCB37	-0.0835	-287	PPB	30 Jun 2025 14:06:14	SMPL	-	7471B	
U	Q2436-04 - 1	TP-86	0.0699	578	PPB	30 Jun 2025 14:08:35	SMPL	-	7471B	
U	Q2436-05 - 1	TP-84	0.0692	574	PPB	30 Jun 2025 14:10:50	SMPL	-	7471B	
U	Q2436-06 - 1	TP-83	0.1473	1014	PPB	30 Jun 2025 14:13:06	SMPL	-	7471B	
U	Q2436-07 - 1	TP-87	0.1441	996	PPB	30 Jun 2025 14:15:21	SMPL	-	7471B	
U	Q2436-08 - 1	TP-100	0.1331	934	PPB	30 Jun 2025 14:17:37	SMPL	-	7471B	
U	Q2436-09 - 1	TP-99	0.0444	434	PPB	30 Jun 2025 14:19:53	SMPL	-	7471B	
U	Q2436-10 - 1	TP-82	0.0680	567	PPB	30 Jun 2025 14:22:09	SMPL	-	7471B	
U	Q2442-01 - 1	500-3B	-0.0273	30	PPB	30 Jun 2025 14:24:26	SMPL	-	7471B	
U	Q2452-01 - 1	TP-5	0.0176	283	PPB	30 Jun 2025 14:26:43	SMPL	-	7471B	
U	Q2452-05 - 1	EP-2	0.2379	1525	PPB	30 Jun 2025 14:28:59	SMPL	-	7471B	
U	CCV38 - 1	CCV38	5.1650	29297	PPB	30 Jun 2025 14:31:14	SMPL	-	7471B	
U	CCB38 - 1	CCB38	-0.0857	-299	PPB	30 Jun 2025 14:33:29	SMPL	-	7471B	
U	Q2452-05DUP - 1	EP-2DUP	0.2401	1537	PPB	30 Jun 2025 14:35:49	SMPL	-	7471B	
U	Q2452-05MS - 1	EP-2MS	4.0539	23034	PPB	30 Jun 2025 14:46:32	SMPL	-	7471B	
U	Q2452-05MSD - 1	EP-2MSD	4.3220	24545	PPB	30 Jun 2025 14:48:46	SMPL	-	7471B	
U	Q2452-05LX5 - 1		0.0006	187	PPB	30 Jun 2025 14:51:05	SMPL	-	7471B	
U	Q2452-05A - 1		3.8937	22131	PPB	30 Jun 2025 14:53:24	SMPL	-	7471B	
U	PB168665BS - 1	LCSS	3.6207	20592	PPB	30 Jun 2025 14:55:40	SMPL	-	7471B	
U	CCV39 - 1	CCV39	4.9976	28353	PPB	30 Jun 2025 14:57:59	SMPL	-	7471B	
U	CCB39 - 1	CCB39	-0.0722	-223	PPB	30 Jun 2025 15:00:19	SMPL	-	7471B	