

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV12	Mercury	3.64	4.0	91	90 - 110	CV	06/30/2025	13:29	LB136329

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV36	Mercury	4.99	5.0	100	90 - 110	CV	06/30/2025	13:33	LB136329
CCV37	Mercury	5.08	5.0	102	90 - 110	CV	06/30/2025	14:03	LB136329
CCV38	Mercury	5.17	5.0	103	90 - 110	CV	06/30/2025	14:31	LB136329
CCV39	Mercury	5.00	5.0	100	90 - 110	CV	06/30/2025	14:57	LB136329

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q2436
Contract: CAMP02 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7650	8000	96	90 - 110	P	07/01/2025	13:30	LB136344
	Antimony	4240	4000	106	90 - 110	P	07/01/2025	13:30	LB136344
	Arsenic	4150	4000	104	90 - 110	P	07/01/2025	13:30	LB136344
	Barium	7710	8000	96	90 - 110	P	07/01/2025	13:30	LB136344
	Beryllium	195	200	97	90 - 110	P	07/01/2025	13:30	LB136344
	Cadmium	2040	2000	102	90 - 110	P	07/01/2025	13:30	LB136344
	Calcium	19100	20000	95	90 - 110	P	07/01/2025	13:30	LB136344
	Chromium	815	800	102	90 - 110	P	07/01/2025	13:30	LB136344
	Cobalt	2050	2000	102	90 - 110	P	07/01/2025	13:30	LB136344
	Copper	1050	1000	105	90 - 110	P	07/01/2025	13:30	LB136344
	Iron	4030	4000	101	90 - 110	P	07/01/2025	13:30	LB136344
	Lead	4000	4000	100	90 - 110	P	07/01/2025	13:30	LB136344
	Magnesium	18900	20000	95	90 - 110	P	07/01/2025	13:30	LB136344
	Manganese	1910	2000	96	90 - 110	P	07/01/2025	13:30	LB136344
	Nickel	2050	2000	103	90 - 110	P	07/01/2025	13:30	LB136344
	Potassium	20600	20000	103	90 - 110	P	07/01/2025	13:30	LB136344
	Selenium	4200	4000	105	90 - 110	P	07/01/2025	13:30	LB136344
	Silver	903	1000	90	90 - 110	P	07/01/2025	13:30	LB136344
	Sodium	20600	20000	103	90 - 110	P	07/01/2025	13:30	LB136344
	Thallium	3900	4000	98	90 - 110	P	07/01/2025	13:30	LB136344
	Vanadium	1900	2000	95	90 - 110	P	07/01/2025	13:30	LB136344
	Zinc	2050	2000	102	90 - 110	P	07/01/2025	13:30	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	99.9	100	100	80 - 120	P	07/01/2025	13:34	LB136344
	Antimony	51.6	50.0	103	80 - 120	P	07/01/2025	13:34	LB136344
	Arsenic	22.8	20.0	114	80 - 120	P	07/01/2025	13:34	LB136344
	Barium	91.3	100	91	80 - 120	P	07/01/2025	13:34	LB136344
	Beryllium	6.06	6.0	101	80 - 120	P	07/01/2025	13:34	LB136344
	Cadmium	5.90	6.0	98	80 - 120	P	07/01/2025	13:34	LB136344
	Calcium	1990	2000	99	80 - 120	P	07/01/2025	13:34	LB136344
	Chromium	10.7	10.0	107	80 - 120	P	07/01/2025	13:34	LB136344
	Cobalt	30.0	30.0	100	80 - 120	P	07/01/2025	13:34	LB136344
	Copper	21.3	20.0	107	80 - 120	P	07/01/2025	13:34	LB136344
	Iron	106	100	106	80 - 120	P	07/01/2025	13:34	LB136344
	Lead	12.0	12.0	100	80 - 120	P	07/01/2025	13:34	LB136344
	Magnesium	2030	2000	102	80 - 120	P	07/01/2025	13:34	LB136344
	Manganese	20.6	20.0	103	80 - 120	P	07/01/2025	13:34	LB136344
	Nickel	40.9	40.0	102	80 - 120	P	07/01/2025	13:34	LB136344
	Potassium	1980	2000	99	80 - 120	P	07/01/2025	13:34	LB136344
	Selenium	23.7	20.0	118	80 - 120	P	07/01/2025	13:34	LB136344
	Silver	9.72	10.0	97	80 - 120	P	07/01/2025	13:34	LB136344
	Sodium	2010	2000	101	80 - 120	P	07/01/2025	13:34	LB136344
	Thallium	41.3	40.0	103	80 - 120	P	07/01/2025	13:34	LB136344
	Vanadium	39.2	40.0	98	80 - 120	P	07/01/2025	13:34	LB136344
	Zinc	41.8	40.0	105	80 - 120	P	07/01/2025	13:34	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436

Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9760	10000	98	90 - 110	P	07/01/2025	14:09	LB136344
	Antimony	4990	5000	100	90 - 110	P	07/01/2025	14:09	LB136344
	Arsenic	5000	5000	100	90 - 110	P	07/01/2025	14:09	LB136344
	Barium	9930	10000	99	90 - 110	P	07/01/2025	14:09	LB136344
	Beryllium	244	250	97	90 - 110	P	07/01/2025	14:09	LB136344
	Cadmium	2500	2500	100	90 - 110	P	07/01/2025	14:09	LB136344
	Calcium	24300	25000	97	90 - 110	P	07/01/2025	14:09	LB136344
	Chromium	1010	1000	101	90 - 110	P	07/01/2025	14:09	LB136344
	Cobalt	2480	2500	99	90 - 110	P	07/01/2025	14:09	LB136344
	Copper	1250	1250	100	90 - 110	P	07/01/2025	14:09	LB136344
	Iron	5180	5000	104	90 - 110	P	07/01/2025	14:09	LB136344
	Lead	4980	5000	100	90 - 110	P	07/01/2025	14:09	LB136344
	Magnesium	24200	25000	97	90 - 110	P	07/01/2025	14:09	LB136344
	Manganese	2450	2500	98	90 - 110	P	07/01/2025	14:09	LB136344
	Nickel	2480	2500	99	90 - 110	P	07/01/2025	14:09	LB136344
	Potassium	25900	25000	104	90 - 110	P	07/01/2025	14:09	LB136344
	Selenium	5020	5000	100	90 - 110	P	07/01/2025	14:09	LB136344
	Silver	1270	1250	102	90 - 110	P	07/01/2025	14:09	LB136344
	Sodium	26100	25000	104	90 - 110	P	07/01/2025	14:09	LB136344
	Thallium	5050	5000	101	90 - 110	P	07/01/2025	14:09	LB136344
	Vanadium	2440	2500	98	90 - 110	P	07/01/2025	14:09	LB136344
	Zinc	2550	2500	102	90 - 110	P	07/01/2025	14:09	LB136344
CCV02	Aluminum	9380	10000	94	90 - 110	P	07/01/2025	15:42	LB136344
	Antimony	4900	5000	98	90 - 110	P	07/01/2025	15:42	LB136344
	Arsenic	4910	5000	98	90 - 110	P	07/01/2025	15:42	LB136344
	Barium	9330	10000	93	90 - 110	P	07/01/2025	15:42	LB136344
	Beryllium	230	250	92	90 - 110	P	07/01/2025	15:42	LB136344
	Cadmium	2400	2500	96	90 - 110	P	07/01/2025	15:42	LB136344
	Calcium	23200	25000	93	90 - 110	P	07/01/2025	15:42	LB136344
	Chromium	960	1000	96	90 - 110	P	07/01/2025	15:42	LB136344
	Cobalt	2390	2500	96	90 - 110	P	07/01/2025	15:42	LB136344
	Copper	1220	1250	98	90 - 110	P	07/01/2025	15:42	LB136344
	Iron	4960	5000	99	90 - 110	P	07/01/2025	15:42	LB136344
	Lead	4780	5000	96	90 - 110	P	07/01/2025	15:42	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436

Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	22900	25000	92	90 - 110	P	07/01/2025	15:42	LB136344
	Manganese	2370	2500	95	90 - 110	P	07/01/2025	15:42	LB136344
	Nickel	2400	2500	96	90 - 110	P	07/01/2025	15:42	LB136344
	Potassium	25000	25000	100	90 - 110	P	07/01/2025	15:42	LB136344
	Selenium	4950	5000	99	90 - 110	P	07/01/2025	15:42	LB136344
	Silver	1220	1250	98	90 - 110	P	07/01/2025	15:42	LB136344
	Sodium	25900	25000	104	90 - 110	P	07/01/2025	15:42	LB136344
	Thallium	4660	5000	93	90 - 110	P	07/01/2025	15:42	LB136344
	Vanadium	2340	2500	94	90 - 110	P	07/01/2025	15:42	LB136344
	Zinc	2470	2500	99	90 - 110	P	07/01/2025	15:42	LB136344
CCV03	Aluminum	9310	10000	93	90 - 110	P	07/01/2025	16:37	LB136344
	Antimony	4960	5000	99	90 - 110	P	07/01/2025	16:37	LB136344
	Arsenic	4900	5000	98	90 - 110	P	07/01/2025	16:37	LB136344
	Barium	9410	10000	94	90 - 110	P	07/01/2025	16:37	LB136344
	Beryllium	228	250	91	90 - 110	P	07/01/2025	16:37	LB136344
	Cadmium	2350	2500	94	90 - 110	P	07/01/2025	16:37	LB136344
	Calcium	23100	25000	93	90 - 110	P	07/01/2025	16:37	LB136344
	Chromium	941	1000	94	90 - 110	P	07/01/2025	16:37	LB136344
	Cobalt	2370	2500	95	90 - 110	P	07/01/2025	16:37	LB136344
	Copper	1200	1250	96	90 - 110	P	07/01/2025	16:37	LB136344
	Iron	4810	5000	96	90 - 110	P	07/01/2025	16:37	LB136344
	Lead	4730	5000	94	90 - 110	P	07/01/2025	16:37	LB136344
	Magnesium	24400	25000	97	90 - 110	P	07/01/2025	16:37	LB136344
	Manganese	2340	2500	94	90 - 110	P	07/01/2025	16:37	LB136344
	Nickel	2370	2500	95	90 - 110	P	07/01/2025	16:37	LB136344
	Potassium	24200	25000	97	90 - 110	P	07/01/2025	16:37	LB136344
	Selenium	4980	5000	100	90 - 110	P	07/01/2025	16:37	LB136344
	Silver	1210	1250	96	90 - 110	P	07/01/2025	16:37	LB136344
	Sodium	25900	25000	103	90 - 110	P	07/01/2025	16:37	LB136344
	Thallium	4740	5000	95	90 - 110	P	07/01/2025	16:37	LB136344
	Vanadium	2310	2500	93	90 - 110	P	07/01/2025	16:37	LB136344
	Zinc	2440	2500	98	90 - 110	P	07/01/2025	16:37	LB136344
CCV04	Aluminum	9450	10000	94	90 - 110	P	07/01/2025	17:26	LB136344
	Antimony	5160	5000	103	90 - 110	P	07/01/2025	17:26	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q2436
Contract: CAMP02 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5110	5000	102	90 - 110	P	07/01/2025	17:26	LB136344
	Barium	9520	10000	95	90 - 110	P	07/01/2025	17:26	LB136344
	Beryllium	230	250	92	90 - 110	P	07/01/2025	17:26	LB136344
	Cadmium	2380	2500	95	90 - 110	P	07/01/2025	17:26	LB136344
	Calcium	23600	25000	94	90 - 110	P	07/01/2025	17:26	LB136344
	Chromium	932	1000	93	90 - 110	P	07/01/2025	17:26	LB136344
	Cobalt	2400	2500	96	90 - 110	P	07/01/2025	17:26	LB136344
	Copper	1230	1250	99	90 - 110	P	07/01/2025	17:26	LB136344
	Iron	4740	5000	95	90 - 110	P	07/01/2025	17:26	LB136344
	Lead	4780	5000	96	90 - 110	P	07/01/2025	17:26	LB136344
	Magnesium	24700	25000	99	90 - 110	P	07/01/2025	17:26	LB136344
	Manganese	2390	2500	96	90 - 110	P	07/01/2025	17:26	LB136344
	Nickel	2400	2500	96	90 - 110	P	07/01/2025	17:26	LB136344
	Potassium	24000	25000	96	90 - 110	P	07/01/2025	17:26	LB136344
	Selenium	5250	5000	105	90 - 110	P	07/01/2025	17:26	LB136344
	Silver	1200	1250	96	90 - 110	P	07/01/2025	17:26	LB136344
	Sodium	26000	25000	104	90 - 110	P	07/01/2025	17:26	LB136344
	Thallium	4820	5000	96	90 - 110	P	07/01/2025	17:26	LB136344
	Vanadium	2360	2500	94	90 - 110	P	07/01/2025	17:26	LB136344
	Zinc	2420	2500	97	90 - 110	P	07/01/2025	17:26	LB136344
CCV05	Aluminum	9530	10000	95	90 - 110	P	07/01/2025	18:27	LB136344
	Antimony	5380	5000	108	90 - 110	P	07/01/2025	18:27	LB136344
	Arsenic	5290	5000	106	90 - 110	P	07/01/2025	18:27	LB136344
	Barium	9750	10000	98	90 - 110	P	07/01/2025	18:27	LB136344
	Beryllium	231	250	92	90 - 110	P	07/01/2025	18:27	LB136344
	Cadmium	2410	2500	97	90 - 110	P	07/01/2025	18:27	LB136344
	Calcium	23500	25000	94	90 - 110	P	07/01/2025	18:27	LB136344
	Chromium	949	1000	95	90 - 110	P	07/01/2025	18:27	LB136344
	Cobalt	2450	2500	98	90 - 110	P	07/01/2025	18:27	LB136344
	Copper	1270	1250	102	90 - 110	P	07/01/2025	18:27	LB136344
	Iron	4990	5000	100	90 - 110	P	07/01/2025	18:27	LB136344
	Lead	4870	5000	97	90 - 110	P	07/01/2025	18:27	LB136344
	Magnesium	24600	25000	98	90 - 110	P	07/01/2025	18:27	LB136344
	Manganese	2400	2500	96	90 - 110	P	07/01/2025	18:27	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2450	2500	98	90 - 110	P	07/01/2025	18:27	LB136344
	Potassium	25100	25000	100	90 - 110	P	07/01/2025	18:27	LB136344
	Selenium	5440	5000	109	90 - 110	P	07/01/2025	18:27	LB136344
	Silver	1240	1250	99	90 - 110	P	07/01/2025	18:27	LB136344
	Sodium	27300	25000	109	90 - 110	P	07/01/2025	18:27	LB136344
	Thallium	4780	5000	96	90 - 110	P	07/01/2025	18:27	LB136344
	Vanadium	2370	2500	95	90 - 110	P	07/01/2025	18:27	LB136344
	Zinc	2530	2500	101	90 - 110	P	07/01/2025	18:27	LB136344
CCV06	Aluminum	9270	10000	93	90 - 110	P	07/01/2025	20:05	LB136344
	Antimony	5250	5000	105	90 - 110	P	07/01/2025	20:05	LB136344
	Arsenic	5150	5000	103	90 - 110	P	07/01/2025	20:05	LB136344
	Barium	9560	10000	96	90 - 110	P	07/01/2025	20:05	LB136344
	Beryllium	234	250	94	90 - 110	P	07/01/2025	20:05	LB136344
	Cadmium	2320	2500	93	90 - 110	P	07/01/2025	20:05	LB136344
	Calcium	23100	25000	92	90 - 110	P	07/01/2025	20:05	LB136344
	Chromium	918	1000	92	90 - 110	P	07/01/2025	20:05	LB136344
	Cobalt	2370	2500	95	90 - 110	P	07/01/2025	20:05	LB136344
	Copper	1230	1250	98	90 - 110	P	07/01/2025	20:05	LB136344
	Iron	4810	5000	96	90 - 110	P	07/01/2025	20:05	LB136344
	Lead	4670	5000	94	90 - 110	P	07/01/2025	20:05	LB136344
	Magnesium	22500	25000	90	90 - 110	P	07/01/2025	20:05	LB136344
	Manganese	2340	2500	94	90 - 110	P	07/01/2025	20:05	LB136344
	Nickel	2360	2500	94	90 - 110	P	07/01/2025	20:05	LB136344
	Potassium	24300	25000	97	90 - 110	P	07/01/2025	20:05	LB136344
	Selenium	5310	5000	106	90 - 110	P	07/01/2025	20:05	LB136344
	Silver	1200	1250	96	90 - 110	P	07/01/2025	20:05	LB136344
	Sodium	24900	25000	100	90 - 110	P	07/01/2025	20:05	LB136344
	Thallium	4600	5000	92	90 - 110	P	07/01/2025	20:05	LB136344
	Vanadium	2330	2500	93	90 - 110	P	07/01/2025	20:05	LB136344
	Zinc	2430	2500	97	90 - 110	P	07/01/2025	20:05	LB136344
CCV07	Aluminum	9140	10000	91	90 - 110	P	07/01/2025	21:22	LB136344
	Antimony	4990	5000	100	90 - 110	P	07/01/2025	21:22	LB136344
	Arsenic	4870	5000	98	90 - 110	P	07/01/2025	21:22	LB136344
	Barium	9540	10000	95	90 - 110	P	07/01/2025	21:22	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q2436
Contract: CAMP02 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	227	250	91	90 - 110	P	07/01/2025	21:22	LB136344
	Cadmium	2290	2500	92	90 - 110	P	07/01/2025	21:22	LB136344
	Calcium	22800	25000	91	90 - 110	P	07/01/2025	21:22	LB136344
	Chromium	933	1000	93	90 - 110	P	07/01/2025	21:22	LB136344
	Cobalt	2330	2500	93	90 - 110	P	07/01/2025	21:22	LB136344
	Copper	1190	1250	95	90 - 110	P	07/01/2025	21:22	LB136344
	Iron	5110	5000	102	90 - 110	P	07/01/2025	21:22	LB136344
	Lead	4610	5000	92	90 - 110	P	07/01/2025	21:22	LB136344
	Magnesium	24000	25000	96	90 - 110	P	07/01/2025	21:22	LB136344
	Manganese	2300	2500	92	90 - 110	P	07/01/2025	21:22	LB136344
	Nickel	2320	2500	93	90 - 110	P	07/01/2025	21:22	LB136344
	Potassium	25600	25000	102	90 - 110	P	07/01/2025	21:22	LB136344
	Selenium	4970	5000	99	90 - 110	P	07/01/2025	21:22	LB136344
	Silver	1220	1250	98	90 - 110	P	07/01/2025	21:22	LB136344
	Sodium	27400	25000	110	90 - 110	P	07/01/2025	21:22	LB136344
	Thallium	4600	5000	92	90 - 110	P	07/01/2025	21:22	LB136344
	Vanadium	2300	2500	92	90 - 110	P	07/01/2025	21:22	LB136344
	Zinc	2480	2500	99	90 - 110	P	07/01/2025	21:22	LB136344
CCV08	Aluminum	9280	10000	93	90 - 110	P	07/01/2025	21:50	LB136344
	Antimony	5110	5000	102	90 - 110	P	07/01/2025	21:50	LB136344
	Arsenic	5040	5000	101	90 - 110	P	07/01/2025	21:50	LB136344
	Barium	9530	10000	95	90 - 110	P	07/01/2025	21:50	LB136344
	Beryllium	238	250	95	90 - 110	P	07/01/2025	21:50	LB136344
	Cadmium	2350	2500	94	90 - 110	P	07/01/2025	21:50	LB136344
	Calcium	23100	25000	92	90 - 110	P	07/01/2025	21:50	LB136344
	Chromium	939	1000	94	90 - 110	P	07/01/2025	21:50	LB136344
	Cobalt	2370	2500	95	90 - 110	P	07/01/2025	21:50	LB136344
	Copper	1230	1250	98	90 - 110	P	07/01/2025	21:50	LB136344
	Iron	5080	5000	102	90 - 110	P	07/01/2025	21:50	LB136344
	Lead	4690	5000	94	90 - 110	P	07/01/2025	21:50	LB136344
	Magnesium	23500	25000	94	90 - 110	P	07/01/2025	21:50	LB136344
	Manganese	2320	2500	93	90 - 110	P	07/01/2025	21:50	LB136344
	Nickel	2370	2500	95	90 - 110	P	07/01/2025	21:50	LB136344
	Potassium	25700	25000	103	90 - 110	P	07/01/2025	21:50	LB136344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: CDM Smith SDG No.: Q2436
Contract: CAMP02 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5170	5000	103	90 - 110	P	07/01/2025	21:50	LB136344
	Silver	1230	1250	98	90 - 110	P	07/01/2025	21:50	LB136344
	Sodium	23900	25000	96	90 - 110	P	07/01/2025	21:50	LB136344
	Thallium	4710	5000	94	90 - 110	P	07/01/2025	21:50	LB136344
	Vanadium	2330	2500	93	90 - 110	P	07/01/2025	21:50	LB136344
	Zinc	2530	2500	101	90 - 110	P	07/01/2025	21:50	LB136344