



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	89	70 - 130	CV	06/30/2025	13:38	LB136329
CRI01	Aluminum	95.2	100	95	65 - 135	P	07/01/2025	13:48	LB136344
	Antimony	49.5	50.0	99	65 - 135	P	07/01/2025	13:48	LB136344
	Arsenic	23.6	20.0	118	65 - 135	P	07/01/2025	13:48	LB136344
	Barium	87.2	100	87	65 - 135	P	07/01/2025	13:48	LB136344
	Beryllium	5.86	6.0	98	65 - 135	P	07/01/2025	13:48	LB136344
	Cadmium	5.88	6.0	98	65 - 135	P	07/01/2025	13:48	LB136344
	Calcium	1970	2000	99	65 - 135	P	07/01/2025	13:48	LB136344
	Chromium	10.7	10.0	107	65 - 135	P	07/01/2025	13:48	LB136344
	Cobalt	30.0	30.0	100	65 - 135	P	07/01/2025	13:48	LB136344
	Copper	21.5	20.0	107	65 - 135	P	07/01/2025	13:48	LB136344
	Iron	111	100	111	65 - 135	P	07/01/2025	13:48	LB136344
	Lead	11.1	12.0	93	65 - 135	P	07/01/2025	13:48	LB136344
	Magnesium	2030	2000	101	65 - 135	P	07/01/2025	13:48	LB136344
	Manganese	19.7	20.0	98	65 - 135	P	07/01/2025	13:48	LB136344
	Nickel	40.7	40.0	102	65 - 135	P	07/01/2025	13:48	LB136344
	Potassium	2100	2000	105	65 - 135	P	07/01/2025	13:48	LB136344
	Selenium	18.4	20.0	92	65 - 135	P	07/01/2025	13:48	LB136344
	Silver	9.62	10.0	96	65 - 135	P	07/01/2025	13:48	LB136344
	Sodium	2070	2000	104	65 - 135	P	07/01/2025	13:48	LB136344
	Thallium	41.5	40.0	104	65 - 135	P	07/01/2025	13:48	LB136344
	Vanadium	38.8	40.0	97	65 - 135	P	07/01/2025	13:48	LB136344
	Zinc	41.6	40.0	104	65 - 135	P	07/01/2025	13:48	LB136344