

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168648BS							
Aluminum	mg/Kg	100	91.1		91	80 - 120	P
Antimony	mg/Kg	40.0	38.9		97	80 - 120	P
Arsenic	mg/Kg	40.0	37.5		94	80 - 120	P
Barium	mg/Kg	10.0	8.43		84	80 - 120	P
Beryllium	mg/Kg	10.0	9.28		93	80 - 120	P
Cadmium	mg/Kg	10.0	9.19		92	80 - 120	P
Calcium	mg/Kg	50.0	46.8	J	94	80 - 120	P
Chromium	mg/Kg	20.0	19.0		95	80 - 120	P
Cobalt	mg/Kg	10.0	9.41		94	80 - 120	P
Copper	mg/Kg	15.0	14.8		99	80 - 120	P
Iron	mg/Kg	150	142		95	80 - 120	P
Lead	mg/Kg	50.0	46.1		92	80 - 120	P
Magnesium	mg/Kg	100	91.5	J	92	80 - 120	P
Manganese	mg/Kg	10.0	9.41		94	80 - 120	P
Nickel	mg/Kg	25.0	23.5		94	80 - 120	P
Potassium	mg/Kg	500	470		94	80 - 120	P
Selenium	mg/Kg	100	95.8		96	80 - 120	P
Silver	mg/Kg	3.8	3.52		93	80 - 120	P
Sodium	mg/Kg	150	148		99	80 - 120	P
Thallium	mg/Kg	100	94.2		94	80 - 120	P
Vanadium	mg/Kg	15.0	13.6		91	80 - 120	P
Zinc	mg/Kg	10.0	9.98		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2436
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2436 **SAS No.:** Q2436

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168665BS Mercury	mg/Kg	0.27	0.24		90	80 - 120	CV