

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG No.: Q2436
DataFile: FF016043.D Analyst Name: YP\AJ Analyst Date: 06-30-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	65882919	131766	134918	2.336