

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** CDM Smith
Lab Code: CHEM **Cas No:** Q2436 **SAS No :** Q2436 **SDG No:** Q2436
Matrix Spike - EPA Sample No : PB168658BS **Datafile:** FF016046.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6662	0	4925	74	68-131