

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2436
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030881.D **Time Analyzed:** 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.655	9.568	9.768	755.240	712.500	6.0
2,4,5-TP(Silvex)	9.360	9.275	9.475	708.050	712.500	-0.6
2,4-D	8.474	8.386	8.586	682.290	705.000	-3.2
2,4-DB	10.232	10.147	10.347	812.680	712.500	14.1
2,4-DCAA	7.339	7.249	7.449	691.520	750.000	-7.8
DICAMBA	7.529	7.439	7.639	690.430	705.000	-2.1
DICHLORPROP	8.242	8.153	8.353	644.410	705.000	-8.6
Dinoseb	11.448	11.366	11.566	697.130	705.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2436
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030881.D **Time Analyzed:** 09:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.356	10.259	10.459	675.040	712.500	-5.3
2,4,5-TP(Silvex)	9.930	9.833	10.033	667.540	712.500	-6.3
2,4-D	9.025	8.927	9.127	624.960	705.000	-11.4
2,4-DB	10.924	10.827	11.027	656.470	712.500	-7.9
2,4-DCAA	7.768	7.670	7.870	676.460	750.000	-9.8
DICAMBA	7.971	7.872	8.072	655.070	705.000	-7.1
DICHLORPROP	8.689	8.591	8.791	633.220	705.000	-10.2
Dinoseb	11.309	11.211	11.411	644.720	705.000	-8.6