

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2436  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/02/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS030893.D **Time Analyzed:** 14:47

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.655  | 9.568     | 9.768  | 765.480            | 712.500           | 7.4  |
| 2,4,5-TP(Silvex) | 9.360  | 9.275     | 9.475  | 709.250            | 712.500           | -0.5 |
| 2,4-D            | 8.473  | 8.386     | 8.586  | 688.390            | 705.000           | -2.4 |
| 2,4-DB           | 10.233 | 10.147    | 10.347 | 807.800            | 712.500           | 13.4 |
| 2,4-DCAA         | 7.338  | 7.249     | 7.449  | 711.720            | 750.000           | -5.1 |
| DICAMBA          | 7.528  | 7.439     | 7.639  | 661.150            | 705.000           | -6.2 |
| DICHLORPROP      | 8.242  | 8.153     | 8.353  | 641.370            | 705.000           | -9.0 |
| Dinoseb          | 11.450 | 11.366    | 11.566 | 683.640            | 705.000           | -3.0 |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2436  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/02/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS030893.D **Time Analyzed:** 14:47

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                  |        | FROM      | TO     |                    |                   |       |
| 2,4,5-T          | 10.356 | 10.259    | 10.459 | 682.460            | 712.500           | -4.2  |
| 2,4,5-TP(Silvex) | 9.931  | 9.833     | 10.033 | 669.160            | 712.500           | -6.1  |
| 2,4-D            | 9.025  | 8.927     | 9.127  | 632.690            | 705.000           | -10.3 |
| 2,4-DB           | 10.924 | 10.827    | 11.027 | 653.090            | 712.500           | -8.3  |
| 2,4-DCAA         | 7.768  | 7.670     | 7.870  | 683.580            | 750.000           | -8.9  |
| DICAMBA          | 7.970  | 7.872     | 8.072  | 664.460            | 705.000           | -5.8  |
| DICHLORPROP      | 8.690  | 8.591     | 8.791  | 628.270            | 705.000           | -10.9 |
| Dinoseb          | 11.310 | 11.211    | 11.411 | 625.550            | 705.000           | -11.3 |