

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2436
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030899.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.655	9.568	9.768	769.470	712.500	8.0
2,4,5-TP(Silvex)	9.360	9.275	9.475	715.120	712.500	0.4
2,4-D	8.473	8.386	8.586	691.350	705.000	-1.9
2,4-DB	10.231	10.147	10.347	804.090	712.500	12.9
2,4-DCAA	7.339	7.249	7.449	709.980	750.000	-5.3
DICAMBA	7.528	7.439	7.639	692.000	705.000	-1.8
DICHLORPROP	8.241	8.153	8.353	647.910	705.000	-8.1
Dinoseb	11.449	11.366	11.566	707.210	705.000	0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2436
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/18/2025 06/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/02/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS030899.D **Time Analyzed:** 17:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.356	10.259	10.459	687.900	712.500	-3.5
2,4,5-TP(Silvex)	9.930	9.833	10.033	678.240	712.500	-4.8
2,4-D	9.024	8.927	9.127	640.000	705.000	-9.2
2,4-DB	10.923	10.827	11.027	660.360	712.500	-7.3
2,4-DCAA	7.767	7.670	7.870	686.620	750.000	-8.5
DICAMBA	7.970	7.872	8.072	667.720	705.000	-5.3
DICHLORPROP	8.689	8.591	8.791	643.660	705.000	-8.7
Dinoseb	11.308	11.211	11.411	657.150	705.000	-6.8