

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 18:08 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1268-1 (1)   | 8.69       | 8.70      | 8.60      | 8.80  | 0.01       |
| Aroclor-1268-2 (2)   | 8.79       | 8.79      | 8.69      | 8.89  | 0.01       |
| Aroclor-1268-3 (3)   | 9.01       | 9.02      | 8.92      | 9.12  | 0.01       |
| Aroclor-1268-4 (4)   | 9.43       | 9.44      | 9.34      | 9.54  | 0.01       |
| Aroclor-1268-5 (5)   | 9.84       | 9.85      | 9.75      | 9.95  | 0.01       |
| Tetrachloro-m-xylene | 4.49       | 4.49      | 4.39      | 4.59  | 0.00       |
| Decachlorobiphenyl   | 10.18      | 10.18     | 10.08     | 10.28 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 18:08 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1268-1       | (1) | 7.67       | 7.68      | 7.58      | 7.78 | 0.01       |
| Aroclor-1268-2       | (2) | 7.74       | 7.75      | 7.65      | 7.85 | 0.01       |
| Aroclor-1268-3       | (3) | 7.94       | 7.95      | 7.85      | 8.05 | 0.01       |
| Aroclor-1268-4       | (4) | 8.24       | 8.25      | 8.15      | 8.35 | 0.01       |
| Aroclor-1268-5       | (5) | 8.53       | 8.54      | 8.44      | 8.64 | 0.01       |
| Tetrachloro-m-xylene |     | 3.78       | 3.79      | 3.69      | 3.89 | 0.01       |
| Decachlorobiphenyl   |     | 8.79       | 8.80      | 8.70      | 8.90 | 0.01       |