

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 23:34 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	8.69	8.70	8.60	8.80	0.01
Aroclor-1268-2	(2)	8.78	8.79	8.69	8.89	0.01
Aroclor-1268-3	(3)	9.01	9.02	8.92	9.12	0.01
Aroclor-1268-4	(4)	9.43	9.44	9.34	9.54	0.01
Aroclor-1268-5	(5)	9.84	9.85	9.75	9.95	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.17	10.18	10.08	10.28	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/27/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 23:34 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1268-1	(1)	7.67	7.68	7.58	7.78	0.01
Aroclor-1268-2	(2)	7.74	7.75	7.65	7.85	0.01
Aroclor-1268-3	(3)	7.94	7.95	7.85	8.05	0.01
Aroclor-1268-4	(4)	8.24	8.25	8.15	8.35	0.02
Aroclor-1268-5	(5)	8.53	8.54	8.44	8.64	0.01
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.02