

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/28/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 04:29 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	5.64	5.64	5.54	5.74	0.00
Aroclor-1242-2	(2)	5.66	5.67	5.57	5.77	0.01
Aroclor-1242-3	(3)	5.72	5.73	5.63	5.83	0.01
Aroclor-1242-4	(4)	5.82	5.83	5.73	5.93	0.01
Aroclor-1242-5	(5)	6.55	6.56	6.46	6.66	0.01
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.18	10.19	10.09	10.29	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

Continuing Calib Date: 06/28/2025 Initial Calibration Date(s): 06/17/2025 06/17/2025

Continuing Calib Time: 04:29 Initial Calibration Time(s): 10:04 20:10

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.86	4.87	4.77	4.97	0.01
Aroclor-1242-2	(2)	4.88	4.89	4.79	4.99	0.01
Aroclor-1242-3	(3)	5.06	5.06	4.96	5.16	0.00
Aroclor-1242-4	(4)	5.14	5.15	5.05	5.25	0.01
Aroclor-1242-5	(5)	5.66	5.67	5.57	5.77	0.01
Tetrachloro-m-xylene		3.78	3.79	3.69	3.89	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01