

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP073329.D Time Analyzed: 16:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.642	5.546	5.746	409.590	500.000	-18.1
Aroclor-1016-2	5.663	5.567	5.767	441.080	500.000	-11.8
Aroclor-1016-3	5.725	5.629	5.829	420.560	500.000	-15.9
Aroclor-1016-4	5.822	5.727	5.927	441.350	500.000	-11.7
Aroclor-1016-5	6.114	6.019	6.219	429.890	500.000	-14.0
Aroclor-1260-1	7.230	7.136	7.336	455.300	500.000	-8.9
Aroclor-1260-2	7.485	7.389	7.589	452.950	500.000	-9.4
Aroclor-1260-3	7.842	7.747	7.947	431.750	500.000	-13.7
Aroclor-1260-4	8.066	7.971	8.171	421.940	500.000	-15.6
Aroclor-1260-5	8.383	8.288	8.488	416.460	500.000	-16.7
Decachlorobiphenyl	10.179	10.087	10.287	44.090	50.000	-11.8
Tetrachloro-m-xylene	4.490	4.394	4.594	42.330	50.000	-15.3

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP073329.D Time Analyzed: 16:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.863	4.768	4.968	412.820	500.000	-17.4
Aroclor-1016-2	4.881	4.786	4.986	411.940	500.000	-17.6
Aroclor-1016-3	5.057	4.962	5.162	406.130	500.000	-18.8
Aroclor-1016-4	5.099	5.004	5.204	405.400	500.000	-18.9
Aroclor-1016-5	5.312	5.218	5.418	418.410	500.000	-16.3
Aroclor-1260-1	6.343	6.249	6.449	412.060	500.000	-17.6
Aroclor-1260-2	6.531	6.438	6.638	425.960	500.000	-14.8
Aroclor-1260-3	6.683	6.589	6.789	408.460	500.000	-18.3
Aroclor-1260-4	7.153	7.060	7.260	416.930	500.000	-16.6
Aroclor-1260-5	7.395	7.302	7.502	411.320	500.000	-17.7
Decachlorobiphenyl	8.788	8.697	8.897	43.670	50.000	-12.7
Tetrachloro-m-xylene	3.784	3.688	3.888	41.990	50.000	-16.0