

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP073345.D Time Analyzed: 22:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.546	5.746	432.490	500.000	-13.5
Aroclor-1016-2	5.659	5.567	5.767	460.330	500.000	-7.9
Aroclor-1016-3	5.721	5.629	5.829	456.640	500.000	-8.7
Aroclor-1016-4	5.818	5.727	5.927	479.780	500.000	-4.0
Aroclor-1016-5	6.110	6.019	6.219	451.140	500.000	-9.8
Aroclor-1260-1	7.227	7.136	7.336	472.570	500.000	-5.5
Aroclor-1260-2	7.480	7.389	7.589	452.070	500.000	-9.6
Aroclor-1260-3	7.838	7.747	7.947	442.770	500.000	-11.4
Aroclor-1260-4	8.062	7.971	8.171	443.420	500.000	-11.3
Aroclor-1260-5	8.380	8.288	8.488	435.790	500.000	-12.8
Decachlorobiphenyl	10.174	10.087	10.287	45.500	50.000	-9.0
Tetrachloro-m-xylene	4.487	4.394	4.594	43.270	50.000	-13.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/27/2025

Lab Sample No.: AR1660CCC500 Data File : PP073345.D Time Analyzed: 22:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.862	4.768	4.968	446.220	500.000	-10.8
Aroclor-1016-2	4.880	4.786	4.986	449.690	500.000	-10.1
Aroclor-1016-3	5.056	4.962	5.162	449.220	500.000	-10.2
Aroclor-1016-4	5.098	5.004	5.204	450.630	500.000	-9.9
Aroclor-1016-5	5.311	5.218	5.418	464.140	500.000	-7.2
Aroclor-1260-1	6.341	6.249	6.449	464.610	500.000	-7.1
Aroclor-1260-2	6.529	6.438	6.638	475.500	500.000	-4.9
Aroclor-1260-3	6.681	6.589	6.789	462.110	500.000	-7.6
Aroclor-1260-4	7.151	7.060	7.260	462.150	500.000	-7.6
Aroclor-1260-5	7.393	7.302	7.502	465.070	500.000	-7.0
Decachlorobiphenyl	8.786	8.697	8.897	50.420	50.000	0.8
Tetrachloro-m-xylene	3.783	3.688	3.888	44.590	50.000	-10.8