

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062725\
 Data File : PP073338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2025 19:29
 Operator : YP\AJ
 Sample : Q2436-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-86

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/30/2025
 Supervised By :mohammad ahmed 07/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 23:23:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 18 08:17:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.487	3.783	38509417	33981032	18.244	18.860
2) SA Decachlor...	10.174	8.786	26033719	23313304	15.234	17.467m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062725\
Data File : PP073338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2025 19:29
Operator : YP\AJ
Sample : Q2436-04
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TP-86


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 07/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 23:23:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 18 08:17:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.487 min
Delta R.T.: -0.007 min
Response: 38509417
Conc: 18.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-86


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 07/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: -0.005 min
Response: 33981032
Conc: 18.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.174 min
Delta R.T.: -0.013 min
Response: 26033719
Conc: 15.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.786 min
Delta R.T.: -0.011 min
Response: 23313304
Conc: 17.47 ng/ml m