

Surrogate Summary

SDG No.: Q2436

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP072990.D	PIBLK-PP072990.D	Tetrachloro-m-xylene	1	20	17.3	86		60	140
		Decachlorobiphenyl	1	20	18.3	91		60	140
		Tetrachloro-m-xylene	2	20	18.0	90		60	140
		Decachlorobiphenyl	2	20	17.2	86		60	140
I.BLK-PP073320.D	PIBLK-PP073320.D	Tetrachloro-m-xylene	1	20	18.5	93		60	140
		Decachlorobiphenyl	1	20	19.0	95		60	140
		Tetrachloro-m-xylene	2	20	18.3	91		60	140
		Decachlorobiphenyl	2	20	20.4	102		60	140
PB168636BL	PB168636BL	Tetrachloro-m-xylene	1	20	18.5	92		32	144
		Decachlorobiphenyl	1	20	18.2	91		32	175
		Tetrachloro-m-xylene	2	20	18.4	92		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
PB168636BS	PB168636BS	Tetrachloro-m-xylene	1	20	19.2	96		32	144
		Decachlorobiphenyl	1	20	18.7	93		32	175
		Tetrachloro-m-xylene	2	20	18.8	94		32	144
		Decachlorobiphenyl	2	20	20.3	101		32	175
Q2430-01MS	MH-E/FMS	Tetrachloro-m-xylene	1	20	20.0	100		32	144
		Decachlorobiphenyl	1	20	18.0	90		32	175
		Tetrachloro-m-xylene	2	20	19.1	96		32	144
		Decachlorobiphenyl	2	20	20.1	100		32	175
Q2430-01MSD	MH-E/FMSD	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	17.7	89		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
		Decachlorobiphenyl	2	20	19.0	95		32	175
I.BLK-PP073334.D	PIBLK-PP073334.D	Tetrachloro-m-xylene	1	20	17.0	85		60	140
		Decachlorobiphenyl	1	20	17.3	87		60	140
		Tetrachloro-m-xylene	2	20	17.4	87		60	140
		Decachlorobiphenyl	2	20	19.1	95		60	140
Q2436-01	TP-70	Tetrachloro-m-xylene	1	20	17.1	86		32	144
		Decachlorobiphenyl	1	20	13.8	69		32	175
		Tetrachloro-m-xylene	2	20	17.6	88		32	144
		Decachlorobiphenyl	2	20	15.9	79		32	175
Q2436-02	TP-69	Tetrachloro-m-xylene	1	20	15.9	80		32	144
		Decachlorobiphenyl	1	20	12.0	60		32	175
		Tetrachloro-m-xylene	2	20	16.8	84		32	144
		Decachlorobiphenyl	2	20	14.2	71		32	175
Q2436-03	TP-85	Tetrachloro-m-xylene	1	20	14.8	74		32	144
		Decachlorobiphenyl	1	20	11.6	58		32	175
		Tetrachloro-m-xylene	2	20	15.2	76		32	144
		Decachlorobiphenyl	2	20	13.7	68		32	175
Q2436-04	TP-86	Tetrachloro-m-xylene	1	20	18.2	91		32	144

Surrogate Summary

SDG No.: Q2436

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2436-04	TP-86	Decachlorobiphenyl	1	20	15.2	76		32	175
		Tetrachloro-m-xylene	2	20	18.9	94		32	144
Q2436-05	TP-84	Decachlorobiphenyl	2	20	17.5	87		32	175
		Tetrachloro-m-xylene	1	20	17.5	87		32	144
		Decachlorobiphenyl	1	20	15.1	75		32	175
		Tetrachloro-m-xylene	2	20	19.3	97		32	144
		Decachlorobiphenyl	2	20	17.4	87		32	175
		Tetrachloro-m-xylene	1	20	16.6	83		32	144
Q2436-06	TP-83	Decachlorobiphenyl	1	20	14.5	72		32	175
		Tetrachloro-m-xylene	2	20	17.6	88		32	144
		Decachlorobiphenyl	2	20	16.7	83		32	175
		Tetrachloro-m-xylene	1	20	16.7	83		32	144
Q2436-07	TP-87	Decachlorobiphenyl	1	20	13.7	68		32	175
		Tetrachloro-m-xylene	2	20	17.8	89		32	144
		Decachlorobiphenyl	2	20	15.7	79		32	175
		Tetrachloro-m-xylene	1	20	17.4	87		32	144
Q2436-08	TP-100	Decachlorobiphenyl	1	20	13.6	68		32	175
		Tetrachloro-m-xylene	2	20	18.6	93		32	144
		Decachlorobiphenyl	2	20	15.1	76		32	175
		Tetrachloro-m-xylene	1	20	17.8	89		32	144
Q2436-09	TP-99	Decachlorobiphenyl	1	20	15.7	79		32	175
		Tetrachloro-m-xylene	2	20	18.6	93		32	144
		Decachlorobiphenyl	2	20	17.7	88		32	175
		Tetrachloro-m-xylene	1	20	17.2	86		32	144
Q2436-10	TP-82	Decachlorobiphenyl	1	20	16.1	81		32	175
		Tetrachloro-m-xylene	2	20	18.4	92		32	144
		Decachlorobiphenyl	2	20	18.1	90		32	175
		Tetrachloro-m-xylene	1	20	17.4	87		60	140
I.BLK-PP073350.D	PIBLK-PP073350.D	Decachlorobiphenyl	1	20	17.8	89		60	140
		Tetrachloro-m-xylene	2	20	18.6	93		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
		Tetrachloro-m-xylene	1	20	17.4	87		60	140