

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.920	50.000	15.8
4,4'-DDE	6.194	6.096	6.296	54.700	50.000	9.4
4,4'-DDT	7.019	6.921	7.121	42.000	50.000	-16.0
Aldrin	5.270	5.171	5.371	56.890	50.000	13.8
alpha-BHC	3.999	3.899	4.099	58.040	50.000	16.1
alpha-Chlordane	6.026	5.927	6.127	56.470	50.000	12.9
beta-BHC	4.515	4.415	4.615	54.780	50.000	9.6
Decachlorobiphenyl	9.070	8.972	9.172	50.550	50.000	1.1
delta-BHC	4.763	4.664	4.864	59.190	50.000	18.4
Dieldrin	6.345	6.247	6.447	55.790	50.000	11.6
Endosulfan I	6.073	5.975	6.175	55.380	50.000	10.8
Endosulfan II	6.785	6.686	6.886	54.540	50.000	9.1
Endosulfan sulfate	7.148	7.049	7.249	52.730	50.000	5.5
Endrin	6.573	6.475	6.675	50.190	50.000	0.4
Endrin aldehyde	6.913	6.815	7.015	50.620	50.000	1.2
Endrin ketone	7.628	7.530	7.730	52.990	50.000	6.0
gamma-BHC (Lindane)	4.330	4.230	4.430	56.990	50.000	14.0
gamma-Chlordane	5.945	5.846	6.046	57.950	50.000	15.9
Heptachlor	4.929	4.829	5.029	51.400	50.000	2.8
Heptachlor epoxide	5.689	5.591	5.791	54.280	50.000	8.6
Methoxychlor	7.491	7.393	7.593	42.560	50.000	-14.9
Tetrachloro-m-xylene	3.550	3.450	3.650	57.800	50.000	15.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL01 Date Analyzed: 06/27/2025

Lab Sample No.: PSTDCCC050 Data File : PD089202.D Time Analyzed: 21:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	54.200	50.000	8.4
4,4'-DDE	5.373	5.275	5.475	53.740	50.000	7.5
4,4'-DDT	6.181	6.084	6.284	41.310	50.000	-17.4
Aldrin	4.367	4.269	4.469	55.520	50.000	11.0
alpha-BHC	3.392	3.293	3.493	56.420	50.000	12.8
alpha-Chlordane	5.189	5.091	5.291	54.340	50.000	8.7
beta-BHC	4.024	3.926	4.126	54.120	50.000	8.2
Decachlorobiphenyl	8.070	7.972	8.172	50.420	50.000	0.8
delta-BHC	4.261	4.162	4.362	54.930	50.000	9.9
Dieldrin	5.511	5.413	5.613	53.700	50.000	7.4
Endosulfan I	5.245	5.147	5.347	54.090	50.000	8.2
Endosulfan II	6.079	5.981	6.181	53.260	50.000	6.5
Endosulfan sulfate	6.480	6.383	6.583	51.330	50.000	2.7
Endrin	5.787	5.689	5.889	50.370	50.000	0.7
Endrin aldehyde	6.256	6.159	6.359	49.550	50.000	-0.9
Endrin ketone	6.989	6.892	7.092	51.450	50.000	2.9
gamma-BHC (Lindane)	3.729	3.630	3.830	55.740	50.000	11.5
gamma-Chlordane	5.124	5.026	5.226	55.510	50.000	11.0
Heptachlor	4.082	3.983	4.183	50.310	50.000	0.6
Heptachlor epoxide	4.871	4.773	4.973	55.210	50.000	10.4
Methoxychlor	6.752	6.655	6.855	44.610	50.000	-10.8
Tetrachloro-m-xylene	2.881	2.780	2.980	58.420	50.000	16.8