

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD089214.D Time Analyzed: 00:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	57.920	50.000	15.8
4,4'-DDE	6.195	6.096	6.296	55.080	50.000	10.2
4,4'-DDT	7.019	6.921	7.121	45.370	50.000	-9.3
Aldrin	5.270	5.171	5.371	57.530	50.000	15.1
alpha-BHC	3.999	3.899	4.099	58.300	50.000	16.6
alpha-Chlordane	6.026	5.927	6.127	56.920	50.000	13.8
beta-BHC	4.515	4.415	4.615	54.940	50.000	9.9
Decachlorobiphenyl	9.070	8.972	9.172	51.410	50.000	2.8
delta-BHC	4.764	4.664	4.864	59.760	50.000	19.5
Dieldrin	6.346	6.247	6.447	55.900	50.000	11.8
Endosulfan I	6.073	5.975	6.175	55.830	50.000	11.7
Endosulfan II	6.785	6.686	6.886	55.060	50.000	10.1
Endosulfan sulfate	7.148	7.049	7.249	53.430	50.000	6.9
Endrin	6.573	6.475	6.675	51.730	50.000	3.5
Endrin aldehyde	6.913	6.815	7.015	53.120	50.000	6.2
Endrin ketone	7.628	7.530	7.730	54.640	50.000	9.3
gamma-BHC (Lindane)	4.330	4.230	4.430	57.210	50.000	14.4
gamma-Chlordane	5.945	5.846	6.046	58.500	50.000	17.0
Heptachlor	4.929	4.829	5.029	53.650	50.000	7.3
Heptachlor epoxide	5.690	5.591	5.791	55.310	50.000	10.6
Methoxychlor	7.491	7.393	7.593	48.170	50.000	-3.7
Tetrachloro-m-xylene	3.550	3.450	3.650	57.830	50.000	15.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL02 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD089214.D Time Analyzed: 00:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	54.820	50.000	9.6
4,4'-DDE	5.374	5.275	5.475	55.020	50.000	10.0
4,4'-DDT	6.182	6.084	6.284	45.730	50.000	-8.5
Aldrin	4.368	4.269	4.469	56.320	50.000	12.6
alpha-BHC	3.393	3.293	3.493	56.820	50.000	13.6
alpha-Chlordane	5.190	5.091	5.291	55.330	50.000	10.7
beta-BHC	4.025	3.926	4.126	54.570	50.000	9.1
Decachlorobiphenyl	8.071	7.972	8.172	53.550	50.000	7.1
delta-BHC	4.261	4.162	4.362	55.720	50.000	11.4
Dieldrin	5.512	5.413	5.613	55.070	50.000	10.1
Endosulfan I	5.246	5.147	5.347	55.100	50.000	10.2
Endosulfan II	6.078	5.981	6.181	54.780	50.000	9.6
Endosulfan sulfate	6.481	6.383	6.583	53.190	50.000	6.4
Endrin	5.788	5.689	5.889	52.970	50.000	5.9
Endrin aldehyde	6.258	6.159	6.359	52.380	50.000	4.8
Endrin ketone	6.990	6.892	7.092	54.180	50.000	8.4
gamma-BHC (Lindane)	3.729	3.630	3.830	56.410	50.000	12.8
gamma-Chlordane	5.125	5.026	5.226	56.400	50.000	12.8
Heptachlor	4.082	3.983	4.183	52.730	50.000	5.5
Heptachlor epoxide	4.872	4.773	4.973	56.350	50.000	12.7
Methoxychlor	6.753	6.655	6.855	50.280	50.000	0.6
Tetrachloro-m-xylene	2.881	2.780	2.980	58.840	50.000	17.7