

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD089226.D Time Analyzed: 03:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	58.230	50.000	16.5
4,4'-DDE	6.195	6.096	6.296	55.570	50.000	11.1
4,4'-DDT	7.020	6.921	7.121	47.730	50.000	-4.5
Aldrin	5.271	5.171	5.371	58.470	50.000	16.9
alpha-BHC	3.999	3.899	4.099	59.140	50.000	18.3
alpha-Chlordane	6.026	5.927	6.127	57.740	50.000	15.5
beta-BHC	4.515	4.415	4.615	55.540	50.000	11.1
Decachlorobiphenyl	9.071	8.972	9.172	52.610	50.000	5.2
delta-BHC	4.764	4.664	4.864	60.790	50.000	21.6
Dieldrin	6.346	6.247	6.447	56.690	50.000	13.4
Endosulfan I	6.074	5.975	6.175	56.550	50.000	13.1
Endosulfan II	6.785	6.686	6.886	55.400	50.000	10.8
Endosulfan sulfate	7.148	7.049	7.249	54.120	50.000	8.2
Endrin	6.573	6.475	6.675	52.870	50.000	5.7
Endrin aldehyde	6.914	6.815	7.015	54.050	50.000	8.1
Endrin ketone	7.629	7.530	7.730	55.310	50.000	10.6
gamma-BHC (Lindane)	4.330	4.230	4.430	58.070	50.000	16.1
gamma-Chlordane	5.945	5.846	6.046	59.180	50.000	18.4
Heptachlor	4.928	4.829	5.029	55.280	50.000	10.6
Heptachlor epoxide	5.690	5.591	5.791	55.990	50.000	12.0
Methoxychlor	7.492	7.393	7.593	51.020	50.000	2.0
Tetrachloro-m-xylene	3.550	3.450	3.650	58.680	50.000	17.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL03 Date Analyzed: 06/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD089226.D Time Analyzed: 03:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	55.100	50.000	10.2
4,4'-DDE	5.374	5.275	5.475	55.740	50.000	11.5
4,4'-DDT	6.182	6.084	6.284	47.680	50.000	-4.6
Aldrin	4.367	4.269	4.469	57.260	50.000	14.5
alpha-BHC	3.392	3.293	3.493	57.920	50.000	15.8
alpha-Chlordane	5.189	5.091	5.291	56.100	50.000	12.2
beta-BHC	4.024	3.926	4.126	55.320	50.000	10.6
Decachlorobiphenyl	8.071	7.972	8.172	54.260	50.000	8.5
delta-BHC	4.261	4.162	4.362	56.750	50.000	13.5
Dieldrin	5.511	5.413	5.613	55.900	50.000	11.8
Endosulfan I	5.245	5.147	5.347	55.680	50.000	11.4
Endosulfan II	6.079	5.981	6.181	55.070	50.000	10.1
Endosulfan sulfate	6.481	6.383	6.583	53.830	50.000	7.7
Endrin	5.788	5.689	5.889	54.690	50.000	9.4
Endrin aldehyde	6.257	6.159	6.359	53.180	50.000	6.4
Endrin ketone	6.989	6.892	7.092	54.840	50.000	9.7
gamma-BHC (Lindane)	3.729	3.630	3.830	57.410	50.000	14.8
gamma-Chlordane	5.124	5.026	5.226	57.110	50.000	14.2
Heptachlor	4.082	3.983	4.183	54.310	50.000	8.6
Heptachlor epoxide	4.871	4.773	4.973	57.050	50.000	14.1
Methoxychlor	6.753	6.655	6.855	52.720	50.000	5.4
Tetrachloro-m-xylene	2.881	2.780	2.980	59.590	50.000	19.2