

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD089230.D Time Analyzed: 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.605	6.805	53.380	50.000	6.8
4,4'-DDE	6.201	6.096	6.296	50.960	50.000	1.9
4,4'-DDT	7.026	6.921	7.121	42.730	50.000	-14.5
Aldrin	5.277	5.171	5.371	52.490	50.000	5.0
alpha-BHC	4.005	3.899	4.099	53.210	50.000	6.4
alpha-Chlordane	6.032	5.927	6.127	51.250	50.000	2.5
beta-BHC	4.521	4.415	4.615	50.740	50.000	1.5
Decachlorobiphenyl	9.078	8.972	9.172	48.640	50.000	-2.7
delta-BHC	4.770	4.664	4.864	54.000	50.000	8.0
Dieldrin	6.352	6.247	6.447	51.790	50.000	3.6
Endosulfan I	6.080	5.975	6.175	51.350	50.000	2.7
Endosulfan II	6.791	6.686	6.886	51.290	50.000	2.6
Endosulfan sulfate	7.155	7.049	7.249	49.650	50.000	-0.7
Endrin	6.580	6.475	6.675	46.870	50.000	-6.3
Endrin aldehyde	6.920	6.815	7.015	49.280	50.000	-1.4
Endrin ketone	7.634	7.530	7.730	51.330	50.000	2.7
gamma-BHC (Lindane)	4.336	4.230	4.430	52.320	50.000	4.6
gamma-Chlordane	5.951	5.846	6.046	52.780	50.000	5.6
Heptachlor	4.935	4.829	5.029	50.190	50.000	0.4
Heptachlor epoxide	5.696	5.591	5.791	50.880	50.000	1.8
Methoxychlor	7.498	7.393	7.593	57.870	50.000	15.7
Tetrachloro-m-xylene	3.555	3.450	3.650	52.550	50.000	5.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL04 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD089230.D Time Analyzed: 09:56

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.930	5.830	6.030	49.690	50.000	-0.6
4,4'-DDE	5.376	5.275	5.475	49.570	50.000	-0.9
4,4'-DDT	6.184	6.084	6.284	41.990	50.000	-16.0
Aldrin	4.369	4.269	4.469	50.790	50.000	1.6
alpha-BHC	3.393	3.293	3.493	51.090	50.000	2.2
alpha-Chlordane	5.191	5.091	5.291	49.510	50.000	-1.0
beta-BHC	4.026	3.926	4.126	49.850	50.000	-0.3
Decachlorobiphenyl	8.073	7.972	8.172	49.620	50.000	-0.8
delta-BHC	4.263	4.162	4.362	49.960	50.000	-0.1
Dieldrin	5.513	5.413	5.613	49.540	50.000	-0.9
Endosulfan I	5.247	5.147	5.347	49.140	50.000	-1.7
Endosulfan II	6.081	5.981	6.181	49.360	50.000	-1.3
Endosulfan sulfate	6.483	6.383	6.583	48.200	50.000	-3.6
Endrin	5.790	5.689	5.889	44.720	50.000	-10.6
Endrin aldehyde	6.260	6.159	6.359	48.270	50.000	-3.5
Endrin ketone	6.992	6.892	7.092	49.890	50.000	-0.2
gamma-BHC (Lindane)	3.730	3.630	3.830	50.660	50.000	1.3
gamma-Chlordane	5.126	5.026	5.226	49.650	50.000	-0.7
Heptachlor	4.083	3.983	4.183	47.330	50.000	-5.3
Heptachlor epoxide	4.873	4.773	4.973	50.950	50.000	1.9
Methoxychlor	6.755	6.655	6.855	55.470	50.000	10.9
Tetrachloro-m-xylene	2.880	2.780	2.980	52.470	50.000	4.9