

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL05 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD089241.D Time Analyzed: 17:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.605	6.805	52.240	50.000	4.5
4,4'-DDE	6.202	6.096	6.296	50.680	50.000	1.4
4,4'-DDT	7.027	6.921	7.121	42.510	50.000	-15.0
Aldrin	5.277	5.171	5.371	53.100	50.000	6.2
alpha-BHC	4.005	3.899	4.099	54.360	50.000	8.7
alpha-Chlordane	6.033	5.927	6.127	51.160	50.000	2.3
beta-BHC	4.522	4.415	4.615	51.930	50.000	3.9
Decachlorobiphenyl	9.079	8.972	9.172	48.610	50.000	-2.8
delta-BHC	4.771	4.664	4.864	54.890	50.000	9.8
Dieldrin	6.353	6.247	6.447	51.080	50.000	2.2
Endosulfan I	6.080	5.975	6.175	51.360	50.000	2.7
Endosulfan II	6.792	6.686	6.886	50.340	50.000	0.7
Endosulfan sulfate	7.155	7.049	7.249	48.940	50.000	-2.1
Endrin	6.580	6.475	6.675	47.410	50.000	-5.2
Endrin aldehyde	6.921	6.815	7.015	48.190	50.000	-3.6
Endrin ketone	7.636	7.530	7.730	48.760	50.000	-2.5
gamma-BHC (Lindane)	4.336	4.230	4.430	54.200	50.000	8.4
gamma-Chlordane	5.952	5.846	6.046	52.240	50.000	4.5
Heptachlor	4.936	4.829	5.029	51.190	50.000	2.4
Heptachlor epoxide	5.697	5.591	5.791	50.920	50.000	1.8
Methoxychlor	7.498	7.393	7.593	57.540	50.000	15.1
Tetrachloro-m-xylene	3.556	3.450	3.650	53.550	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/17/2025 06/17/2025

Client Sample No.: CCAL05 Date Analyzed: 06/30/2025

Lab Sample No.: PSTDCCC050 Data File : PD089241.D Time Analyzed: 17:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.931	5.830	6.030	49.480	50.000	-1.0
4,4'-DDE	5.376	5.275	5.475	49.650	50.000	-0.7
4,4'-DDT	6.184	6.084	6.284	42.410	50.000	-15.2
Aldrin	4.369	4.269	4.469	51.750	50.000	3.5
alpha-BHC	3.393	3.293	3.493	52.530	50.000	5.1
alpha-Chlordane	5.191	5.091	5.291	49.660	50.000	-0.7
beta-BHC	4.025	3.926	4.126	50.600	50.000	1.2
Decachlorobiphenyl	8.073	7.972	8.172	45.140	50.000	-9.7
delta-BHC	4.262	4.162	4.362	51.290	50.000	2.6
Dieldrin	5.514	5.413	5.613	49.820	50.000	-0.4
Endosulfan I	5.247	5.147	5.347	49.320	50.000	-1.4
Endosulfan II	6.081	5.981	6.181	48.420	50.000	-3.2
Endosulfan sulfate	6.483	6.383	6.583	47.530	50.000	-4.9
Endrin	5.790	5.689	5.889	46.410	50.000	-7.2
Endrin aldehyde	6.260	6.159	6.359	46.990	50.000	-6.0
Endrin ketone	6.992	6.892	7.092	46.930	50.000	-6.1
gamma-BHC (Lindane)	3.729	3.630	3.830	52.020	50.000	4.0
gamma-Chlordane	5.127	5.026	5.226	49.970	50.000	-0.1
Heptachlor	4.083	3.983	4.183	49.310	50.000	-1.4
Heptachlor epoxide	4.873	4.773	4.973	51.760	50.000	3.5
Methoxychlor	6.755	6.655	6.855	56.930	50.000	13.9
Tetrachloro-m-xylene	2.880	2.780	2.980	53.540	50.000	7.1