

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062825\
 Data File : PD089226.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jun 2025 03:12
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 03:51:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 18 05:39:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.550	2.881	167.4E6	1010.3E6	58.684	59.594
28)	SA Decachlor...	9.071	8.071	206.6E6	1072.7E6	52.612	54.261
Target Compounds							
2)	A alpha-BHC	3.999	3.392	344.7E6	1552.2E6	59.140	57.925
3)	MA gamma-BHC...	4.330	3.729	324.1E6	1421.4E6	58.073	57.415
4)	MA Heptachlor	4.928	4.082	301.4E6	1359.9E6	55.275	54.314
5)	MB Aldrin	5.271	4.367	310.6E6	1386.7E6	58.470	57.262
6)	B beta-BHC	4.515	4.024	122.6E6	594.4E6	55.545	55.316
7)	B delta-BHC	4.764	4.261	312.0E6	1411.5E6	60.788	56.747
8)	B Heptachlo...	5.690	4.871	270.1E6	1249.2E6	55.987	57.045
9)	A Endosulfan I	6.074	5.245	254.9E6	1160.4E6	56.551	55.677
10)	B gamma-Chl...	5.945	5.124	283.3E6	1366.3E6	59.177	57.112
11)	B alpha-Chl...	6.026	5.189	279.4E6	1284.4E6	57.736	56.097
12)	B 4,4'-DDE	6.195	5.374	242.4E6	1277.4E6	55.572	55.742
13)	MA Dieldrin	6.346	5.511	272.0E6	1297.3E6	56.693	55.899
14)	MA Endrin	6.573	5.788	218.1E6	1187.4E6	52.865	54.685
15)	B Endosulfa...	6.785	6.079	223.6E6	1112.3E6	55.397	55.075
16)	A 4,4'-DDD	6.704	5.928	199.9E6	1054.1E6	58.232	55.101
17)	MA 4,4'-DDT	7.020	6.182	180.8E6	966.9E6	47.734	47.677
18)	B Endrin al...	6.914	6.257	165.6E6	814.3E6	54.049	53.180
19)	B Endosulfa...	7.148	6.481	207.7E6	1056.1E6	54.123	53.825
20)	A Methoxychlor	7.492	6.753	103.1E6	567.6E6	51.019	52.716
21)	B Endrin ke...	7.629	6.989	225.3E6	1184.5E6	55.314	54.835
22)	Mirex	8.112	7.183	162.3E6	887.9E6	53.401	53.236

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD062825\
Data File : PD089226.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jun 2025 03:12
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 03:51:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD061825.M
Quant Title : GC Extractables
QLast Update : Wed Jun 18 05:39:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

