

LAB CHRONICLE

OrderID:	Q2436	OrderDate:	6/26/2025 3:41:00 PM
Client:	CDM Smith	Project:	South River WM Replacement
Contact:	Marcie Ann Encinas	Location:	D51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2436-01	TP-70	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/27/25	06/26/25
Q2436-02	TP-69	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/27/25	06/26/25
Q2436-03	TP-85	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/30/25	06/26/25
Q2436-04	TP-86	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/30/25	06/26/25
Q2436-05	TP-84	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/30/25	06/26/25
Q2436-06	TP-83	SOIL	VOC-TCLVOA-10	8260D	06/25/25		06/27/25	06/26/25
Q2436-07	TP-87	SOIL	VOC-TCLVOA-10	8260D	06/26/25		06/27/25	06/26/25
Q2436-08	TP-100	SOIL	VOC-TCLVOA-10	8260D	06/26/25		06/27/25	06/26/25
Q2436-09	TP-99	SOIL	VOC-TCLVOA-10	8260D	06/26/25		06/27/25	06/26/25
Q2436-10	TP-82	SOIL	VOC-TCLVOA-10	8260D	06/26/25		06/27/25	06/26/25

Hit Summary Sheet
SW-846

SDG No.: Q2436
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	TP-70							
Q2436-01	TP-70	SOIL	Acetone	18.5	J	4.10	21.5	ug/Kg
Q2436-01	TP-70	SOIL	Methylene Chloride	4.10	J	3.00	8.60	ug/Kg
			Total Voc :	22.6				
Q2436-01	TP-70	SOIL	Hexane, 3-methyl-	* 4.70	J	0	0	ug/Kg
Q2436-01	TP-70	SOIL	Heptane, 3-methyl-	* 4.40	J	0	0	ug/Kg
Q2436-01	TP-70	SOIL	2-Ethyl-oxetane	* 4.80	J	0	0	ug/Kg
			Total Tics :	13.9				
			Total Concentration:	36.5				
Client ID:	TP-69							
Q2436-02	TP-69	SOIL	Acetone	77.9		4.40	23.4	ug/Kg
Q2436-02	TP-69	SOIL	Methylene Chloride	5.90	J	3.30	9.40	ug/Kg
Q2436-02	TP-69	SOIL	2-Butanone	9.80	J	6.10	23.4	ug/Kg
			Total Voc :	93.6				
			Total Concentration:	93.6				
Client ID:	TP-85							
Q2436-03	TP-85	SOIL	Acetone	33.3		4.10	21.7	ug/Kg
Q2436-03	TP-85	SOIL	Carbon Disulfide	0.99	J	0.92	4.30	ug/Kg
Q2436-03	TP-85	SOIL	Methylene Chloride	4.90	J	3.10	8.70	ug/Kg
			Total Voc :	39.2				
			Total Concentration:	39.2				
Client ID:	TP-86							
Q2436-04	TP-86	SOIL	Methylene Chloride	6.10	J	3.30	9.30	ug/Kg
			Total Voc :	6.10				
Q2436-04	TP-86	SOIL	Diethyl Ether	* 1.80	J	0.72	4.70	ug/Kg
			Total Tics :	1.80				
			Total Concentration:	7.90				
Client ID:	TP-84							
Q2436-05	TP-84	SOIL	Methylene Chloride	4.10	J	3.10	8.80	ug/Kg
			Total Voc :	4.10				
			Total Concentration:	4.10				
Client ID:	TP-100							
Q2436-08	TP-100	SOIL	Methylene Chloride	3.40	J	2.90	8.20	ug/Kg
			Total Voc :	3.40				
			Total Concentration:	3.40				
Client ID:	TP-82							
Q2436-10	TP-82	SOIL	Methylene Chloride	3.60	J	3.30	9.50	ug/Kg
			Total Voc :	3.60				
			Total Concentration:	3.60				



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-70	SDG No.:	Q2436
Lab Sample ID:	Q2436-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82.1
Sample Wt/Vol:	7.07 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022856.D	1		06/27/25 12:59	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.98	U	0.98	4.30	ug/Kg
74-87-3	Chloromethane	0.98	U	0.98	4.30	ug/Kg
75-01-4	Vinyl Chloride	0.68	U	0.68	4.30	ug/Kg
74-83-9	Bromomethane	0.92	U	0.92	4.30	ug/Kg
75-00-3	Chloroethane	1.10	U	1.10	4.30	ug/Kg
75-69-4	Trichlorofluoromethane	1.00	U	1.00	4.30	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.91	U	0.91	4.30	ug/Kg
75-35-4	1,1-Dichloroethene	0.86	U	0.86	4.30	ug/Kg
67-64-1	Acetone	18.5	J	4.10	21.5	ug/Kg
75-15-0	Carbon Disulfide	0.91	U	0.91	4.30	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.63	U	0.63	4.30	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.30	ug/Kg
75-09-2	Methylene Chloride	4.10	J	3.00	8.60	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.74	U	0.74	4.30	ug/Kg
75-34-3	1,1-Dichloroethane	0.69	U	0.69	4.30	ug/Kg
110-82-7	Cyclohexane	0.68	U	0.68	4.30	ug/Kg
78-93-3	2-Butanone	5.60	U	5.60	21.5	ug/Kg
56-23-5	Carbon Tetrachloride	0.84	U	0.84	4.30	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.65	U	0.65	4.30	ug/Kg
74-97-5	Bromochloromethane	0.99	U	0.99	4.30	ug/Kg
67-66-3	Chloroform	0.72	U	0.72	4.30	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.80	U	0.80	4.30	ug/Kg
108-87-2	Methylcyclohexane	0.78	U	0.78	4.30	ug/Kg
71-43-2	Benzene	0.68	U	0.68	4.30	ug/Kg
107-06-2	1,2-Dichloroethane	0.68	U	0.68	4.30	ug/Kg
79-01-6	Trichloroethene	0.70	U	0.70	4.30	ug/Kg
78-87-5	1,2-Dichloropropane	0.78	U	0.78	4.30	ug/Kg
75-27-4	Bromodichloromethane	0.67	U	0.67	4.30	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.10	U	3.10	21.5	ug/Kg
108-88-3	Toluene	0.67	U	0.67	4.30	ug/Kg

A
B
C
D
E
F
G

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022856.D	1		06/27/25 12:59	VY062725

[illegible]

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-70	SDG No.:	Q2436
Lab Sample ID:	Q2436-01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82.1
Sample Wt/Vol:	7.07 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022856.D	1		06/27/25 12:59	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
1010386-40-2	2-Ethyl-oxetane	4.80	J		5.65	ug/Kg
000589-34-4	Hexane, 3-methyl-	4.70	J		7.92	ug/Kg
000589-81-1	Heptane, 3-methyl-	4.40	J		9.89	ug/Kg

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-69	SDG No.:	Q2436
Lab Sample ID:	Q2436-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82
Sample Wt/Vol:	6.51 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022857.D	1		06/27/25 13:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	4.70	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	4.70	ug/Kg
75-01-4	Vinyl Chloride	0.74	U	0.74	4.70	ug/Kg
74-83-9	Bromomethane	1.00	U	1.00	4.70	ug/Kg
75-00-3	Chloroethane	1.20	U	1.20	4.70	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.70	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.99	U	0.99	4.70	ug/Kg
75-35-4	1,1-Dichloroethene	0.94	U	0.94	4.70	ug/Kg
67-64-1	Acetone	77.9		4.40	23.4	ug/Kg
75-15-0	Carbon Disulfide	0.99	U	0.99	4.70	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.68	U	0.68	4.70	ug/Kg
79-20-9	Methyl Acetate	1.40	U	1.40	4.70	ug/Kg
75-09-2	Methylene Chloride	5.90	J	3.30	9.40	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.81	U	0.81	4.70	ug/Kg
75-34-3	1,1-Dichloroethane	0.75	U	0.75	4.70	ug/Kg
110-82-7	Cyclohexane	0.74	U	0.74	4.70	ug/Kg
78-93-3	2-Butanone	9.80	J	6.10	23.4	ug/Kg
56-23-5	Carbon Tetrachloride	0.91	U	0.91	4.70	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.70	U	0.70	4.70	ug/Kg
74-97-5	Bromochloromethane	1.10	U	1.10	4.70	ug/Kg
67-66-3	Chloroform	0.79	U	0.79	4.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.87	U	0.87	4.70	ug/Kg
108-87-2	Methylcyclohexane	0.85	U	0.85	4.70	ug/Kg
71-43-2	Benzene	0.74	U	0.74	4.70	ug/Kg
107-06-2	1,2-Dichloroethane	0.74	U	0.74	4.70	ug/Kg
79-01-6	Trichloroethene	0.76	U	0.76	4.70	ug/Kg
78-87-5	1,2-Dichloropropane	0.85	U	0.85	4.70	ug/Kg
75-27-4	Bromodichloromethane	0.73	U	0.73	4.70	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.40	U	3.40	23.4	ug/Kg
108-88-3	Toluene	0.73	U	0.73	4.70	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-69	SDG No.:	Q2436
Lab Sample ID:	Q2436-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82
Sample Wt/Vol:	6.51 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022857.D	1		06/27/25 13:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.61	U	0.61	4.70	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.58	U	0.58	4.70	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.86	U	0.86	4.70	ug/Kg
591-78-6	2-Hexanone	3.50	U	3.50	23.4	ug/Kg
124-48-1	Dibromochloromethane	0.81	U	0.81	4.70	ug/Kg
106-93-4	1,2-Dibromoethane	0.82	U	0.82	4.70	ug/Kg
127-18-4	Tetrachloroethene	0.98	U	0.98	4.70	ug/Kg
108-90-7	Chlorobenzene	0.85	U	0.85	4.70	ug/Kg
100-41-4	Ethyl Benzene	0.63	U	0.63	4.70	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	9.40	ug/Kg
95-47-6	o-Xylene	0.77	U	0.77	4.70	ug/Kg
100-42-5	Styrene	0.67	U	0.67	4.70	ug/Kg
75-25-2	Bromoform	0.81	U	0.81	4.70	ug/Kg
98-82-8	Isopropylbenzene	0.73	U	0.73	4.70	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.70	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.60	U	1.60	4.70	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.50	U	1.50	4.70	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.40	U	1.40	4.70	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.70	U	1.70	4.70	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.80	U	2.80	4.70	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.00	U	3.00	4.70	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.8		63 - 155	94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	49.5		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		17 - 146	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	341000	7.713			
540-36-3	1,4-Difluorobenzene	626000	8.622			
3114-55-4	Chlorobenzene-d5	598000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	255000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-69	SDG No.:	Q2436
Lab Sample ID:	Q2436-02	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	82
Sample Wt/Vol:	6.51 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022857.D	1		06/27/25 13:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-85	SDG No.:	Q2436
Lab Sample ID:	Q2436-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.1
Sample Wt/Vol:	6.77 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022877.D	1		06/30/25 13:32	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.99	U	0.99	4.30	ug/Kg
74-87-3	Chloromethane	0.99	U	0.99	4.30	ug/Kg
75-01-4	Vinyl Chloride	0.69	U	0.69	4.30	ug/Kg
74-83-9	Bromomethane	0.93	U	0.93	4.30	ug/Kg
75-00-3	Chloroethane	1.10	U	1.10	4.30	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.30	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.92	U	0.92	4.30	ug/Kg
75-35-4	1,1-Dichloroethene	0.87	U	0.87	4.30	ug/Kg
67-64-1	Acetone	33.3		4.10	21.7	ug/Kg
75-15-0	Carbon Disulfide	0.99	J	0.92	4.30	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.63	U	0.63	4.30	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.30	ug/Kg
75-09-2	Methylene Chloride	4.90	J	3.10	8.70	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.75	U	0.75	4.30	ug/Kg
75-34-3	1,1-Dichloroethane	0.69	U	0.69	4.30	ug/Kg
110-82-7	Cyclohexane	0.69	U	0.69	4.30	ug/Kg
78-93-3	2-Butanone	5.70	U	5.70	21.7	ug/Kg
56-23-5	Carbon Tetrachloride	0.84	U	0.84	4.30	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.65	U	0.65	4.30	ug/Kg
74-97-5	Bromochloromethane	1.00	U	1.00	4.30	ug/Kg
67-66-3	Chloroform	0.73	U	0.73	4.30	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.81	U	0.81	4.30	ug/Kg
108-87-2	Methylcyclohexane	0.79	U	0.79	4.30	ug/Kg
71-43-2	Benzene	0.69	U	0.69	4.30	ug/Kg
107-06-2	1,2-Dichloroethane	0.69	U	0.69	4.30	ug/Kg
79-01-6	Trichloroethene	0.70	U	0.70	4.30	ug/Kg
78-87-5	1,2-Dichloropropane	0.79	U	0.79	4.30	ug/Kg
75-27-4	Bromodichloromethane	0.68	U	0.68	4.30	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.10	U	3.10	21.7	ug/Kg
108-88-3	Toluene	0.68	U	0.68	4.30	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-85	SDG No.:	Q2436
Lab Sample ID:	Q2436-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.1
Sample Wt/Vol:	6.77 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022877.D	1		06/30/25 13:32	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.56	U	0.56	4.30	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.54	U	0.54	4.30	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.80	U	0.80	4.30	ug/Kg
591-78-6	2-Hexanone	3.20	U	3.20	21.7	ug/Kg
124-48-1	Dibromochloromethane	0.76	U	0.76	4.30	ug/Kg
106-93-4	1,2-Dibromoethane	0.76	U	0.76	4.30	ug/Kg
127-18-4	Tetrachloroethene	0.91	U	0.91	4.30	ug/Kg
108-90-7	Chlorobenzene	0.79	U	0.79	4.30	ug/Kg
100-41-4	Ethyl Benzene	0.58	U	0.58	4.30	ug/Kg
179601-23-1	m/p-Xylenes	1.10	U	1.10	8.70	ug/Kg
95-47-6	o-Xylene	0.71	U	0.71	4.30	ug/Kg
100-42-5	Styrene	0.62	U	0.62	4.30	ug/Kg
75-25-2	Bromoform	0.75	U	0.75	4.30	ug/Kg
98-82-8	Isopropylbenzene	0.68	U	0.68	4.30	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.30	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.50	U	1.50	4.30	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.40	U	1.40	4.30	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.30	U	1.30	4.30	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.60	U	1.60	4.30	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.60	U	2.60	4.30	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.80	U	2.80	4.30	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	45.5		63 - 155	91%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	49.9		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		17 - 146	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	373000	7.707			
540-36-3	1,4-Difluorobenzene	691000	8.61			
3114-55-4	Chlorobenzene-d5	658000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	284000	13.34			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-85	SDG No.:	Q2436
Lab Sample ID:	Q2436-03	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.1
Sample Wt/Vol:	6.77	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022877.D	1		06/30/25 13:32	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-86	SDG No.:	Q2436
Lab Sample ID:	Q2436-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.7
Sample Wt/Vol:	6.03 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022878.D	1		06/30/25 14:28	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	4.70	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	4.70	ug/Kg
75-01-4	Vinyl Chloride	0.74	U	0.74	4.70	ug/Kg
74-83-9	Bromomethane	1.00	U	1.00	4.70	ug/Kg
75-00-3	Chloroethane	1.20	U	1.20	4.70	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.70	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.99	U	0.99	4.70	ug/Kg
75-35-4	1,1-Dichloroethene	0.93	U	0.93	4.70	ug/Kg
67-64-1	Acetone	4.40	U	4.40	23.4	ug/Kg
75-15-0	Carbon Disulfide	0.99	U	0.99	4.70	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.68	U	0.68	4.70	ug/Kg
79-20-9	Methyl Acetate	1.40	U	1.40	4.70	ug/Kg
75-09-2	Methylene Chloride	6.10	J	3.30	9.30	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.80	U	0.80	4.70	ug/Kg
75-34-3	1,1-Dichloroethane	0.75	U	0.75	4.70	ug/Kg
110-82-7	Cyclohexane	0.74	U	0.74	4.70	ug/Kg
78-93-3	2-Butanone	6.10	U	6.10	23.4	ug/Kg
56-23-5	Carbon Tetrachloride	0.91	U	0.91	4.70	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.70	U	0.70	4.70	ug/Kg
74-97-5	Bromochloromethane	1.10	U	1.10	4.70	ug/Kg
67-66-3	Chloroform	0.79	U	0.79	4.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.87	U	0.87	4.70	ug/Kg
108-87-2	Methylcyclohexane	0.85	U	0.85	4.70	ug/Kg
71-43-2	Benzene	0.74	U	0.74	4.70	ug/Kg
107-06-2	1,2-Dichloroethane	0.74	U	0.74	4.70	ug/Kg
79-01-6	Trichloroethene	0.76	U	0.76	4.70	ug/Kg
78-87-5	1,2-Dichloropropane	0.85	U	0.85	4.70	ug/Kg
75-27-4	Bromodichloromethane	0.73	U	0.73	4.70	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.30	U	3.30	23.4	ug/Kg
108-88-3	Toluene	0.73	U	0.73	4.70	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-86	SDG No.:	Q2436
Lab Sample ID:	Q2436-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.7
Sample Wt/Vol:	6.03	Units:	g
Soil Aliquot Vol:		Final Vol:	5000 uL
		Test:	VOC-TCLVOA-10
GC Column:	RXI-624	Level :	LOW
Prep Method :	ID : 0.25		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022878.D	1		06/30/25 14:28	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.61	U	0.61	4.70	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.58	U	0.58	4.70	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.86	U	0.86	4.70	ug/Kg
591-78-6	2-Hexanone	3.40	U	3.40	23.4	ug/Kg
124-48-1	Dibromochloromethane	0.81	U	0.81	4.70	ug/Kg
106-93-4	1,2-Dibromoethane	0.82	U	0.82	4.70	ug/Kg
127-18-4	Tetrachloroethene	0.98	U	0.98	4.70	ug/Kg
108-90-7	Chlorobenzene	0.85	U	0.85	4.70	ug/Kg
100-41-4	Ethyl Benzene	0.63	U	0.63	4.70	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	9.30	ug/Kg
95-47-6	o-Xylene	0.77	U	0.77	4.70	ug/Kg
100-42-5	Styrene	0.66	U	0.66	4.70	ug/Kg
75-25-2	Bromoform	0.80	U	0.80	4.70	ug/Kg
98-82-8	Isopropylbenzene	0.73	U	0.73	4.70	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.70	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.60	U	1.60	4.70	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.50	U	1.50	4.70	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.40	U	1.40	4.70	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.70	U	1.70	4.70	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.80	U	2.80	4.70	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.00	U	3.00	4.70	ug/Kg

SURROGATES

17060-07-0	1,2-Dichloroethane-d4	48.2	63 - 155	96%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7	70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	50.2	74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.9	17 - 146	116%	SPK: 50

INTERNAL STANDARDS

363-72-4	Pentafluorobenzene	346000	7.707
540-36-3	1,4-Difluorobenzene	640000	8.616
3114-55-4	Chlorobenzene-d5	634000	11.414
3855-82-1	1,4-Dichlorobenzene-d4	290000	13.347

TENTATIVE IDENTIFIED COMPOUNDS

A
B
C
D
E
F
G

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-86	SDG No.:	Q2436
Lab Sample ID:	Q2436-04	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	88.7
Sample Wt/Vol:	6.03 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022878.D	1		06/30/25 14:28	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
60-29-7	Diethyl Ether	1.80	J		3.45	ug/Kg

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-84	SDG No.:	Q2436
Lab Sample ID:	Q2436-05	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.3
Sample Wt/Vol:	6.13 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022879.D	1		06/30/25 14:52	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.00	U	1.00	4.40	ug/Kg
74-87-3	Chloromethane	1.00	U	1.00	4.40	ug/Kg
75-01-4	Vinyl Chloride	0.70	U	0.70	4.40	ug/Kg
74-83-9	Bromomethane	0.95	U	0.95	4.40	ug/Kg
75-00-3	Chloroethane	1.10	U	1.10	4.40	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.40	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.94	U	0.94	4.40	ug/Kg
75-35-4	1,1-Dichloroethene	0.88	U	0.88	4.40	ug/Kg
67-64-1	Acetone	4.20	U	4.20	22.1	ug/Kg
75-15-0	Carbon Disulfide	0.94	U	0.94	4.40	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.65	U	0.65	4.40	ug/Kg
79-20-9	Methyl Acetate	1.40	U	1.40	4.40	ug/Kg
75-09-2	Methylene Chloride	4.10	J	3.10	8.80	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.76	U	0.76	4.40	ug/Kg
75-34-3	1,1-Dichloroethane	0.71	U	0.71	4.40	ug/Kg
110-82-7	Cyclohexane	0.70	U	0.70	4.40	ug/Kg
78-93-3	2-Butanone	5.80	U	5.80	22.1	ug/Kg
56-23-5	Carbon Tetrachloride	0.86	U	0.86	4.40	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.66	U	0.66	4.40	ug/Kg
74-97-5	Bromochloromethane	1.00	U	1.00	4.40	ug/Kg
67-66-3	Chloroform	0.74	U	0.74	4.40	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.82	U	0.82	4.40	ug/Kg
108-87-2	Methylcyclohexane	0.80	U	0.80	4.40	ug/Kg
71-43-2	Benzene	0.70	U	0.70	4.40	ug/Kg
107-06-2	1,2-Dichloroethane	0.70	U	0.70	4.40	ug/Kg
79-01-6	Trichloroethene	0.72	U	0.72	4.40	ug/Kg
78-87-5	1,2-Dichloropropane	0.80	U	0.80	4.40	ug/Kg
75-27-4	Bromodichloromethane	0.69	U	0.69	4.40	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.20	U	3.20	22.1	ug/Kg
108-88-3	Toluene	0.69	U	0.69	4.40	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-84	SDG No.:	Q2436
Lab Sample ID:	Q2436-05	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.3
Sample Wt/Vol:	6.13 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022879.D	1		06/30/25 14:52	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.57	U	0.57	4.40	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.55	U	0.55	4.40	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.81	U	0.81	4.40	ug/Kg
591-78-6	2-Hexanone	3.30	U	3.30	22.1	ug/Kg
124-48-1	Dibromochloromethane	0.77	U	0.77	4.40	ug/Kg
106-93-4	1,2-Dibromoethane	0.78	U	0.78	4.40	ug/Kg
127-18-4	Tetrachloroethene	0.93	U	0.93	4.40	ug/Kg
108-90-7	Chlorobenzene	0.80	U	0.80	4.40	ug/Kg
100-41-4	Ethyl Benzene	0.59	U	0.59	4.40	ug/Kg
179601-23-1	m/p-Xylenes	1.10	U	1.10	8.80	ug/Kg
95-47-6	o-Xylene	0.72	U	0.72	4.40	ug/Kg
100-42-5	Styrene	0.63	U	0.63	4.40	ug/Kg
75-25-2	Bromoform	0.76	U	0.76	4.40	ug/Kg
98-82-8	Isopropylbenzene	0.69	U	0.69	4.40	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.40	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.50	U	1.50	4.40	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.40	U	1.40	4.40	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.30	U	1.30	4.40	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.60	U	1.60	4.40	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.60	U	2.60	4.40	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.80	U	2.80	4.40	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.9		63 - 155	100%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	50.6		74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.6		17 - 146	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	324000	7.707			
540-36-3	1,4-Difluorobenzene	612000	8.61			
3114-55-4	Chlorobenzene-d5	599000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	250000	13.34			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-84	SDG No.:	Q2436
Lab Sample ID:	Q2436-05	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.3
Sample Wt/Vol:	6.13	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022879.D	1		06/30/25 14:52	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-83	SDG No.:	Q2436
Lab Sample ID:	Q2436-06	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.6
Sample Wt/Vol:	5.19 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022861.D	1		06/27/25 14:56	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.20	U	1.20	5.30	ug/Kg
74-87-3	Chloromethane	1.20	U	1.20	5.30	ug/Kg
75-01-4	Vinyl Chloride	0.84	U	0.84	5.30	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.30	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.30	ug/Kg
75-69-4	Trichlorofluoromethane	1.30	U	1.30	5.30	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.30	ug/Kg
75-35-4	1,1-Dichloroethene	1.10	U	1.10	5.30	ug/Kg
67-64-1	Acetone	5.00	U	5.00	26.6	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.30	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.78	U	0.78	5.30	ug/Kg
79-20-9	Methyl Acetate	1.60	U	1.60	5.30	ug/Kg
75-09-2	Methylene Chloride	3.80	U	3.80	10.6	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.91	U	0.91	5.30	ug/Kg
75-34-3	1,1-Dichloroethane	0.85	U	0.85	5.30	ug/Kg
110-82-7	Cyclohexane	0.84	U	0.84	5.30	ug/Kg
78-93-3	2-Butanone	7.00	U	7.00	26.6	ug/Kg
56-23-5	Carbon Tetrachloride	1.00	U	1.00	5.30	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.80	U	0.80	5.30	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.30	ug/Kg
67-66-3	Chloroform	0.89	U	0.89	5.30	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.99	U	0.99	5.30	ug/Kg
108-87-2	Methylcyclohexane	0.97	U	0.97	5.30	ug/Kg
71-43-2	Benzene	0.84	U	0.84	5.30	ug/Kg
107-06-2	1,2-Dichloroethane	0.84	U	0.84	5.30	ug/Kg
79-01-6	Trichloroethene	0.86	U	0.86	5.30	ug/Kg
78-87-5	1,2-Dichloropropane	0.97	U	0.97	5.30	ug/Kg
75-27-4	Bromodichloromethane	0.83	U	0.83	5.30	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.80	U	3.80	26.6	ug/Kg
108-88-3	Toluene	0.83	U	0.83	5.30	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-83	SDG No.:	Q2436
Lab Sample ID:	Q2436-06	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.6
Sample Wt/Vol:	5.19 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022861.D	1		06/27/25 14:56	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.69	U	0.69	5.30	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.66	U	0.66	5.30	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.98	U	0.98	5.30	ug/Kg
591-78-6	2-Hexanone	3.90	U	3.90	26.6	ug/Kg
124-48-1	Dibromochloromethane	0.93	U	0.93	5.30	ug/Kg
106-93-4	1,2-Dibromoethane	0.94	U	0.94	5.30	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.30	ug/Kg
108-90-7	Chlorobenzene	0.97	U	0.97	5.30	ug/Kg
100-41-4	Ethyl Benzene	0.71	U	0.71	5.30	ug/Kg
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.6	ug/Kg
95-47-6	o-Xylene	0.87	U	0.87	5.30	ug/Kg
100-42-5	Styrene	0.75	U	0.75	5.30	ug/Kg
75-25-2	Bromoform	0.91	U	0.91	5.30	ug/Kg
98-82-8	Isopropylbenzene	0.83	U	0.83	5.30	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.30	U	1.30	5.30	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.80	U	1.80	5.30	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.70	U	1.70	5.30	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.30	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	2.00	U	2.00	5.30	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.20	U	3.20	5.30	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.40	U	3.40	5.30	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.5		63 - 155	93%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		70 - 134	99%	SPK: 50
2037-26-5	Toluene-d8	49.6		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		17 - 146	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	337000	7.713			
540-36-3	1,4-Difluorobenzene	630000	8.615			
3114-55-4	Chlorobenzene-d5	584000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	232000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/25/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-83	SDG No.:	Q2436
Lab Sample ID:	Q2436-06	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	90.6
Sample Wt/Vol:	5.19	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022861.D	1		06/27/25 14:56	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-87	SDG No.:	Q2436
Lab Sample ID:	Q2436-07	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	89.9
Sample Wt/Vol:	6.71 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022862.D	1		06/27/25 15:19	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.94	U	0.94	4.10	ug/Kg
74-87-3	Chloromethane	0.94	U	0.94	4.10	ug/Kg
75-01-4	Vinyl Chloride	0.65	U	0.65	4.10	ug/Kg
74-83-9	Bromomethane	0.89	U	0.89	4.10	ug/Kg
75-00-3	Chloroethane	1.00	U	1.00	4.10	ug/Kg
75-69-4	Trichlorofluoromethane	1.00	U	1.00	4.10	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.88	U	0.88	4.10	ug/Kg
75-35-4	1,1-Dichloroethene	0.83	U	0.83	4.10	ug/Kg
67-64-1	Acetone	3.90	U	3.90	20.7	ug/Kg
75-15-0	Carbon Disulfide	0.88	U	0.88	4.10	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.61	U	0.61	4.10	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.10	ug/Kg
75-09-2	Methylene Chloride	2.90	U	2.90	8.30	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.71	U	0.71	4.10	ug/Kg
75-34-3	1,1-Dichloroethane	0.66	U	0.66	4.10	ug/Kg
110-82-7	Cyclohexane	0.65	U	0.65	4.10	ug/Kg
78-93-3	2-Butanone	5.40	U	5.40	20.7	ug/Kg
56-23-5	Carbon Tetrachloride	0.80	U	0.80	4.10	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.62	U	0.62	4.10	ug/Kg
74-97-5	Bromochloromethane	0.95	U	0.95	4.10	ug/Kg
67-66-3	Chloroform	0.70	U	0.70	4.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.77	U	0.77	4.10	ug/Kg
108-87-2	Methylcyclohexane	0.75	U	0.75	4.10	ug/Kg
71-43-2	Benzene	0.65	U	0.65	4.10	ug/Kg
107-06-2	1,2-Dichloroethane	0.65	U	0.65	4.10	ug/Kg
79-01-6	Trichloroethene	0.67	U	0.67	4.10	ug/Kg
78-87-5	1,2-Dichloropropane	0.75	U	0.75	4.10	ug/Kg
75-27-4	Bromodichloromethane	0.65	U	0.65	4.10	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.00	U	3.00	20.7	ug/Kg
108-88-3	Toluene	0.65	U	0.65	4.10	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-87	SDG No.:	Q2436
Lab Sample ID:	Q2436-07	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	89.9
Sample Wt/Vol:	6.71 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022862.D	1		06/27/25 15:19	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.54	U	0.54	4.10	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.51	U	0.51	4.10	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.76	U	0.76	4.10	ug/Kg
591-78-6	2-Hexanone	3.10	U	3.10	20.7	ug/Kg
124-48-1	Dibromochloromethane	0.72	U	0.72	4.10	ug/Kg
106-93-4	1,2-Dibromoethane	0.73	U	0.73	4.10	ug/Kg
127-18-4	Tetrachloroethene	0.87	U	0.87	4.10	ug/Kg
108-90-7	Chlorobenzene	0.75	U	0.75	4.10	ug/Kg
100-41-4	Ethyl Benzene	0.56	U	0.56	4.10	ug/Kg
179601-23-1	m/p-Xylenes	1.00	U	1.00	8.30	ug/Kg
95-47-6	o-Xylene	0.68	U	0.68	4.10	ug/Kg
100-42-5	Styrene	0.59	U	0.59	4.10	ug/Kg
75-25-2	Bromoform	0.71	U	0.71	4.10	ug/Kg
98-82-8	Isopropylbenzene	0.65	U	0.65	4.10	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.00	U	1.00	4.10	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.40	U	1.40	4.10	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.30	U	1.30	4.10	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.20	U	1.20	4.10	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.50	U	1.50	4.10	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.50	U	2.50	4.10	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.60	U	2.60	4.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	39.1		63 - 155	78%	SPK: 50
1868-53-7	Dibromofluoromethane	47.0		70 - 134	94%	SPK: 50
2037-26-5	Toluene-d8	49.1		74 - 123	98%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.2		17 - 146	90%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	349000	7.707			
540-36-3	1,4-Difluorobenzene	633000	8.616			
3114-55-4	Chlorobenzene-d5	541000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	204000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-87	SDG No.:	Q2436
Lab Sample ID:	Q2436-07	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	89.9
Sample Wt/Vol:	6.71	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022862.D	1		06/27/25 15:19	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-100	SDG No.:	Q2436
Lab Sample ID:	Q2436-08	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.6
Sample Wt/Vol:	7.14 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022863.D	1		06/27/25 15:43	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.93	U	0.93	4.10	ug/Kg
74-87-3	Chloromethane	0.93	U	0.93	4.10	ug/Kg
75-01-4	Vinyl Chloride	0.65	U	0.65	4.10	ug/Kg
74-83-9	Bromomethane	0.88	U	0.88	4.10	ug/Kg
75-00-3	Chloroethane	1.00	U	1.00	4.10	ug/Kg
75-69-4	Trichlorofluoromethane	0.99	U	0.99	4.10	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.87	U	0.87	4.10	ug/Kg
75-35-4	1,1-Dichloroethene	0.82	U	0.82	4.10	ug/Kg
67-64-1	Acetone	3.90	U	3.90	20.5	ug/Kg
75-15-0	Carbon Disulfide	0.87	U	0.87	4.10	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.60	U	0.60	4.10	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.10	ug/Kg
75-09-2	Methylene Chloride	3.40	J	2.90	8.20	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.70	U	0.70	4.10	ug/Kg
75-34-3	1,1-Dichloroethane	0.65	U	0.65	4.10	ug/Kg
110-82-7	Cyclohexane	0.65	U	0.65	4.10	ug/Kg
78-93-3	2-Butanone	5.40	U	5.40	20.5	ug/Kg
56-23-5	Carbon Tetrachloride	0.79	U	0.79	4.10	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.61	U	0.61	4.10	ug/Kg
74-97-5	Bromochloromethane	0.94	U	0.94	4.10	ug/Kg
67-66-3	Chloroform	0.69	U	0.69	4.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.76	U	0.76	4.10	ug/Kg
108-87-2	Methylcyclohexane	0.74	U	0.74	4.10	ug/Kg
71-43-2	Benzene	0.65	U	0.65	4.10	ug/Kg
107-06-2	1,2-Dichloroethane	0.65	U	0.65	4.10	ug/Kg
79-01-6	Trichloroethene	0.66	U	0.66	4.10	ug/Kg
78-87-5	1,2-Dichloropropane	0.74	U	0.74	4.10	ug/Kg
75-27-4	Bromodichloromethane	0.64	U	0.64	4.10	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.90	U	2.90	20.5	ug/Kg
108-88-3	Toluene	0.64	U	0.64	4.10	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-100	SDG No.:	Q2436
Lab Sample ID:	Q2436-08	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.6
Sample Wt/Vol:	7.14 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022863.D	1		06/27/25 15:43	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.53	U	0.53	4.10	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.51	U	0.51	4.10	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.75	U	0.75	4.10	ug/Kg
591-78-6	2-Hexanone	3.00	U	3.00	20.5	ug/Kg
124-48-1	Dibromochloromethane	0.71	U	0.71	4.10	ug/Kg
106-93-4	1,2-Dibromoethane	0.72	U	0.72	4.10	ug/Kg
127-18-4	Tetrachloroethene	0.86	U	0.86	4.10	ug/Kg
108-90-7	Chlorobenzene	0.74	U	0.74	4.10	ug/Kg
100-41-4	Ethyl Benzene	0.55	U	0.55	4.10	ug/Kg
179601-23-1	m/p-Xylenes	1.00	U	1.00	8.20	ug/Kg
95-47-6	o-Xylene	0.67	U	0.67	4.10	ug/Kg
100-42-5	Styrene	0.58	U	0.58	4.10	ug/Kg
75-25-2	Bromoform	0.70	U	0.70	4.10	ug/Kg
98-82-8	Isopropylbenzene	0.64	U	0.64	4.10	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.99	U	0.99	4.10	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.40	U	1.40	4.10	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.30	U	1.30	4.10	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.20	U	1.20	4.10	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.50	U	1.50	4.10	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.40	U	2.40	4.10	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.60	U	2.60	4.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.1		63 - 155	96%	SPK: 50
1868-53-7	Dibromofluoromethane	49.9		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	50.4		74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.0		17 - 146	110%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	312000	7.707			
540-36-3	1,4-Difluorobenzene	578000	8.616			
3114-55-4	Chlorobenzene-d5	554000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	240000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-100	SDG No.:	Q2436
Lab Sample ID:	Q2436-08	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	85.6
Sample Wt/Vol:	7.14 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022863.D	1		06/27/25 15:43	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-99	SDG No.:	Q2436
Lab Sample ID:	Q2436-09	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.2
Sample Wt/Vol:	5.48 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022864.D	1		06/27/25 16:06	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	4.90	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	4.90	ug/Kg
75-01-4	Vinyl Chloride	0.78	U	0.78	4.90	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	4.90	ug/Kg
75-00-3	Chloroethane	1.20	U	1.20	4.90	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	4.90	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.00	U	1.00	4.90	ug/Kg
75-35-4	1,1-Dichloroethene	0.99	U	0.99	4.90	ug/Kg
67-64-1	Acetone	4.70	U	4.70	24.7	ug/Kg
75-15-0	Carbon Disulfide	1.00	U	1.00	4.90	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.72	U	0.72	4.90	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	4.90	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	9.90	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.85	U	0.85	4.90	ug/Kg
75-34-3	1,1-Dichloroethane	0.79	U	0.79	4.90	ug/Kg
110-82-7	Cyclohexane	0.78	U	0.78	4.90	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	24.7	ug/Kg
56-23-5	Carbon Tetrachloride	0.96	U	0.96	4.90	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.74	U	0.74	4.90	ug/Kg
74-97-5	Bromochloromethane	1.10	U	1.10	4.90	ug/Kg
67-66-3	Chloroform	0.83	U	0.83	4.90	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.92	U	0.92	4.90	ug/Kg
108-87-2	Methylcyclohexane	0.90	U	0.90	4.90	ug/Kg
71-43-2	Benzene	0.78	U	0.78	4.90	ug/Kg
107-06-2	1,2-Dichloroethane	0.78	U	0.78	4.90	ug/Kg
79-01-6	Trichloroethene	0.80	U	0.80	4.90	ug/Kg
78-87-5	1,2-Dichloropropane	0.90	U	0.90	4.90	ug/Kg
75-27-4	Bromodichloromethane	0.77	U	0.77	4.90	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.50	U	3.50	24.7	ug/Kg
108-88-3	Toluene	0.77	U	0.77	4.90	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-99	SDG No.:	Q2436
Lab Sample ID:	Q2436-09	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.2
Sample Wt/Vol:	5.48 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022864.D	1		06/27/25 16:06	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.64	U	0.64	4.90	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.61	U	0.61	4.90	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.91	U	0.91	4.90	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	24.7	ug/Kg
124-48-1	Dibromochloromethane	0.86	U	0.86	4.90	ug/Kg
106-93-4	1,2-Dibromoethane	0.87	U	0.87	4.90	ug/Kg
127-18-4	Tetrachloroethene	1.00	U	1.00	4.90	ug/Kg
108-90-7	Chlorobenzene	0.90	U	0.90	4.90	ug/Kg
100-41-4	Ethyl Benzene	0.66	U	0.66	4.90	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	9.90	ug/Kg
95-47-6	o-Xylene	0.81	U	0.81	4.90	ug/Kg
100-42-5	Styrene	0.70	U	0.70	4.90	ug/Kg
75-25-2	Bromoform	0.85	U	0.85	4.90	ug/Kg
98-82-8	Isopropylbenzene	0.77	U	0.77	4.90	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	4.90	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	4.90	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.50	U	1.50	4.90	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.40	U	1.40	4.90	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	4.90	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.90	U	2.90	4.90	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.10	U	3.10	4.90	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.2		63 - 155	100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.0		74 - 123	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		17 - 146	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	301000	7.713			
540-36-3	1,4-Difluorobenzene	558000	8.616			
3114-55-4	Chlorobenzene-d5	537000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	231000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-99	SDG No.:	Q2436
Lab Sample ID:	Q2436-09	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92.2
Sample Wt/Vol:	5.48	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022864.D	1		06/27/25 16:06	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-82	SDG No.:	Q2436
Lab Sample ID:	Q2436-10	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92
Sample Wt/Vol:	5.75 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022865.D	1		06/27/25 16:30	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	4.70	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	4.70	ug/Kg
75-01-4	Vinyl Chloride	0.75	U	0.75	4.70	ug/Kg
74-83-9	Bromomethane	1.00	U	1.00	4.70	ug/Kg
75-00-3	Chloroethane	1.20	U	1.20	4.70	ug/Kg
75-69-4	Trichlorofluoromethane	1.10	U	1.10	4.70	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.00	U	1.00	4.70	ug/Kg
75-35-4	1,1-Dichloroethene	0.95	U	0.95	4.70	ug/Kg
67-64-1	Acetone	4.50	U	4.50	23.6	ug/Kg
75-15-0	Carbon Disulfide	1.00	U	1.00	4.70	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.69	U	0.69	4.70	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	4.70	ug/Kg
75-09-2	Methylene Chloride	3.60	J	3.30	9.50	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.81	U	0.81	4.70	ug/Kg
75-34-3	1,1-Dichloroethane	0.76	U	0.76	4.70	ug/Kg
110-82-7	Cyclohexane	0.75	U	0.75	4.70	ug/Kg
78-93-3	2-Butanone	6.20	U	6.20	23.6	ug/Kg
56-23-5	Carbon Tetrachloride	0.92	U	0.92	4.70	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.71	U	0.71	4.70	ug/Kg
74-97-5	Bromochloromethane	1.10	U	1.10	4.70	ug/Kg
67-66-3	Chloroform	0.79	U	0.79	4.70	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.88	U	0.88	4.70	ug/Kg
108-87-2	Methylcyclohexane	0.86	U	0.86	4.70	ug/Kg
71-43-2	Benzene	0.75	U	0.75	4.70	ug/Kg
107-06-2	1,2-Dichloroethane	0.75	U	0.75	4.70	ug/Kg
79-01-6	Trichloroethene	0.77	U	0.77	4.70	ug/Kg
78-87-5	1,2-Dichloropropane	0.86	U	0.86	4.70	ug/Kg
75-27-4	Bromodichloromethane	0.74	U	0.74	4.70	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.40	U	3.40	23.6	ug/Kg
108-88-3	Toluene	0.74	U	0.74	4.70	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-82	SDG No.:	Q2436
Lab Sample ID:	Q2436-10	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92
Sample Wt/Vol:	5.75 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022865.D	1		06/27/25 16:30	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.61	U	0.61	4.70	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.59	U	0.59	4.70	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.87	U	0.87	4.70	ug/Kg
591-78-6	2-Hexanone	3.50	U	3.50	23.6	ug/Kg
124-48-1	Dibromochloromethane	0.82	U	0.82	4.70	ug/Kg
106-93-4	1,2-Dibromoethane	0.83	U	0.83	4.70	ug/Kg
127-18-4	Tetrachloroethene	0.99	U	0.99	4.70	ug/Kg
108-90-7	Chlorobenzene	0.86	U	0.86	4.70	ug/Kg
100-41-4	Ethyl Benzene	0.63	U	0.63	4.70	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	9.50	ug/Kg
95-47-6	o-Xylene	0.78	U	0.78	4.70	ug/Kg
100-42-5	Styrene	0.67	U	0.67	4.70	ug/Kg
75-25-2	Bromoform	0.81	U	0.81	4.70	ug/Kg
98-82-8	Isopropylbenzene	0.74	U	0.74	4.70	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.10	U	1.10	4.70	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.60	U	1.60	4.70	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.50	U	1.50	4.70	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.40	U	1.40	4.70	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.70	U	1.70	4.70	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.80	U	2.80	4.70	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.00	U	3.00	4.70	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.0		63 - 155	98%	SPK: 50
1868-53-7	Dibromofluoromethane	50.5		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.5		74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.6		17 - 146	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	319000	7.707			
540-36-3	1,4-Difluorobenzene	604000	8.615			
3114-55-4	Chlorobenzene-d5	583000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	251000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	06/26/25
Project:	South River WM Replacement	Date Received:	06/26/25
Client Sample ID:	TP-82	SDG No.:	Q2436
Lab Sample ID:	Q2436-10	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	92
Sample Wt/Vol:	5.75	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022865.D	1		06/27/25 16:30	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q2436**

Client: **CDM Smith**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits	
							Low	High
Q2436-01	TP-70	1,2-Dichloroethane-d4	50	41.9	84		63	155
		Dibromofluoromethane	50	47.8	96		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	49.6	99		17	146
Q2436-02	TP-69	1,2-Dichloroethane-d4	50	46.8	94		63	155
		Dibromofluoromethane	50	49.9	100		70	134
		Toluene-d8	50	49.5	99		74	123
		4-Bromofluorobenzene	50	54.2	108		17	146
Q2436-03	TP-85	1,2-Dichloroethane-d4	50	45.5	91		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	49.9	100		74	123
		4-Bromofluorobenzene	50	54.2	108		17	146
Q2436-04	TP-86	1,2-Dichloroethane-d4	50	48.2	96		63	155
		Dibromofluoromethane	50	49.7	99		70	134
		Toluene-d8	50	50.2	100		74	123
		4-Bromofluorobenzene	50	57.9	116		17	146
Q2436-05	TP-84	1,2-Dichloroethane-d4	50	49.9	100		63	155
		Dibromofluoromethane	50	49.9	100		70	134
		Toluene-d8	50	50.6	101		74	123
		4-Bromofluorobenzene	50	54.6	109		17	146
Q2436-06	TP-83	1,2-Dichloroethane-d4	50	46.5	93		63	155
		Dibromofluoromethane	50	49.7	99		70	134
		Toluene-d8	50	49.6	99		74	123
		4-Bromofluorobenzene	50	50.7	101		17	146
Q2436-07	TP-87	1,2-Dichloroethane-d4	50	39.1	78		63	155
		Dibromofluoromethane	50	47.0	94		70	134
		Toluene-d8	50	49.1	98		74	123
		4-Bromofluorobenzene	50	45.2	90		17	146
Q2436-08	TP-100	1,2-Dichloroethane-d4	50	48.1	96		63	155
		Dibromofluoromethane	50	49.9	100		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	55.0	110		17	146
Q2436-09	TP-99	1,2-Dichloroethane-d4	50	50.2	100		63	155
		Dibromofluoromethane	50	50.7	101		70	134
		Toluene-d8	50	50.0	100		74	123
		4-Bromofluorobenzene	50	54.3	109		17	146
Q2436-10	TP-82	1,2-Dichloroethane-d4	50	49.0	98		63	155
		Dibromofluoromethane	50	50.5	101		70	134
		Toluene-d8	50	50.5	101		74	123
		4-Bromofluorobenzene	50	54.6	109		17	146
VY0627SBL01	VY0627SBL01	1,2-Dichloroethane-d4	50	44.5	89		63	155
		Dibromofluoromethane	50	49.1	98		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	51.2	102		17	146
VY0627SBS01	VY0627SBS01	1,2-Dichloroethane-d4	50	48.5	97		63	155
		Dibromofluoromethane	50	50.3	101		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	48.4	97		17	146
VY0627SBSD01	VY0627SBSD01	1,2-Dichloroethane-d4	50	51.8	104		63	155
		Dibromofluoromethane	50	51.6	103		70	134

Surrogate Summary

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits	
							Low	High
VY0627SBSD01	VY0627SBSD01	Toluene-d8	50	51.0	102		74	123
		4-Bromofluorobenzene	50	49.2	98		17	146
VY0630SBL01	VY0630SBL01	1,2-Dichloroethane-d4	50	44.2	88		63	155
		Dibromofluoromethane	50	50.1	100		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	54.2	108		17	146
VY0630SBS01	VY0630SBS01	1,2-Dichloroethane-d4	50	46.6	93		63	155
		Dibromofluoromethane	50	48.0	96		70	134
		Toluene-d8	50	48.5	97		74	123
		4-Bromofluorobenzene	50	47.0	94		17	146

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022854.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0627SBS01	Dichlorodifluoromethane	20	20.3	ug/Kg	102			64	136	
	Chloromethane	20	21.4	ug/Kg	107			52	151	
	Vinyl chloride	20	19.1	ug/Kg	96			56	148	
	Bromomethane	20	22.9	ug/Kg	115			58	141	
	Chloroethane	20	19.8	ug/Kg	99			69	130	
	Trichlorofluoromethane	20	18.5	ug/Kg	93			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/Kg	103			81	123	
	1,1-Dichloroethene	20	20.2	ug/Kg	101			79	121	
	Acetone	100	120	ug/Kg	120			40	171	
	Carbon disulfide	20	19.9	ug/Kg	100			59	130	
	Methyl tert-butyl Ether	20	19.0	ug/Kg	95			77	129	
	Methyl Acetate	20	15.2	ug/Kg	76			69	149	
	Methylene Chloride	20	23.3	ug/Kg	117			72	131	
	trans-1,2-Dichloroethene	20	19.9	ug/Kg	100			80	123	
	1,1-Dichloroethane	20	20.1	ug/Kg	101			82	123	
	Cyclohexane	20	20.1	ug/Kg	101			76	122	
	2-Butanone	100	94.9	ug/Kg	95			69	131	
	Carbon Tetrachloride	20	19.7	ug/Kg	99			76	129	
	cis-1,2-Dichloroethene	20	19.8	ug/Kg	99			82	123	
	Bromochloromethane	20	19.8	ug/Kg	99			80	127	
	Chloroform	20	19.9	ug/Kg	100			82	125	
	1,1,1-Trichloroethane	20	20.1	ug/Kg	101			80	126	
	Methylcyclohexane	20	20.1	ug/Kg	101			77	123	
	Benzene	20	19.9	ug/Kg	100			84	121	
	1,2-Dichloroethane	20	19.0	ug/Kg	95			81	126	
	Trichloroethene	20	19.9	ug/Kg	100			83	122	
	1,2-Dichloropropane	20	19.7	ug/Kg	99			83	122	
	Bromodichloromethane	20	19.2	ug/Kg	96			82	123	
	4-Methyl-2-Pentanone	100	85.5	ug/Kg	86			70	135	
	Toluene	20	19.7	ug/Kg	99			83	122	
	t-1,3-Dichloropropene	20	18.6	ug/Kg	93			78	124	
	cis-1,3-Dichloropropene	20	19.9	ug/Kg	100			81	122	
	1,1,2-Trichloroethane	20	18.7	ug/Kg	94			82	125	
	2-Hexanone	100	89.6	ug/Kg	90			66	138	
	Dibromochloromethane	20	18.6	ug/Kg	93			79	125	
	1,2-Dibromoethane	20	18.4	ug/Kg	92			80	125	
	Tetrachloroethene	20	20.1	ug/Kg	101			83	125	
	Chlorobenzene	20	20.1	ug/Kg	101			84	122	
	Ethyl Benzene	20	20.5	ug/Kg	103			82	124	
	m/p-Xylenes	40	40.3	ug/Kg	101			83	124	
	o-Xylene	20	20.0	ug/Kg	100			83	123	
	Styrene	20	19.4	ug/Kg	97			82	124	
	Bromoform	20	17.8	ug/Kg	89			75	127	
	Isopropylbenzene	20	21.3	ug/Kg	106			82	124	
	1,1,2,2-Tetrachloroethane	20	19.1	ug/Kg	96			77	127	
	1,3-Dichlorobenzene	20	20.1	ug/Kg	101			83	122	
	1,4-Dichlorobenzene	20	20.0	ug/Kg	100			84	121	
	1,2-Dichlorobenzene	20	19.5	ug/Kg	98			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022854.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0627SBS01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/Kg	88			66	134	
	1,2,4-Trichlorobenzene	20	18.7	ug/Kg	94			78	127	
	1,2,3-Trichlorobenzene	20	18.3	ug/Kg	92			70	137	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022855.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0627SBSD01	Dichlorodifluoromethane	20	20.9	ug/Kg	104	2		64	136	20
	Chloromethane	20	22.0	ug/Kg	110	3		52	151	20
	Vinyl chloride	20	19.1	ug/Kg	96	0		56	148	20
	Bromomethane	20	22.7	ug/Kg	114	1		58	141	20
	Chloroethane	20	18.7	ug/Kg	94	5		69	130	20
	Trichlorofluoromethane	20	17.9	ug/Kg	90	3		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/Kg	102	1		81	123	20
	1,1-Dichloroethene	20	21.5	ug/Kg	108	7		79	121	20
	Acetone	100	120	ug/Kg	120	0		40	171	20
	Carbon disulfide	20	20.9	ug/Kg	104	4		59	130	20
	Methyl tert-butyl Ether	20	21.5	ug/Kg	108	13		77	129	20
	Methyl Acetate	20	17.8	ug/Kg	89	16		69	149	20
	Methylene Chloride	20	24.9	ug/Kg	125	7		72	131	20
	trans-1,2-Dichloroethene	20	21.6	ug/Kg	108	8		80	123	20
	1,1-Dichloroethane	20	22.1	ug/Kg	111	9		82	123	20
	Cyclohexane	20	20.4	ug/Kg	102	1		76	122	20
	2-Butanone	100	110	ug/Kg	110	15		69	131	20
	Carbon Tetrachloride	20	19.8	ug/Kg	99	0		76	129	20
	cis-1,2-Dichloroethene	20	21.7	ug/Kg	109	10		82	123	20
	Bromochloromethane	20	21.0	ug/Kg	105	6		80	127	20
	Chloroform	20	21.5	ug/Kg	108	8		82	125	20
	1,1,1-Trichloroethane	20	21.1	ug/Kg	106	5		80	126	20
	Methylcyclohexane	20	19.7	ug/Kg	99	2		77	123	20
	Benzene	20	20.8	ug/Kg	104	4		84	121	20
	1,2-Dichloroethane	20	20.9	ug/Kg	104	9		81	126	20
	Trichloroethene	20	20.4	ug/Kg	102	2		83	122	20
	1,2-Dichloropropane	20	21.1	ug/Kg	106	7		83	122	20
	Bromodichloromethane	20	20.8	ug/Kg	104	8		82	123	20
	4-Methyl-2-Pentanone	100	94.0	ug/Kg	94	9		70	135	20
	Toluene	20	20.5	ug/Kg	103	4		83	122	20
	t-1,3-Dichloropropene	20	20.4	ug/Kg	102	9		78	124	20
	cis-1,3-Dichloropropene	20	21.1	ug/Kg	106	6		81	122	20
	1,1,2-Trichloroethane	20	20.3	ug/Kg	102	8		82	125	20
	2-Hexanone	100	95.1	ug/Kg	95	5		66	138	20
	Dibromochloromethane	20	20.1	ug/Kg	101	8		79	125	20
	1,2-Dibromoethane	20	20.0	ug/Kg	100	8		80	125	20
	Tetrachloroethene	20	20.5	ug/Kg	103	2		83	125	20
	Chlorobenzene	20	21.0	ug/Kg	105	4		84	122	20
	Ethyl Benzene	20	20.9	ug/Kg	104	1		82	124	20
	m/p-Xylenes	40	40.8	ug/Kg	102	1		83	124	20
	o-Xylene	20	20.7	ug/Kg	104	4		83	123	20
	Styrene	20	20.3	ug/Kg	102	5		82	124	20
	Bromoform	20	19.0	ug/Kg	95	7		75	127	20
	Isopropylbenzene	20	21.6	ug/Kg	108	2		82	124	20
	1,1,2,2-Tetrachloroethane	20	20.0	ug/Kg	100	4		77	127	20
	1,3-Dichlorobenzene	20	21.0	ug/Kg	105	4		83	122	20
	1,4-Dichlorobenzene	20	21.2	ug/Kg	106	6		84	121	20
	1,2-Dichlorobenzene	20	20.4	ug/Kg	102	4		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022855.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0627SBSD01	1,2-Dibromo-3-Chloropropane	20	18.7	ug/Kg	94	7		66	134	20
	1,2,4-Trichlorobenzene	20	20.0	ug/Kg	100	6		78	127	20
	1,2,3-Trichlorobenzene	20	19.4	ug/Kg	97	5		70	137	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022874.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0630SBS01	Dichlorodifluoromethane	20	20.1	ug/Kg	101			64	136	
	Chloromethane	20	21.7	ug/Kg	109			52	151	
	Vinyl chloride	20	19.4	ug/Kg	97			56	148	
	Bromomethane	20	20.1	ug/Kg	101			58	141	
	Chloroethane	20	19.2	ug/Kg	96			69	130	
	Trichlorofluoromethane	20	18.7	ug/Kg	94			69	134	
	1,1,2-Trichlorotrifluoroethane	20	20.8	ug/Kg	104			81	123	
	1,1-Dichloroethene	20	20.6	ug/Kg	103			79	121	
	Acetone	100	130	ug/Kg	130			40	171	
	Carbon disulfide	20	20.3	ug/Kg	102			59	130	
	Methyl tert-butyl Ether	20	18.8	ug/Kg	94			77	129	
	Methyl Acetate	20	18.5	ug/Kg	93			69	149	
	Methylene Chloride	20	21.7	ug/Kg	109			72	131	
	trans-1,2-Dichloroethene	20	20.2	ug/Kg	101			80	123	
	1,1-Dichloroethane	20	20.4	ug/Kg	102			82	123	
	Cyclohexane	20	20.8	ug/Kg	104			76	122	
	2-Butanone	100	110	ug/Kg	110			69	131	
	Carbon Tetrachloride	20	20.0	ug/Kg	100			76	129	
	cis-1,2-Dichloroethene	20	20.0	ug/Kg	100			82	123	
	Bromochloromethane	20	19.6	ug/Kg	98			80	127	
	Chloroform	20	19.9	ug/Kg	100			82	125	
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101			80	126	
	Methylcyclohexane	20	21.1	ug/Kg	106			77	123	
	Benzene	20	20.1	ug/Kg	101			84	121	
	1,2-Dichloroethane	20	19.4	ug/Kg	97			81	126	
	Trichloroethene	20	19.6	ug/Kg	98			83	122	
	1,2-Dichloropropane	20	20.3	ug/Kg	102			83	122	
	Bromodichloromethane	20	19.4	ug/Kg	97			82	123	
	4-Methyl-2-Pentanone	100	90.8	ug/Kg	91			70	135	
	Toluene	20	19.7	ug/Kg	99			83	122	
	t-1,3-Dichloropropene	20	18.8	ug/Kg	94			78	124	
	cis-1,3-Dichloropropene	20	19.8	ug/Kg	99			81	122	
	1,1,2-Trichloroethane	20	19.1	ug/Kg	96			82	125	
	2-Hexanone	100	98.0	ug/Kg	98			66	138	
	Dibromochloromethane	20	18.7	ug/Kg	94			79	125	
	1,2-Dibromoethane	20	18.6	ug/Kg	93			80	125	
	Tetrachloroethene	20	18.3	ug/Kg	92			83	125	
	Chlorobenzene	20	19.9	ug/Kg	100			84	122	
	Ethyl Benzene	20	20.1	ug/Kg	101			82	124	
	m/p-Xylenes	40	40.2	ug/Kg	101			83	124	
	o-Xylene	20	19.8	ug/Kg	99			83	123	
	Styrene	20	19.4	ug/Kg	97			82	124	
	Bromoform	20	18.3	ug/Kg	92			75	127	
	Isopropylbenzene	20	21.1	ug/Kg	106			82	124	
	1,1,2,2-Tetrachloroethane	20	20.6	ug/Kg	103			77	127	
	1,3-Dichlorobenzene	20	20.3	ug/Kg	102			83	122	
	1,4-Dichlorobenzene	20	20.1	ug/Kg	101			84	121	
	1,2-Dichlorobenzene	20	19.5	ug/Kg	98			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2436

Client: CDM Smith

Analytical Method: SW8260D

Datafile : VY022874.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0630SBS01	1,2-Dibromo-3-Chloropropane	20	18.0	ug/Kg	90			66	134	
	1,2,4-Trichlorobenzene	20	19.6	ug/Kg	98			78	127	
	1,2,3-Trichlorobenzene	20	19.0	ug/Kg	95			70	137	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0627SBL01

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436

SAS No.: Q2436 SDG NO.: Q2436

Lab File ID: VY022853.D

Lab Sample ID: VY0627SBL01

Date Analyzed: 06/27/2025

Time Analyzed: 11:32

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0627SBS01	VY0627SBS01	VY022854.D	06/27/2025
VY0627SBSD01	VY0627SBSD01	VY022855.D	06/27/2025
TP-70	Q2436-01	VY022856.D	06/27/2025
TP-69	Q2436-02	VY022857.D	06/27/2025
TP-83	Q2436-06	VY022861.D	06/27/2025
TP-87	Q2436-07	VY022862.D	06/27/2025
TP-100	Q2436-08	VY022863.D	06/27/2025
TP-99	Q2436-09	VY022864.D	06/27/2025
TP-82	Q2436-10	VY022865.D	06/27/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0630SBL01

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q2436

SAS No.: Q2436 SDG NO.: Q2436

Lab File ID: VY022873.D

Lab Sample ID: VY0630SBL01

Date Analyzed: 06/30/2025

Time Analyzed: 11:39

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0630SBS01	VY0630SBS01	VY022874.D	06/30/2025
TP-85	Q2436-03	VY022877.D	06/30/2025
TP-86	Q2436-04	VY022878.D	06/30/2025
TP-84	Q2436-05	VY022879.D	06/30/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022775.D BFB Injection Date: 06/23/2025
Instrument ID: MSVOA_Y BFB Injection Time: 10:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.8
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	81.9
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.2 (95.5) 1
177	5.0 - 9.0% of mass 176	5.1 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY022776.D	06/23/2025	13:38
VSTDIC010	VSTDIC010	VY022777.D	06/23/2025	14:00
VSTDIC020	VSTDIC020	VY022778.D	06/23/2025	14:23
VSTDIC050	VSTDIC050	VY022779.D	06/23/2025	14:46
VSTDIC100	VSTDIC100	VY022780.D	06/23/2025	15:08
VSTDIC150	VSTDIC150	VY022781.D	06/23/2025	15:31

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022851.D BFB Injection Date: 06/27/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:31
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.9 (1.1) 1
174	50.0 - 100.0% of mass 95	80.5
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	77.3 (96) 1
177	5.0 - 9.0% of mass 176	5.1 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022852.D	06/27/2025	11:00
VY0627SBL01	VY0627SBL01	VY022853.D	06/27/2025	11:32
VY0627SBS01	VY0627SBS01	VY022854.D	06/27/2025	12:00
VY0627SBSD01	VY0627SBSD01	VY022855.D	06/27/2025	12:22
TP-70	Q2436-01	VY022856.D	06/27/2025	12:59
TP-69	Q2436-02	VY022857.D	06/27/2025	13:22
TP-83	Q2436-06	VY022861.D	06/27/2025	14:56
TP-87	Q2436-07	VY022862.D	06/27/2025	15:19
TP-100	Q2436-08	VY022863.D	06/27/2025	15:43
TP-99	Q2436-09	VY022864.D	06/27/2025	16:06
TP-82	Q2436-10	VY022865.D	06/27/2025	16:30

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022871.D BFB Injection Date: 06/30/2025
Instrument ID: MSVOA_Y BFB Injection Time: 08:42
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.8
75	30.0 - 60.0% of mass 95	53.9
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1) 1
174	50.0 - 100.0% of mass 95	84.6
175	5.0 - 9.0% of mass 174	6.3 (7.4) 1
176	95.0 - 101.0% of mass 174	80.8 (95.5) 1
177	5.0 - 9.0% of mass 176	5.3 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY022872.D	06/30/2025	11:03
VY0630SBL01	VY0630SBL01	VY022873.D	06/30/2025	11:39
VY0630SBS01	VY0630SBS01	VY022874.D	06/30/2025	12:09
TP-85	Q2436-03	VY022877.D	06/30/2025	13:32
TP-86	Q2436-04	VY022878.D	06/30/2025	14:28
TP-84	Q2436-05	VY022879.D	06/30/2025	14:52

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022852.D Date Analyzed: 06/27/2025
Instrument ID: MSVOA_Y Time Analyzed: 11:00
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	460120	7.71	754723	8.62	647658	11.42
UPPER LIMIT	920240	8.213	1509450	9.122	1295320	11.92
LOWER LIMIT	230060	7.213	377362	8.122	323829	10.92
EPA SAMPLE NO.						
TP-70	370411	7.71	661194	8.62	597332	11.42
TP-69	340803	7.71	625770	8.62	598373	11.42
TP-83	337210	7.71	630445	8.62	584339	11.41
TP-87	348864	7.71	632840	8.62	540565	11.41
TP-100	312427	7.71	578070	8.62	554007	11.41
TP-99	301121	7.71	558142	8.62	536522	11.41
TP-82	318610	7.71	604014	8.62	582898	11.41
VY0627SBL01	358693	7.71	655911	8.62	603519	11.42
VY0627SBS01	438526	7.71	740220	8.62	622783	11.42
VY0627SBSD01	392671	7.71	686703	8.62	587941	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022852.D Date Analyzed: 06/27/2025
Instrument ID: MSVOA_Y Time Analyzed: 11:00
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	319930	13.347				
UPPER LIMIT	639860	13.847				
LOWER LIMIT	159965	12.847				
EPA SAMPLE NO.						
TP-70	238564	13.35				
TP-69	255220	13.35				
TP-83	232255	13.35				
TP-87	204342	13.35				
TP-100	240211	13.35				
TP-99	230605	13.35				
TP-82	250591	13.35				
VY0627SBL01	250491	13.35				
VY0627SBS01	290353	13.35				
VY0627SBSD01	274040	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
Lab File ID: VY022872.D Date Analyzed: 06/30/2025
Instrument ID: MSVOA_Y Time Analyzed: 11:03
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	450935	7.71	753538	8.62	651538	11.41
UPPER LIMIT	901870	8.207	1507080	9.115	1303080	11.914
LOWER LIMIT	225468	7.207	376769	8.115	325769	10.914
EPA SAMPLE NO.						
TP-85	372972	7.71	690751	8.61	657778	11.41
TP-86	346216	7.71	639767	8.62	633746	11.41
TP-84	323652	7.71	612070	8.61	599073	11.41
VY0630SBL01	371114	7.71	672085	8.62	635637	11.41
VY0630SBS01	473861	7.71	796065	8.62	673720	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
 Lab File ID: VY022872.D Date Analyzed: 06/30/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	313014	13.346				
UPPER LIMIT	626028	13.846				
LOWER LIMIT	156507	12.846				
EPA SAMPLE NO.						
TP-85	284435	13.34				
TP-86	290240	13.35				
TP-84	250413	13.34				
VY0630SBL01	274914	13.35				
VY0630SBS01	318118	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBL01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022853.D	1		06/27/25 11:32	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	CDM Smith		Date Collected:	
Project:	South River WM Replacement		Date Received:	
Client Sample ID:	VY0627SBL01		SDG No.:	Q2436
Lab Sample ID:	VY0627SBL01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022853.D	1		06/27/25 11:32	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	44.5		63 - 155	89%	SPK: 50
1868-53-7	Dibromofluoromethane	49.2		70 - 134	98%	SPK: 50
2037-26-5	Toluene-d8	49.3		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.2		17 - 146	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	359000	7.713			
540-36-3	1,4-Difluorobenzene	656000	8.622			
3114-55-4	Chlorobenzene-d5	604000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	250000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBL01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022853.D	1		06/27/25 11:32	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBL01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022873.D	1		06/30/25 11:39	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	U	1.10	5.00	ug/Kg
74-87-3	Chloromethane	1.10	U	1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	0.79	U	0.79	5.00	ug/Kg
74-83-9	Bromomethane	1.10	U	1.10	5.00	ug/Kg
75-00-3	Chloroethane	1.30	U	1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	1.20	U	1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	U	1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	1.00	U	1.00	5.00	ug/Kg
67-64-1	Acetone	4.70	U	4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	1.10	U	1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.73	U	0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	1.50	U	1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	3.50	U	3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.86	U	0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	0.80	U	0.80	5.00	ug/Kg
110-82-7	Cyclohexane	0.79	U	0.79	5.00	ug/Kg
78-93-3	2-Butanone	6.50	U	6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	0.97	U	0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	1.20	U	1.20	5.00	ug/Kg
67-66-3	Chloroform	0.84	U	0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	0.91	U	0.91	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	0.91	U	0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	0.78	U	0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3.60	U	3.60	25.0	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBL01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022873.D	1		06/30/25 11:39	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.65	U	0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.62	U	0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.92	U	0.92	5.00	ug/Kg
591-78-6	2-Hexanone	3.70	U	3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	0.87	U	0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	0.88	U	0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	1.10	U	1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	0.91	U	0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	1.20	U	1.20	10.0	ug/Kg
95-47-6	o-Xylene	0.82	U	0.82	5.00	ug/Kg
100-42-5	Styrene	0.71	U	0.71	5.00	ug/Kg
75-25-2	Bromoform	0.86	U	0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1.20	U	1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.70	U	1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.60	U	1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.50	U	1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.80	U	1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	3.00	U	3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	3.20	U	3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	44.2		63 - 155	88%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	49.3		74 - 123	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		17 - 146	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	371000	7.707			
540-36-3	1,4-Difluorobenzene	672000	8.615			
3114-55-4	Chlorobenzene-d5	636000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	275000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBL01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBL01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022873.D	1		06/30/25 11:39	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBS01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022854.D	1		06/27/25 12:00	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.3		1.10	5.00	ug/Kg
74-87-3	Chloromethane	21.4		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.1		0.79	5.00	ug/Kg
74-83-9	Bromomethane	22.9		1.10	5.00	ug/Kg
75-00-3	Chloroethane	19.8		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	18.5		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.6		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	20.2		1.00	5.00	ug/Kg
67-64-1	Acetone	120		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	19.9		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	19.0		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	15.2		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	23.3		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	19.9		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	20.1		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	20.1		0.79	5.00	ug/Kg
78-93-3	2-Butanone	94.9		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	19.7		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	19.8		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	19.8		1.20	5.00	ug/Kg
67-66-3	Chloroform	19.9		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.1		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	20.1		0.91	5.00	ug/Kg
71-43-2	Benzene	19.9		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	19.0		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	19.9		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	19.7		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	19.2		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	85.5		3.60	25.0	ug/Kg
108-88-3	Toluene	19.7		0.78	5.00	ug/Kg

Report of Analysis

Client:	CDM Smith		Date Collected:	
Project:	South River WM Replacement		Date Received:	
Client Sample ID:	VY0627SBS01		SDG No.:	Q2436
Lab Sample ID:	VY0627SBS01		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022854.D	1		06/27/25 12:00	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	18.6		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	19.9		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	18.7		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	89.6		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	18.6		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	18.4		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	20.1		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	20.1		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.5		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.3		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.0		0.82	5.00	ug/Kg
100-42-5	Styrene	19.4		0.71	5.00	ug/Kg
75-25-2	Bromoform	17.8		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.3		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	19.1		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.1		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.0		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	19.5		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	17.5		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	18.7		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	18.3		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.5		63 - 155	97%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	50.4		74 - 123	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.4		17 - 146	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	439000	7.713			
540-36-3	1,4-Difluorobenzene	740000	8.616			
3114-55-4	Chlorobenzene-d5	623000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	290000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBS01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022854.D	1		06/27/25 12:00	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBS01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022874.D	1		06/30/25 12:09	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.1		1.10	5.00	ug/Kg
74-87-3	Chloromethane	21.7		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.4		0.79	5.00	ug/Kg
74-83-9	Bromomethane	20.1		1.10	5.00	ug/Kg
75-00-3	Chloroethane	19.2		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	18.7		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.8		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	20.6		1.00	5.00	ug/Kg
67-64-1	Acetone	130		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	20.3		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	18.8		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	18.5		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	21.7		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	20.2		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	20.4		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	20.8		0.79	5.00	ug/Kg
78-93-3	2-Butanone	110		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	20.0		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	20.0		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	19.6		1.20	5.00	ug/Kg
67-66-3	Chloroform	19.9		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.2		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	21.1		0.91	5.00	ug/Kg
71-43-2	Benzene	20.1		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	19.4		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	19.6		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	20.3		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	19.4		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	90.8		3.60	25.0	ug/Kg
108-88-3	Toluene	19.7		0.78	5.00	ug/Kg

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBS01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022874.D	1		06/30/25 12:09	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	18.8		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	19.8		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.1		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	98.0		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	18.7		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	18.6		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	18.3		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	19.9		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.1		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.2		1.20	10.0	ug/Kg
95-47-6	o-Xylene	19.8		0.82	5.00	ug/Kg
100-42-5	Styrene	19.4		0.71	5.00	ug/Kg
75-25-2	Bromoform	18.3		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.1		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	20.6		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	20.3		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	20.1		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	19.5		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.0		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	19.6		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.0		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.6		63 - 155	93%	SPK: 50
1868-53-7	Dibromofluoromethane	48.0		70 - 134	96%	SPK: 50
2037-26-5	Toluene-d8	48.5		74 - 123	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.0		17 - 146	94%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	474000	7.707			
540-36-3	1,4-Difluorobenzene	796000	8.616			
3114-55-4	Chlorobenzene-d5	674000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	318000	13.34			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0630SBS01	SDG No.:	Q2436
Lab Sample ID:	VY0630SBS01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022874.D	1		06/30/25 12:09	VY063025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBSD01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBSD01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022855.D	1		06/27/25 12:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	20.9		1.10	5.00	ug/Kg
74-87-3	Chloromethane	22.0		1.10	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.1		0.79	5.00	ug/Kg
74-83-9	Bromomethane	22.7		1.10	5.00	ug/Kg
75-00-3	Chloroethane	18.7		1.30	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	17.9		1.20	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	20.4		1.10	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	21.5		1.00	5.00	ug/Kg
67-64-1	Acetone	120		4.70	25.0	ug/Kg
75-15-0	Carbon Disulfide	20.9		1.10	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	21.5		0.73	5.00	ug/Kg
79-20-9	Methyl Acetate	17.8		1.50	5.00	ug/Kg
75-09-2	Methylene Chloride	24.9		3.50	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	21.6		0.86	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	22.1		0.80	5.00	ug/Kg
110-82-7	Cyclohexane	20.4		0.79	5.00	ug/Kg
78-93-3	2-Butanone	110		6.50	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	19.8		0.97	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	21.7		0.75	5.00	ug/Kg
74-97-5	Bromochloromethane	21.0		1.20	5.00	ug/Kg
67-66-3	Chloroform	21.5		0.84	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.1		0.93	5.00	ug/Kg
108-87-2	Methylcyclohexane	19.7		0.91	5.00	ug/Kg
71-43-2	Benzene	20.8		0.79	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	20.9		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.4		0.81	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	21.1		0.91	5.00	ug/Kg
75-27-4	Bromodichloromethane	20.8		0.78	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	94.0		3.60	25.0	ug/Kg
108-88-3	Toluene	20.5		0.78	5.00	ug/Kg

Report of Analysis

Client:	CDM Smith		Date Collected:	
Project:	South River WM Replacement		Date Received:	
Client Sample ID:	VY0627SBSD01	SDG No.:	Q2436	
Lab Sample ID:	VY0627SBSD01	Matrix:	SOIL	
Analytical Method:	8260D	% Solid:	100	
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022855.D	1		06/27/25 12:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	20.4		0.65	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	21.1		0.62	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	20.3		0.92	5.00	ug/Kg
591-78-6	2-Hexanone	95.1		3.70	25.0	ug/Kg
124-48-1	Dibromochloromethane	20.1		0.87	5.00	ug/Kg
106-93-4	1,2-Dibromoethane	20.0		0.88	5.00	ug/Kg
127-18-4	Tetrachloroethene	20.5		1.10	5.00	ug/Kg
108-90-7	Chlorobenzene	21.0		0.91	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.9		0.67	5.00	ug/Kg
179601-23-1	m/p-Xylenes	40.8		1.20	10.0	ug/Kg
95-47-6	o-Xylene	20.7		0.82	5.00	ug/Kg
100-42-5	Styrene	20.3		0.71	5.00	ug/Kg
75-25-2	Bromoform	19.0		0.86	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.6		0.78	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	20.0		1.20	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	21.0		1.70	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	21.2		1.60	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.4		1.50	5.00	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	18.7		1.80	5.00	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	20.0		3.00	5.00	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	19.4		3.20	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.8		63 - 155	104%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 - 134	103%	SPK: 50
2037-26-5	Toluene-d8	51.0		74 - 123	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.2		17 - 146	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	393000	7.713			
540-36-3	1,4-Difluorobenzene	687000	8.616			
3114-55-4	Chlorobenzene-d5	588000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	274000	13.346			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	South River WM Replacement	Date Received:	
Client Sample ID:	VY0627SBSD01	SDG No.:	Q2436
Lab Sample ID:	VY0627SBSD01	Matrix:	SOIL
Analytical Method:	8260D	% Solid:	100
Sample Wt/Vol:	5	Units:	g
Soil Aliquot Vol:			uL
GC Column:	RXI-624	ID :	0.25
Prep Method :		Level :	LOW
		Final Vol:	5000
		Test:	VOC-TCLVOA-10

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY022855.D	1		06/27/25 12:22	VY062725

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2436</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q2436</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q2436</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>06/23/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>13:38</u>
			<u>15:31</u>

LAB FILE ID:	RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D			
	RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D			
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.424	0.456	0.474	0.424	0.404	0.384	0.428	7.7
Chloromethane	0.837	0.921	0.865	0.793	0.758	0.724	0.816	8.9
Vinyl Chloride	0.934	1.099	1.091	1.045	0.993	0.958	1.020	6.8
Bromomethane	0.784	0.885	0.854	0.771	0.760	0.756	0.802	6.8
Chloroethane	0.649	0.736	0.722	0.694	0.673	0.640	0.686	5.6
Trichlorofluoromethane	0.999	1.180	1.219	1.166	1.127	1.085	1.129	7
1,1,2-Trichlorotrifluoroethane	0.508	0.560	0.547	0.515	0.492	0.474	0.516	6.3
1,1-Dichloroethene	0.478	0.539	0.524	0.514	0.500	0.483	0.506	4.7
Acetone	0.117	0.124	0.114	0.095	0.096	0.087	0.105	13.9
Carbon Disulfide	1.516	1.705	1.731	1.667	1.625	1.566	1.635	5.1
Methyl tert-butyl Ether	1.173	1.398	1.396	1.435	1.460	1.405	1.378	7.5
Methyl Acetate	0.272	0.358	0.440	0.351	0.353	0.322	0.349	15.7
Methylene Chloride	0.840	0.777	0.664	0.590	0.578	0.548	0.666	17.7
trans-1,2-Dichloroethene	0.521	0.604	0.597	0.592	0.581	0.575	0.578	5.2
1,1-Dichloroethane	0.949	1.075	1.079	1.077	1.055	1.030	1.044	4.8
Cyclohexane	0.998	1.021	0.988	0.946	0.905	0.894	0.959	5.4
2-Butanone	0.145	0.160	0.160	0.153	0.156	0.147	0.154	4.4
Carbon Tetrachloride	0.439	0.498	0.507	0.491	0.492	0.491	0.486	5
cis-1,2-Dichloroethene	0.606	0.689	0.687	0.685	0.687	0.678	0.672	4.8
Bromochloromethane	0.437	0.431	0.437	0.459	0.443	0.427	0.439	2.6
Chloroform	0.986	1.130	1.099	1.096	1.084	1.059	1.076	4.6
1,1,1-Trichloroethane	0.847	0.945	0.973	0.950	0.939	0.923	0.929	4.7
Methylcyclohexane	0.543	0.589	0.610	0.618	0.608	0.611	0.596	4.7
Benzene	1.248	1.433	1.451	1.464	1.467	1.440	1.417	5.9
1,2-Dichloroethane	0.335	0.397	0.402	0.400	0.404	0.392	0.388	6.8
Trichloroethene	0.305	0.364	0.382	0.372	0.360	0.350	0.356	7.6
1,2-Dichloropropane	0.289	0.339	0.345	0.339	0.341	0.337	0.332	6.4
Bromodichloromethane	0.422	0.495	0.496	0.498	0.504	0.498	0.485	6.4
4-Methyl-2-Pentanone	0.168	0.201	0.215	0.226	0.230	0.221	0.210	10.9
Toluene	0.747	0.873	0.908	0.926	0.955	0.954	0.894	8.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>CAMP02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2436</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q2436</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q2436</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>06/23/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>13:38</u>
			<u>15:31</u>

LAB FILE ID:		RRF005 = VY022776.D		RRF010 = VY022777.D		RRF020 = VY022778.D		
		RRF050 = VY022779.D		RRF100 = VY022780.D		RRF150 = VY022781.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.355	0.430	0.438	0.451	0.473	0.473	0.437	10
cis-1,3-Dichloropropene	0.412	0.503	0.523	0.524	0.540	0.538	0.506	9.6
1,1,2-Trichloroethane	0.207	0.249	0.249	0.253	0.255	0.250	0.244	7.4
2-Hexanone	0.115	0.140	0.145	0.151	0.157	0.149	0.143	10.5
Dibromochloromethane	0.260	0.315	0.321	0.329	0.336	0.329	0.315	8.8
1,2-Dibromoethane	0.193	0.231	0.229	0.237	0.244	0.236	0.228	7.8
Tetrachloroethene	0.399	0.465	0.535	0.515	0.473	0.446	0.472	10.3
Chlorobenzene	0.981	1.110	1.131	1.126	1.130	1.114	1.099	5.3
Ethyl Benzene	1.644	1.881	1.971	2.029	2.040	2.018	1.930	7.9
m/p-Xylenes	0.624	0.722	0.759	0.782	0.800	0.791	0.746	8.8
o-Xylene	0.578	0.674	0.708	0.734	0.759	0.765	0.703	10
Styrene	0.926	1.108	1.165	1.249	1.309	1.309	1.178	12.5
Bromoform	0.178	0.204	0.203	0.212	0.225	0.220	0.207	8
Isopropylbenzene	3.354	3.764	3.823	3.778	3.709	3.759	3.698	4.7
1,1,2,2-Tetrachloroethane	0.597	0.659	0.566	0.567	0.594	0.593	0.596	5.6
1,3-Dichlorobenzene	1.546	1.660	1.692	1.708	1.750	1.744	1.683	4.5
1,4-Dichlorobenzene	1.564	1.740	1.688	1.685	1.690	1.666	1.672	3.5
1,2-Dichlorobenzene	1.395	1.488	1.502	1.499	1.515	1.502	1.483	3
1,2-Dibromo-3-Chloropropane	0.102	0.101	0.103	0.103	0.102	0.096	0.101	2.7
1,2,4-Trichlorobenzene	0.778	0.841	0.848	0.843	0.871	0.845	0.838	3.7
1,2,3-Trichlorobenzene	0.679	0.723	0.735	0.728	0.751	0.727	0.724	3.3
1,2-Dichloroethane-d4	0.568	0.550	0.557	0.559	0.571	0.545	0.558	1.8
Dibromofluoromethane	0.306	0.297	0.295	0.304	0.314	0.308	0.304	2.3
Toluene-d8	1.182	1.148	1.186	1.215	1.262	1.247	1.207	3.6
4-Bromofluorobenzene	0.368	0.362	0.370	0.385	0.423	0.421	0.388	7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG No.: Q2436

Instrument ID: MSVOA_Y Calibration Date/Time: 06/27/2025 11:00

Lab File ID: VY022852.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.428		0	20
Chloromethane	0.816	0.857	0.1	5.03	20
Vinyl Chloride	1.020	0.985		-3.43	20
Bromomethane	0.802	0.846		5.49	20
Chloroethane	0.686	0.683		-0.44	20
Trichlorofluoromethane	1.129	1.067		-5.49	20
1,1,2-Trichlorotrifluoroethane	0.516	0.539		4.46	20
1,1-Dichloroethene	0.506	0.535		5.73	20
Acetone	0.105	0.121		15.24	20
Carbon Disulfide	1.635	1.684		3	20
Methyl tert-butyl Ether	1.378	1.416		2.76	20
Methyl Acetate	0.349	0.300		-14.04	20
Methylene Chloride	0.666	0.659		-1.05	20
trans-1,2-Dichloroethene	0.578	0.608		5.19	20
1,1-Dichloroethane	1.044	1.117	0.1	6.99	20
Cyclohexane	0.959	0.990		3.23	20
2-Butanone	0.154	0.156		1.3	20
Carbon Tetrachloride	0.486	0.532		9.47	20
cis-1,2-Dichloroethene	0.672	0.707		5.21	20
Bromochloromethane	0.439	0.420		-4.33	20
Chloroform	1.076	1.114		3.63	20
1,1,1-Trichloroethane	0.929	0.989		6.46	20
Methylcyclohexane	0.596	0.671		12.58	20
Benzene	1.417	1.538		8.54	20
1,2-Dichloroethane	0.388	0.407		4.9	20
Trichloroethene	0.356	0.381		7.02	20
1,2-Dichloropropane	0.332	0.355		6.93	20
Bromodichloromethane	0.485	0.520		7.22	20
4-Methyl-2-Pentanone	0.210	0.212		0.95	20
Toluene	0.894	0.991		10.85	20
t-1,3-Dichloropropene	0.437	0.473		8.24	20
cis-1,3-Dichloropropene	0.506	0.555		9.68	20
1,1,2-Trichloroethane	0.244	0.251		2.87	20
2-Hexanone	0.143	0.151		5.59	20
Dibromochloromethane	0.315	0.331		5.08	20
1,2-Dibromoethane	0.228	0.233		2.19	20
Tetrachloroethene	0.472	0.495		4.87	20
Chlorobenzene	1.099	1.211	0.3	10.19	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG No.: Q2436

Instrument ID: MSVOA_Y Calibration Date/Time: 06/27/2025 11:00

Lab File ID: VY022852.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.214		14.72	20
m/p-Xylenes	0.746	0.854		14.48	20
o-Xylene	0.703	0.804		14.37	20
Styrene	1.178	1.339		13.67	20
Bromoform	0.207	0.211	0.1	1.93	20
Isopropylbenzene	3.698	4.232		14.44	20
1,1,2,2-Tetrachloroethane	0.596	0.622	0.3	4.36	20
1,3-Dichlorobenzene	1.683	1.862		10.64	20
1,4-Dichlorobenzene	1.672	1.820		8.85	20
1,2-Dichlorobenzene	1.483	1.590		7.22	20
1,2-Dibromo-3-Chloropropane	0.101	0.099		-1.98	20
1,2,4-Trichlorobenzene	0.838	0.867		3.46	20
1,2,3-Trichlorobenzene	0.724	0.718		-0.83	20
1,2-Dichloroethane-d4	0.558	0.538		-3.58	20
Dibromofluoromethane	0.304	0.307		0.99	20
Toluene-d8	1.207	1.256		4.06	20
4-Bromofluorobenzene	0.388	0.399		2.84	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG No.: Q2436

Instrument ID: MSVOA_Y Calibration Date/Time: 06/30/2025 11:03

Lab File ID: VY022872.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.428	0.394		-7.94	20
Chloromethane	0.816	0.771	0.1	-5.51	20
Vinyl Chloride	1.020	0.967		-5.2	20
Bromomethane	0.802	0.735		-8.35	20
Chloroethane	0.686	0.666		-2.91	20
Trichlorofluoromethane	1.129	1.080		-4.34	20
1,1,2-Trichlorotrifluoroethane	0.516	0.554		7.36	20
1,1-Dichloroethene	0.506	0.562		11.07	20
Acetone	0.105	0.124		18.09	20
Carbon Disulfide	1.635	1.820		11.31	20
Methyl tert-butyl Ether	1.378	1.494		8.42	20
Methyl Acetate	0.349	0.300		-14.04	20
Methylene Chloride	0.666	0.725		8.86	20
trans-1,2-Dichloroethene	0.578	0.648		12.11	20
1,1-Dichloroethane	1.044	1.203	0.1	15.23	20
Cyclohexane	0.959	1.049		9.39	20
2-Butanone	0.154	0.166		7.79	20
Carbon Tetrachloride	0.486	0.536		10.29	20
cis-1,2-Dichloroethene	0.672	0.765		13.84	20
Bromochloromethane	0.439	0.448		2.05	20
Chloroform	1.076	1.198		11.44	20
1,1,1-Trichloroethane	0.929	1.028		10.66	20
Methylcyclohexane	0.596	0.679		13.93	20
Benzene	1.417	1.617		14.11	20
1,2-Dichloroethane	0.388	0.429		10.57	20
Trichloroethene	0.356	0.404		13.48	20
1,2-Dichloropropane	0.332	0.380		14.46	20
Bromodichloromethane	0.485	0.546		12.58	20
4-Methyl-2-Pentanone	0.210	0.221		5.24	20
Toluene	0.894	1.029		15.1	20
t-1,3-Dichloropropene	0.437	0.496		13.5	20
cis-1,3-Dichloropropene	0.506	0.585		15.61	20
1,1,2-Trichloroethane	0.244	0.262		7.38	20
2-Hexanone	0.143	0.150		4.89	20
Dibromochloromethane	0.315	0.338		7.3	20
1,2-Dibromoethane	0.228	0.244		7.02	20
Tetrachloroethene	0.472	0.539		14.19	20
Chlorobenzene	1.099	1.262	0.3	14.83	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG No.: Q2436

Instrument ID: MSVOA_Y Calibration Date/Time: 06/30/2025 11:03

Lab File ID: VY022872.D Init. Calib. Date(s): 06/23/2025 06/23/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:38 15:31

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.930	2.279		18.08	20
m/p-Xylenes	0.746	0.876		17.43	20
o-Xylene	0.703	0.826		17.5	20
Styrene	1.178	1.368		16.13	20
Bromoform	0.207	0.217	0.1	4.83	20
Isopropylbenzene	3.698	4.437		19.98	20
1,1,2,2-Tetrachloroethane	0.596	0.623	0.3	4.53	20
1,3-Dichlorobenzene	1.683	1.949		15.81	20
1,4-Dichlorobenzene	1.672	1.902		13.76	20
1,2-Dichlorobenzene	1.483	1.663		12.14	20
1,2-Dibromo-3-Chloropropane	0.101	0.098		-2.97	20
1,2,4-Trichlorobenzene	0.838	0.907		8.23	20
1,2,3-Trichlorobenzene	0.724	0.733		1.24	20
1,2-Dichloroethane-d4	0.558	0.541		-3.05	20
Dibromofluoromethane	0.304	0.310		1.97	20
Toluene-d8	1.207	1.238		2.57	20
4-Bromofluorobenzene	0.388	0.386		-0.51	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.