

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2436 SAS No.: Q2436 SDG NO.: Q2436
 Lab File ID: VY022852.D Date Analyzed: 06/27/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	460120	7.71	754723	8.62	647658	11.42
UPPER LIMIT	920240	8.213	1509450	9.122	1295320	11.92
LOWER LIMIT	230060	7.213	377362	8.122	323829	10.92
EPA SAMPLE NO.						
TP-70	370411	7.71	661194	8.62	597332	11.42
TP-69	340803	7.71	625770	8.62	598373	11.42
TP-83	337210	7.71	630445	8.62	584339	11.41
TP-87	348864	7.71	632840	8.62	540565	11.41
TP-100	312427	7.71	578070	8.62	554007	11.41
TP-99	301121	7.71	558142	8.62	536522	11.41
TP-82	318610	7.71	604014	8.62	582898	11.41
VY0627SBL01	358693	7.71	655911	8.62	603519	11.42
VY0627SBS01	438526	7.71	740220	8.62	622783	11.42
VY0627SBSD01	392671	7.71	686703	8.62	587941	11.42

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.